

College of Saint Elizabeth
School of Graduate and Continuing Studies
2 Convent Road, Morristown, NJ 07960

p-3

29-12852-02
03014737**FAX**

To: Kathy Modes From: Donna Howell

Fax: _____ Fax: _____

Phone: _____ Phone: 973-290-4094

Subject: Wipe test - Final Date: 1/23/06

Comments: _____

control # 138124

138124

SURVEY RECORD

Biology Department, College of Saint Elizabeth

Date: 3/27/97**WIPE TEST RESULTS:**

	Location	cpm (^{51}Cr window)	dpm*
1	Background	44	-
2	Fume hood	54	160
3	Sink top	46	32
4	Sink	46	32
5	Work bench	41	-
6	little Centrifuge	34	64
7	has Sterile hood centrifuge	42	-
8	Gamma counter	46	32
9	Waste room floor	51	114
10	vortex	60	260

* dpm = 16.3 (cpm-background)

Wipe test limit for ^{51}Cr is 1000 dpm per 100 cm²**AREA SURVEY RESULTS:**

Location	Exposure (mR/hr)**
Background	.01
Fume hood	.02
Sink top	.02
Sink	.01
Work bench	.02
Centrifuge	.03
Sterile hood	.02
Gamma counter	.03
Waste room floor	.02

**Use Ludlum Model 3 survey meter