

OFFICIAL RECORD COPY

MATERIALS LICENSE

Amendment No. 2

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p>Licensee</p> <p>1. Luis M. Cruz, M.D.</p> <p>2. Dr. Basora #55 Norte Oficina 1 J Mayagüez, Puerto Rico 00680</p>	<p>In accordance with letter dated June 15, 1995</p> <p>3. License Number 52-25102-01</p> <p>is renewed in its entirety to read as follows:</p> <p>4. Expiration Date September 30, 2000</p> <p>5. Docket or Reference No. 030-31733</p>
---	--

<p>6. Byproduct, Source, and/or Special Nuclear Material</p> <p>A. Strontium 90</p>	<p>7. Chemical and/or Physical Form</p> <p>A. Sealed source in an applicator registered pursuant to 10 CFR 32.210 or an equivalent Agreement State Regulation.</p>	<p>8. Maximum Amount that Licensee May Possess at Any One Time Under This License</p> <p>A. 150 millicuries (5.55 GBq)</p>
<p>9. Authorized Use:</p> <p>A. Medical treatment of superficial eye conditions as identified in 10 CFR 35.400(e).</p>		

CONDITIONS

10. Location for use: Dr. Basora #55 Norte
Oficina 1 J
Mayagüez, Puerto Rico 00680
11. Radiation Safety Officer: Luis M. Cruz, M.D.
12. Authorized user: Luis M. Cruz, M.D.
13. Sealed sources containing licensed material shall not be opened or removed from the source holder by the licensee.
14. The licensee shall maintain records of information important to safe and effective decommissioning at Dr. Basora #55 Norte, Oficina 1 J, Mayaguez, Puerto Rico, per the provision of 10 CFR 30.35(g) until this license is terminated by the Commission.

9510030156 950912
PDR ADOCK 03031733
C PDR

NMSS/RGNI MATERIALS-001

ML20

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License Number

52-25102-01

Docket or Reference Number

030-31733

Amendment No. 2

(cont'd)

CONDITIONS

15. The eye applicator device shall have a locked container designed to prevent unauthorized or accidental removal of the sealed source from its shielded position. The eye applicator must be locked in its storage container when in transport, storage or when not under the direct surveillance of an authorized user.
16. Except as specifically provided otherwise in the license and as provided in 10 CFR 35.51, the licensee shall conduct its program in accordance with statements, representations, and procedures contained in the documents including any enclosures, listed below. The Nuclear Regulatory Commission's regulations shall govern unless the statements, representations and procedures in the licensee's application and correspondence are more restrictive than the regulations.

- Application dated June 15, 1995 [renewal, procedures addendum]

FOR THE U. S. NUCLEAR REGULATORY COMMISSION


HECTOR BERMUDEZ

Date

SEP 12 1995

9/12/95

By


Region II, Nuclear Materials Licensing Section
101 Marietta Street, Suite 2900
Atlanta, GA 30323-0199