

## SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

## 1. LICENSEE/LOCATION INSPECTED:

Western Missouri Medical Center  
403 Burkarth Road  
Warrensburg, MO 64093

## 2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission  
Region III  
2443 Warrenville Road, Suite 210  
Lisle, Illinois 60532-4351

REPORT NUMBER(S) 2006-001

## 3. DOCKET NUMBER(S)

030-35258

## 4. LICENSEE NUMBER(S)

24-32222-01

## 5. DATE(S) OF INSPECTION

Feb. 7, 2006

## LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection

- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- ☐ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

## Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title

Printed Name

Signature

Date

LICENSEE'S  
REPRESENTATIVE

NRC INSPECTOR

Deborah A. Piskura



2/7/06

## Docket File Information

SAFETY INSPECTION REPORT  
AND COMPLIANCE INSPECTION

1. LICENSEE <b>Western Missouri Medical Center</b> REPORT NUMBER(S) <b>2006-001</b>		2. NRC/REGIONAL OFFICE <b>Region III</b> <b>2443 Warrenville Road, Suite 210</b> <b>Lisle, IL 60532</b>	
3. DOCKET NUMBER(S) <b>030-35258</b>	4. LICENSE NUMBER(S) <b>24-32222-01</b>	5. DATE(S) OF INSPECTION <b>Feb. 7, 2006</b>	
6. INSPECTION PROCEDURES <b>87130</b>	7. INSPECTION FOCUS AREAS <b>03.01, 03.02, 03.03, 03.04, 03.05, 03.06, 03.07, and 03.08</b>		
<b>SUPPLEMENTAL INSPECTION INFORMATION</b>			
1. PROGRAM CODE(S) <b>02121</b>	2. PRIORITY <b>G 5</b>	3. LICENSEE CONTACT <b>Cathy Jo Parsells, D.O., RSO</b>	4. TELEPHONE NUMBER <b>660.747.2500</b>
<input checked="" type="checkbox"/> Main Office Inspection		Next Inspection Date: <b>Feb. 2011</b>	
<input type="checkbox"/> Field			
<input type="checkbox"/> Temporary Job Site			

## PROGRAM SCOPE

This licensee was a community hospital authorized to use licensed material permitted by Sections 35.100 and 35.200. The nuclear medicine department was staffed with two full-time technologists who performed approximately 50-60 diagnostic studies monthly. The licensee received unit doses from a licensed radiopharmacy. The clinic retained the services of a consulting physicist who audited the radiation safety program on a quarterly basis.

This inspection consisted of interviews with licensee personnel, a review of select records, tour of the nuclear medicine department, and independent measurements. The inspector observed licensee personnel demonstrate dose calibrator QA tests, package surveys, and radiation/contamination area surveys.