

## SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

## 1. LICENSEE/LOCATION INSPECTED

Medical Imaging, Inc.  
9501 North Oak Traffic Way  
Kansas City, MO 64155

## 2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission  
Region III  
2443 Warrenville Road, Suite 210  
Lisle, Illinois 60532-4351

REPORT NUMBER(S) 2006-001

## 3. DOCKET NUMBER(S)

030-18630

## 4. LICENSEE NUMBER(S)

24-24423-01

## 5. DATE(S) OF INSPECTION

Feb. 6, 2006

## LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection

- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- ☐ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

## Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura	<i>Deborah A. Piskura</i>	2/6/06

*RP*

## Docket File Information

SAFETY INSPECTION REPORT  
AND COMPLIANCE INSPECTION

1. LICENSEE

Medical Imaging, Inc.

REPORT NUMBER(S) 2006-001

2. NRC/REGIONAL OFFICE

Region III  
2443 Warrenville Road, Suite 210  
Lisle, IL 60532

3. DOCKET NUMBER(S)

030-18630

4. LICENSE NUMBER(S)

24-24423-01

5. DATE(S) OF INSPECTION

Feb. 6, 2006

6. INSPECTION PROCEDURES

87130 &amp; 87131

7. INSPECTION FOCUS AREAS

03.01, 03.02, 03.03, 03.04, 03.05, 03.06, 03.07, and 03.08

## SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S)

02200

2. PRIORITY

G 3

3. LICENSEE CONTACT

John Stephen Dykstra, D.O.

4. TELEPHONE NUMBER

816.455.0661

☒ Main Office Inspection

Next Inspection Date: Feb. 2009

☐ Field☐ Temporary Job Site

## PROGRAM SCOPE

This licensee was a doctors owned multi-specialty diagnostic office and authorized to use licensed material permitted by Sections 35.100, 35.200 and 35.300. The nuclear medicine department was staffed with one full-time technologist who performed approximately 40-50 diagnostic procedures monthly. Typically, in a year, the clinic administered 3-5 dosages for hyperthyroidism. The licensee received unit doses and I-131 capsules from a licensed radiopharmacy. The clinic retained the services of a consulting physicist who audited the radiation safety program on a quarterly basis.

This inspection consisted of interviews with licensee personnel, a review of select records, tour of the nuclear medicine department, and independent measurements. The inspector observed licensee personnel demonstrate dose calibrator QA tests, and radiation surveys.