| U.S. NUCLEAR REGULATORY COMMISSION<br><br>TELEPHONE CONVERSATION RECORD | Date: 2/8/2006<br><br>Time: 10:00 am |
|---|---|

| Mail Control or Report No(s). | Insp. No. 2006-001 | License No(s). 37-00554-03 | Docket No(s). 03002961 |
|---|---|---|---|

| | |
|---|---|
| Name of Licensee: | Sacred Heart Hospital |
| Name of Participant(s): | Belle Rodrigues, Director, Radiology and Nuclear Medicine |
| Telephone No. | 610-776-5117 |
| Subject:<br>(NOTE: This will be used as the Documents Title in ADAMS) | Additional information hot lab security |
| Summary: | Ms. Rodrigues informed me that the automatic door closure for the Nuclear Medicine Department hot lab was installed today. |
| Action Required: | Place in ADAMS. Review during next inspection. |

Document Availability: X Publicly Available ___ Non-Publicly Available

X Non-Sensitive ___ Non-Sensitive Copyright ___ Sensitive ___ Sensitive Copyright

___ Immediate Release X Normal Release ___ Delay Release Date

Prepared & SISP Review Completed By: ***/ RA /*** Donna M. Janda Date: 2/8/2006