SAINT BARNABAS
HEALTH CARE SYSTEM
Irvington General Hospital

RONALD J. DEL MAURO
President and Chief Executive Officer
Saint Barnabas Health Care System

PATRICIA CARROLL
Executive Director

January 26, 2006

NMSB1

RECEIVED
REGION 1
2006 FEB 15 PM 1:21

United States Nuclear Regulatory Commission
Region I
475 Allendale Road
King of Prussia, PA 19406-1415

Re: License No. 29-07987-02 03002508

To Whom It May Concern:

Please be advised that Irvington General Hospital will cease operation as of January 31, 2006.

Therefore, we are requesting termination of out radioactive materials license.

All radioactive materials have been removed and properly disposed of. Attached please find the closeout survey for our Nuclear Medicine Department.

If you have any questions or require additional information concerning this matter, please contact Tamara Cunningham at (973) 322-4061.

Sincerely,

Patricia Carroll

Patricia Carroll
Executive Director

138428

IRVINGTON MATERIALS-002

SAINT BARNABAS HEALTH CARE SYSTEM

Irvington General Hospital

CLOSE OUT - Jan 26, 2006 SURVEY

RONALD J. DEL MAURO
President and Chief Executive Officer
Saint Barnabas Health Care System

NUCLEAR MEDICINE
DAILY SURVEYS & WEEKLY WIPE TEST



AREA KEY

1. OLD CAMERA & TABLE
2. NEW CAMERA & TABLE
3. DESK
4. PROCESSING STATION
5. PROCESSING STATION NEW CAMERA
6. XENON TRAP
7. MAIN ENTRANCE & INJECTION AREA
8. BACK ENTRANCE
9. FLOOR
10. HOT LAB FLOOR
11. HOT LAB COUNTER
12. L-BLOCK & PREP. AREA
13. SHARPS CONTAINERS
* 14. STRESS LAB. — See attached report

EQUIPMENT - CAPINTEC CAPRAC & BIODEX ATOMLAB 950 WELL **METERS:** A - LUDLUM 14C # S/N 77564 CAL- 3/17/05

ACTIONS LAVELS - HOT LAB - 2 mR/HR & 2000 DPM
CAMERA ROOM - 0.5 mR/HR & 2000 DPM

~~WEEK OF~~ Performed By Vincent M Iannuzzi MS, DABR Medical Physicist

Jan 26, 2006

All areas below are surveyed in mR/hr

Date	Initial	Survey Meter	Check source	BKG mR/hr	1	2	3	4	5	6	7	8	9	10	11	12	13	14
1/26	VE	A	7.0	0.03	.03	.04	.04	.03	.03	.03	.03	.03	.03	.03	.03	.03	.03	.03

Weekly Wipe Test; Results below in dpm **Well Efficiency** 93.59% dpm/cpm

Date	Initial	Well Counter	BKG cpm	Wipes in	1	2	3	4	5	6	7	8	9	10	11	12	13	14
1/26	VE		402	DPM	0	0	13	0	24	0	26	0	0	0	8	17	0	*

IRVINGTON GENERAL – CLOSEOUT SURVEY CARDIAC STRESS AREA

DATE – JANUARY 4, 2006

EQUIPMENT – LUDLUM 14C SURVEY METER
SN – 77524
CALI – 3/17/05

BIODEX ATOM LAB 950 UPTAKE PROBE AND WELL COUNTER
COBALT – 57 EFFICIENCY 93.5%

FLOOR PLAN –



LOCATION BKGRD – 365	CPM	NET DPM
1.SHARPS CONTAINER	366	1
2, WASTE CONTAINER	357	0
3. TREADMILL	372	8
4. BED	363	0
5. FLOOR	358	0
6. EKG MONITOR	381	17

CLOSEOUT SURVEY CARDIAC STRESS AREA - CONTINUED

LOCATION	CPM	NET DPM
7. CRASH CART	376	12
8. FLOOR	374	10
9. FLOOR	388	25
10. FLOOR	374	10

PERFORMED BY:

Vincent M Immerso 11/4/2016

Vincent M. Immerso, MS., DABR.
Medical Physicist

138428

SAINT BARNABAS
HEALTH CARE SYSTEM
Newark Beth Israel Medical Center
201 Lyons Avenue at Osborne Terrace
Newark, New Jersey 07112

DNMS

7001 2510 0005 9688 4620

UNITED STATES POSTAGE
PITNEY BOWES
02 1A $ 04.64⁰
0004394369 FEB 13 2006
MAILED FROM ZIP CODE 07112

US Nuclear Regulatory Commission
Region I
475 Allendale Road
King of Prussia, PA 19406-1415

19406+1431-99 C002

This is to acknowledge the receipt of your letter/application dated 1/26/2006 (RECEIVED 2/15/2006) and to inform you that the initial processing which includes an administrative review has been performed.

☑ TERMINATION 29-07987-02
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

☐ Please provide to this office within 30 days of your receipt of this card

---

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 138428.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.

NRC FORM 532 (RI)
(6-96)

Sincerely,
Licensing Assistance Team Leader