

From: Richard McKinley
To: Sharon@npc-link.com
Date: 3/9/06 11:06AM
Subject: Bluefield Regional Medical Center renewal

License No.:47-19142-01
Docket No:03017038
Control No: 137905

Please note that you may not reply to this letter by return e-mail. Your reply must be in writing by letter or facsimile (610-337-5269). If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your application.

Dear Sharon:

With regard to your response facsimile of February 22, 2006, I need the following additional information:

- 1) In Item 3 you referenced a fax sent on 12/21/05, with room diagrams. This fax was for the HDR room. I need equivalent diagrams for the rooms where you put patients who are undergoing LDR brachytherapy or I-131 treatments.
- 2) Are the emergency source containers that you use for LDR sources of the same kind that you use for the HDR source?
- 3) Please confirm that your spot procedures for the HDR include confirming that the activity printed out from the HDR matches that obtained from an independent decay check.

Thank you very much.

Sincerely,
Richard McKinley
Health Physicist
U.S. NRC, RI

CC: RWm1

Mail Envelope Properties (441053E0.103 : 16 : 27049)

Subject: Bluefield Regional Medical Center, 47-19142-01, 03017038, Control
No. 137905
Creation Date: Thu, Mar 9, 2006 11:12 AM
From: Richard McKinley
Created By: RWM1@nrc.gov

Recipients

nrc.gov
kpl_po.KP_DO
LAT (LAT)

Post Office
kpl_po.KP_DO

Route
nrc.gov

Files	Size	Date & Time
MESSAGE	1547	Thursday, March 9, 2006 11:12 AM
Bluefield Regional Medical Center renewal.mlm		2024 Thursday, March 9,
2006 11:07 AM		

Options

Expiration Date: None
Priority: Standard
Reply Requested: No
Return Notification: None

Concealed Subject: No
Security: Standard