

FACSIMILE TRANSMISSION

To: Licensing Assistance Team Leader, Control Number 138339

From: Charles L. Chipley III, MS, DABR

Date: March 13, 2006


Number of Pages: 4 (incl. this cover sheet)

Q-8

55-17986-d
03013764

1) Per our conversation, please find the State of Maryland amendment 118 to radioactive materials license #MD-07-011-01, which follows. You will please note that it is addressed to me as Radiation Safety Officer, and that license condition 11A states that the radiation protection program shall be under my supervision.

2) I hope that this will help expedite the amendment of our NRC radioactive materials license number 55-17986-01, with control number above assigned.



-Charles L. Chipley III, MS, DABR
Licensed Medical Physicist
Schneider Regional Medical Center
9048 Sugar Estate
St. Thomas, VI 00802
340-776-8311 x4441
Fax 340-714-6316

138339

NMSS/RONI MATERIALS-002
TOTAL P.01

**MARYLAND DEPARTMENT OF THE ENVIRONMENT**

1800 Washington Boulevard • Baltimore MD 21230

410-537-3000 • 1-800-633-6101

bert L. Ehrlich, Jr.

Governor

Michael S. Steele

Lt. Governor

Kend P. Philbrick
Acting Secretary

FEB 10 2004

Charles Chipley, Radiation Safety Officer
Sinai Hospital of Baltimore
2401 W. Belvedere Avenue
Baltimore, Maryland 21215

RE: Radioactive Material License #MD-07-011-01

Dear Mr. Chipley:

Your requested amendment to radioactive materials license number MD-07-011-01 is enclosed. Please review it carefully to ensure that it reflects all modifications included in your letter received in this office on November 7, 2003.

Should you require further assistance, please contact Mr. Raymond E. Manley at (410) 537-3301. You may also reach our office toll-free (in Maryland only) by dialing 1-800-633-6101 and requesting extension 3301. Also, you may contact this office via facsimile at (410) 537-3198.

Sincerely,

Roland G. Fletcher, Program Manager III
Radiological Health Program

RGF/REM/BJP/cc

Enclosures: License amendment (118)
Code (02120)



Recycled Paper

www.mde.state.md.us

TTY Users 1-800-735-2258
Via Maryland Relay Service



DEPARTMENT OF THE ENVIRONMENT
RADIOLOGICAL HEALTH PROGRAM
RADIOACTIVE MATERIAL LICENSE

Page 5 of 9

License Number: MD-07-011-01

Amendment Number: 118 (CODE 02120)

LICENSE CONDITIONS

10. The authorized place of use is the licensee's address stated in Item 2. The licensee must notify the Radiological Health Program 30 days prior to vacating a permanent use address as is required by Section D.1301 of COMAR 26.12.01.01.
- 11A. The radiation protection program shall be under the supervision of Charles Chipley.
- 11B. Radioactive material shall be used by, or under the supervision of individuals designated by the Radiation Safety Committee.
12. The licensee shall comply with all appropriate provisions of COMAR 26.12.01.01 "Regulations for Control of Ionizing Radiation."
- 13A. Each sealed source containing radioactive material, other than Hydrogen-3 with a half-life greater than thirty (30) days and in any form other than gas shall be tested for leakage and/or contamination at intervals not to exceed six (6) months. In the absence of a certificate from a transferor indicating that a test has been made within six (6) months prior to the transfer, the sealed source shall not be put into use until tested. If there is reason to suspect that a sealed source might have been damaged, or might be leaking, it shall be tested for leakage before further use.
- 13B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the sealed source or from the surfaces of a device in which the sealed source is permanently mounted or stored on which one might expect contamination to accumulate.
- 13C. Records of leak tests shall be kept in units of microcuries and maintained for inspection by the Department.
- 13D. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Department regulations. A report shall be filed within five (5) days of the test with the Administrator, Radiological Health Program, 1800 Washington Boulevard, Suite 750, Baltimore, Maryland 21230, describing the equipment involved, the test results, and the corrective action taken.
- 13E. Test for leakage and/or contamination shall be performed by the licensee or by other persons specifically authorized by the Department, the U.S. Nuclear Regulatory Commission or another Agreement State to perform such services.



DEPARTMENT OF THE ENVIRONMENT
RADIOLOGICAL HEALTH PROGRAM
RADIOACTIVE MATERIAL LICENSE

Page 9 of 9

License Number: MD-07-011-01

Amendment Number: 118 (CODE 02120)

LICENSE CONDITIONS, CONTINUED

27. Except as specifically provided otherwise by this license, the licensee shall possess and use radioactive material authorized by this license in accordance with statements representations, and procedures contained in:

- Renewal application dated December 17, 2002
- Letter and attachments dated March 20, 2002
- Letter and attachments dated June 19, 2002
- Letter and attachments dated July 1, 2002
- Letter and attachments dated July 5, 2002
- Letter with attachments dated March 6, 2003
- Letter dated August 25, 2003 adding North American Scientific as a vendor for I-125 & Pd-103 seeds.
- Letter with attachment dated November 14, 2003, changing the radiation safety officer.

COMAR 26.12.01.01 "Regulations for Control of Ionizing Radiation" shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

FOR THE MARYLAND DEPARTMENT OF THE ENVIRONMENT

Roland G. Fletcher, Manager
Radiological Health Program

January 16, 2004

BJP

BJP 1/16/04

2/3/04