

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Professional Inspection and Testing Services, Inc.  
56 North Second Street  
Chambersburg, Pennsylvania 17201

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission  
Region I, 475 Allendale Road  
King of Prussia, Pennsylvania 19406-1415

REPORT 2006001

3. DOCKET NUMBER(S)

03032721

4. LICENSE NUMBER(S)

37-28744-01

5. DATE(S) OF INSPECTION

3/16/06

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- ☐ 1. Based on the inspection findings, no violations were identified.
- ☒ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

☐ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- ☒ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.



LICENSE CONDITION 13 REQUIRES ROUTINE LEAK TESTS. FOR TROXLER GAUGES THIS LEAK TEST INTERVAL IS EVERY SIX MONTHS FOR DEVICES NOT IN STORAGE.

CONTRARY TO THIS REQUIREMENT, TROXLER 3430 (SERIAL 21190) NUCLEAR GAUGE WAS ALLOWED TO REMAIN IN SERVICE WITH A LEAK TEST INTERVAL OF OVER 11 MONTHS.

THE LICENSEE AGREES TO CHANGE THE LEAK TEST COLLECTION SYSTEM TO ASSURE LEAK TESTS ARE CONDUCTED IN A TIMELY FASHION. THIS CHANGE WILL BE INTEGRATED WITH THE SIX MONTH INVENTORY REQUIREMENT.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	Rosemarie McNew		3/16/06
NRC INSPECTOR	James Schmidt		3/16/06