

CONTROL # 138496

From: Maha Mikhail, M.D.
Connecticut Multispecialty Group, P.C.
1260 Silas Deane Highway
Suite 106
Wethersfield, CT 06109

Q-8

To: USNRC, Region I Office
NMSB - Medical Licensing
475 Allendale Road
King of Prussia, PA 19406

03008163

Re: Amendment for NRC Materials License # 06-14854-01

Dear Sir or Madam:

The following items are submitted for your review and approval in an amendment.


ITEM 1

The Nuclear Cardiology service intends to use Gd-153 sealed sources for patient attenuation correction of nuclear cardiac studies. Presently, the materials license does not authorize this source type (10 CFR 35.500) and we are requesting that Dr. Maha Mikhail and Dr. Carol Gemayel be listed as the authorized users. The gamma camera vendor has indicated the sources are listed in the SSDR # CA 0406S204S, with an activity of 240 mCi per source. A total of 960 mCi is requested for possession on our license. The sources will be stored in the existing hot lab when sources are exchanged due to expiration.

We appreciate your efforts in reviewing this amendment request and we look forward to a continued safe and effective program with radioactive materials. There is no cost for an amendment, per the fees schedule of 10 CFR 170.31, 7.C.

If you have any questions or desire additional information, please contact Lori Pedemonti, Practice Manager, Connecticut Multispecialty Group, P.C. 860-241-6077.

Respectfully submitted,


Maha Mikhail, M.D.

Post-it® Fax Note	7671	Date	3-24-06	# of pages	1
To	SHIRLEY	From	DENISE CONDRON		
Co./Dept.	NRC	Co.	CMSG		
Phone #		Phone #	(203)288-6860		
Fax #		Fax #			

138496

NMSB/RGNI MATERIALS-002