

March 14, 2006

US Nuclear Regulatory Commission Region IV  
Nuclear Materials Licensing Branch  
611 Ryan Plaza Drive  
Suite 400  
Arlington, Texas 76011-8064

**SUBJECT: Decontamination of Nuclear Medicine Area**  
**LICENSE: 11-27312-01**

This letter is in followup to the decontamination notification letter sent December 15, 2005 to the Nuclear Regulatory Commission, regarding our Boise facility: 190 E Bannock St, Boise, Idaho 83712. Specifically addressing the questions:

- 1- Make and model of survey meter used: Biodex 14C, #99926. The survey instrument was calibrated on 2/25/05 and properly tested before the surveys were performed. Make and model of well counter: Biodex Atom Lab 950, #1201099859.
- 2- Units/area of wipe test, from diagrams: cpm/100 sqcm
- 3- Waste storage survey/wipe test results: attached

For further information, please contact me at 208-706-1412.

Sincerely,

Jefferson Fairbanks, PhD  
Radiation Safety Officer

470802

# **Basement Storage Area**

<b>Area</b>	<b>mR/h</b>	<b>Bkg</b>	<b>Wipe</b>	
1	0.02	546	450	
2	0.02	546	552	
3	0.02	546	549	
4	0.02	546	522	
5	0.02	546	510	

