



1001 EAST SECOND STREET • COUDERSPORT, PA 16915-9762 • 814-274-9300

March 21, 2006

*NMSB1*

U.S. Nuclear Regulatory Commission  
Materials Licensing Branch, Region I  
475 Allendale Road  
King of Prussia, PA 19406

*03003202*

RE: Amendment to Radioactive Material License No. 37-13798-01  
Charles Cole Memorial Hospital

To Whom It May Concern:

We have moved our area of RAM storage as outlined in our amendment request/letter dated February 2, 2006. Please delete our old area from our license. Enclosed is our close-out survey of the area.

If you have any questions or require additional clarification, please do not hesitate to contact me.

Sincerely,

David B. Acker  
Chief Executive Officer

kjn  
Enclosure

RECEIVED  
REGION I  
2006 MAR 24 AM 11:19

*138628*

NMSS/RGN MATERIALS-002  
REF. 138375



1001 EAST SECOND STREET • COUDERSPORT, PA 16915-9762 • 814-274-9300

March 21, 2006

U.S. Nuclear Regulatory Commission  
Materials Licensing Branch, Region I  
475 Allendale Road  
King of Prussia, PA 19406

RE: Close-out Survey for Radioactive Material License No. 37-13798-01  
Charles Cole Memorial Hospital

Gentlemen:

We have decommissioned our old area of RAM storage. Only 35.100 and 200 materials were ever stored in the room. The following is a close-out survey performed on March 15, 2006:

A gross contamination smear was taken on all walls and on the floor, door and doorknob. The smear covered 80% of the surface area of the walls and floor. The smear was counted in a Capintec Caprac well counter with an efficiency for Tc-99m which is greater than 90% and minimum detectable activity of 199. Area radiation level surveys were taken at each smear site. The background for the area surveys was 0.03 mR/hr using an Eberline E-520 GM survey meter last calibrated on November 1, 2005. All readings were 0.03 mR/hr.

Background for the wipe test was 458. CPM of the gross smear was 450. Wipe survey indicated no activity present or 0 dpm. This is consistent with the exclusive storage of Tc-99m in this area with the last use of the room occurring more than two weeks ago.

We have released the room for unrestricted use. If you have any questions or require additional clarification, please do not hesitate to contact me.

Sincerely,

David B. Acker  
Chief Executive Officer

kjn

This is to acknowledge the receipt of your letter/application dated

3/21/2006, and to inform you that the initial processing which includes an administrative review has been performed.

☒ AMEND. 37-13798-01  
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

☐ Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned Mail Control Number 138628.  
When calling to inquire about this action, please refer to this control number.  
You may call us on (610) 337-5398, or 337-5260.