



11438 Cronridge Drive
Suite G
Owings Mills, Maryland 21117
Tel: [410] 363-8344
Fax: [410] 363-9978
E-mail: ctibaltimore@cti-consultants.com

June 13, 2006

United States Nuclear Regulatory Commission
Security and Industrial Branch
Division of Materials Safety
Region 1
475 Allendale Road
King of Prussia
PA 19406-1415

Q-2

Attn: Mr. Sattar Lodhr, PhD

03020195

Re: License Amendment (45-23031-01)

Dear Mr. Lodhr:

Attached are the resent leak tests results provided to me by Chesapeake branch.

Please let me know if I can provide any additional information.

Sincerely,

A. Kuriatnikov
RSO

RECEIVED
REGION 1
2006 JUN 19 AM 10:31

139013

NMOS/RONI MATERIALS-002

CHANTILLY, VA ■ RICHMOND, VA ■ ROCKVILLE, MD ■ CHARLOTTESVILLE, VA ■ BALTIMORE, MD
BLACKSBURG, VA ■ FREDERICKSBURG, VA ■ CHESAPEAKE, VA ■ WOODBRIDGE, VA



North East Technical Services, Inc.

75 Allerton Ct., Suite 4

Westminster, MD 21157

Ph: 410.751.5090 Fax: 410.751.5091

CTI Consultants - Chesapeake

1407 Stephanie Way

Chesapeake, VA 23320

ATTN: Darrell Darrow

Shipping Address: 1407 Stephanie Way

Chesapeake, VA 23320

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3411	Gauge S/N	5155	Leak Test Date	5/30/2006
--------------------	-------------	------------------	-------------	-----------------------	------------------

Source	Reading in microCuries
---------------	-------------------------------

CC-2212	0.000088
---------	----------

CAA-1261	0.00000
----------	---------

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:

Douglas Sims

Date:

6-5-06



#52

North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

CTI Consultants - Chesapeake

*1407 Stephanie Way
Chesapeake, VA 23320
ATTN: Darell Darrow*

*Shipping Address: 1407 Stephanie Way
Chesapeake, VA 23320*

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	5001	Gauge S/N	114402	Leak Test Date	5/30/2006
Source	Reading in microCuries				
1394CM	0.0000556				
NJ04640	0.00000				

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:

Douglas Sims

Date:

6-5-06

Monday, June 05, 2006



#53

*North East Technical Services, Inc.**75 Allerton Ct., Suite 4**Westminster, MD 21157**Ph: 410.751.5090 Fax: 410.751.5091**CTI Consultants - Chesapeake**1407 Stephanie Way**Chesapeake, VA 23320**ATTN: Darrell Darrow**Shipping Address: 1407 Stephanie Way**Chesapeake, VA 23320*

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	<i>5001 EZ</i>	Gauge S/N	<i>114403</i>	Leak Test Date	<i>5/30/2006</i>
Source		Reading in microCuries			
	<i>1396 CM</i>		<i>0.0000648</i>		
	<i>NJ04641</i>		<i>0.00000</i>		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:

Date:

6-5-06

Monday, June 05, 2006



North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

CTI Consultants - Chesapeake

*1407 Stephanie Way
Chesapeake, VA 23320
ATTN: Darell Darrow*

*Shipping Address: 1407 Stephanie Way
Chesapeake, VA 23320*

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3430	Gauge S/N	22632	Leak Test Date	5/23/2006
Source		Reading in microCuries			
75-4388		0.0000538			
47-18476		0.00000			

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:

Douglas Sims

Date:

5-25-06

Thursday, May 25, 2006

This is to acknowledge the receipt of your letter/application dated

6/13/2006, and to inform you that the initial processing which includes an administrative review has been performed.

☒ Admnd. 45-23031-01
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

☐ Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 139013.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.

NRC FORM 532 (RI)
(6-98)

Sincerely,
Licensing Assistance Team Leader