



V i r g i n i a C o m m o n w e a l t h U n i v e r s i t y

Monroe Park Campus

July 14, 2006

Facilities Management

700 West Grace Street, Suite 2200
P.O. Box 842502
Richmond, Virginia 23284-2502

804 828-0769
FAX: 804 828-1288
TDD: 1-800-828-1120
www.vcu.edu/fmd/

U. S. Nuclear Regulatory Commission, Region I
Division of Nuclear Materials Safety
Licensing Assistance Team
Attn: Penny Lanzisera, HP
475 Allendale Road
King of Prussia, PA 19406

MS 16
P-6

Dear Ms. Lanzisera:

RE: NRC Byproduct Materials License 45-00048-17
Docket No. 030-03297
MCN. 138883

With regard to your email to Mary Beth Taormina on July 13, 2006, we are submitting the additional information you requested for our license amendment. We confirm the following conditions:

1. Authorized users will receive the minimal training and experience required by the Radiation Safety Committee.
2. The review of the facilities will ensure that the floor beneath the self-shielded irradiator is adequate to support the weight of the irradiator; the self-shielded irradiator is secured to prevent unauthorized access or removal; and the area where the self-shielded irradiator is located is equipped with an automatically operated fire detection and control system (sprinkler, chemical, or gas) or the location of the area and other controls ensure a low-level radiation risk attributable to fires.
3. Operating and emergency procedures will be developed, implemented, maintained, and distributed and will meet the Criteria in the section entitled, "Radiation Safety program - Operating and Emergency Procedures" in NUREG - 1556, Vol. 5, dated October, 1998. Changes to operating and emergency procedures will be reviewed and approved by the RSC and affected licensee staff will be trained in the procedures before they are implemented.

Should you have any questions or require any additional information, please contact Mary Beth Taormina in our Radiation Safety Section at (804) 828-7097. Thank you again for your prompt assistance with this request.

Sincerely,


Brian J. Ohlinger, PE

Associate Vice President for Facilities Management

Attachments

RECEIVED
REGION 1

13 8883

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION UNIVERSITY
NRC/RGN MATERIALS-002