

U.S. NUCLEAR REGULATORY COMMISSION		Date: 7/18/2006								
TELEPHONE CONVERSATION RECORD		Time: 1432								
Mail Control or Report No(s).	N/A	License No(s). 19-14125-01 Docket No(s). 03007656								
<table style="width: 100%; border: none;"> <tr> <td style="width: 30%; padding: 5px;">Name of Licensee:</td> <td style="padding: 5px;">Environmental Protection Agency, Ft. Meade, Maryland</td> </tr> <tr> <td style="padding: 5px;">Name of Participant(s):</td> <td style="padding: 5px;">Charles "Skip" Weisberg (EPA) Jim Schmidt (NRC)</td> </tr> <tr> <td style="padding: 5px;">Telephone No.</td> <td style="padding: 5px;">410-305-2681</td> </tr> <tr> <td style="padding: 5px;">Subject: (NOTE: This will be used as the Documents Title in ADAMS)</td> <td style="padding: 5px;">Resolution of Leak Test Results</td> </tr> </table>			Name of Licensee:	Environmental Protection Agency, Ft. Meade, Maryland	Name of Participant(s):	Charles "Skip" Weisberg (EPA) Jim Schmidt (NRC)	Telephone No.	410-305-2681	Subject: (NOTE: This will be used as the Documents Title in ADAMS)	Resolution of Leak Test Results
Name of Licensee:	Environmental Protection Agency, Ft. Meade, Maryland									
Name of Participant(s):	Charles "Skip" Weisberg (EPA) Jim Schmidt (NRC)									
Telephone No.	410-305-2681									
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	Resolution of Leak Test Results									
<table style="width: 100%; border: none;"> <tr> <td style="width: 30%; padding: 5px; vertical-align: top;">Summary:</td> <td style="padding: 5px;"> <p>The licensee requested guidance regarding the handling of high leak test results from improperly collected leak tests for three Agilent electron capture cells. The licensee got the high leak test results from smears collected after the Ni-63 foils were inappropriately directly smeared using a Q-Tip. The manufacturer told the licensee that the results by themselves do not indicate a failed Ni-63 foil and instructed the licensee to repeat the smears using the procedure specified in the technical manual to determine if the cells are leaking. The licensee intends to document the event, retake the leak test smears using the correct procedure and report a failed leak test (and take corrective actions) only if the new or subsequent results indicate leaking sources.</p> <p>The NRC staff agreed with the licensee's proposed actions. (Note: this item was discussed with J. Kinneman at the time of the call)</p> </td> </tr> </table>			Summary:	<p>The licensee requested guidance regarding the handling of high leak test results from improperly collected leak tests for three Agilent electron capture cells. The licensee got the high leak test results from smears collected after the Ni-63 foils were inappropriately directly smeared using a Q-Tip. The manufacturer told the licensee that the results by themselves do not indicate a failed Ni-63 foil and instructed the licensee to repeat the smears using the procedure specified in the technical manual to determine if the cells are leaking. The licensee intends to document the event, retake the leak test smears using the correct procedure and report a failed leak test (and take corrective actions) only if the new or subsequent results indicate leaking sources.</p> <p>The NRC staff agreed with the licensee's proposed actions. (Note: this item was discussed with J. Kinneman at the time of the call)</p>						
Summary:	<p>The licensee requested guidance regarding the handling of high leak test results from improperly collected leak tests for three Agilent electron capture cells. The licensee got the high leak test results from smears collected after the Ni-63 foils were inappropriately directly smeared using a Q-Tip. The manufacturer told the licensee that the results by themselves do not indicate a failed Ni-63 foil and instructed the licensee to repeat the smears using the procedure specified in the technical manual to determine if the cells are leaking. The licensee intends to document the event, retake the leak test smears using the correct procedure and report a failed leak test (and take corrective actions) only if the new or subsequent results indicate leaking sources.</p> <p>The NRC staff agreed with the licensee's proposed actions. (Note: this item was discussed with J. Kinneman at the time of the call)</p>									
<table style="width: 100%; border: none;"> <tr> <td style="width: 30%; padding: 5px;">Action Required:</td> <td style="padding: 5px;">Enter this document into ADAMS</td> </tr> </table>			Action Required:	Enter this document into ADAMS						
Action Required:	Enter this document into ADAMS									
<table style="width: 100%; border: none;"> <tr> <td style="width: 30%; padding: 5px;">Document Availability:</td> <td style="padding: 5px;"> <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available </td> </tr> <tr> <td style="padding: 5px;"> <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright </td> <td></td> </tr> <tr> <td style="padding: 5px;"> <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date </td> <td></td> </tr> <tr> <td colspan="2" style="padding: 5px;"> Prepared & SUNSI Review Completed By: J.W.Schmidt / RA / Date: 7/17/2006 </td> </tr> </table>			Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available	<input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright		<input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date		Prepared & SUNSI Review Completed By: J.W.Schmidt / RA / Date: 7/17/2006	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available									
<input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright										
<input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date										
Prepared & SUNSI Review Completed By: J.W.Schmidt / RA / Date: 7/17/2006										