



July 19, 2006

*In alliance with  
The University of Vermont*

Nuclear Materials Safety Section  
Division of Radiation Safety and Safeguards  
United States Nuclear Regulatory Commission Region I  
475 Allendale Road  
King of Prussia, PA 19406

RE: Fletcher Allen Health Care  
License # 44-10187-03  
Mail Control No. 139014

03003289

MS-16  
P-6

This is in response to the questions raised in the phone conversation of July 17, 2006.

1) Gadolinium sources (35.500) were used at Dorset St. The following information was confirmed in a phone conversation with one of the Philips Medical service personnel. When the camera was moved to Tilley Drive by Philips (ADAC), the manufacturer and service provider, the sources were left in place in the camera assembly and moved as a piece. Doug Hoes, our biomedical engineer and Assistant Radiation Safety, was present at all times during this move. The sources are present in the camera at the Tilley Drive site.

2.) As suggested, please remove this request from the current amendment review. The requested information will be provided separately.

3.) Available purchase, waste, and wipe test records demonstrate that only I-125 and Co-57 (which I realize is not currently regulated by NRC) have been used in the laboratory space.

Thank you for your attention to this. If you have questions please contact Marleen M. Moore, M.S., Radiation Safety Officer, at (802) 847-3506 or (802) 847-4845 direct. All correspondence should be copied to Marleen Moore, M.S., Shepardson 2, FAHC.

Sincerely,

A handwritten signature in dark ink, appearing to read "Angeline".

Angeline Marano  
Chief Operating Officer

A handwritten signature in dark ink, appearing to read "Marleen Moore".

Marleen M. Moore, M.S.  
Radiation Safety Officer and Authorized Medical Physicist  
Fletcher Allen Health Care

139014