



**UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION IV
611 RYAN PLAZA DRIVE, SUITE 400
ARLINGTON, TEXAS 76011-4006**

July 24, 2006

Finlay Testing Laboratories, Inc.
ATTN: Gordon Finlay
President & Radiation Safety Officer
707 Kakoi Street
Honolulu, HI 96819

SUBJECT: TERMINATION OF YOUR NRC RADIOACTIVE MATERIALS LICENSE

On May 17, you contacted the U.S. Nuclear Regulatory Commission and indicated that you wished to terminate your NRC radioactive materials license. The NRC staff has reviewed your sealed source leak test results, which documents there was no evidence of leakage of radioactive material from the sealed sources. Based on its review, the staff concluded that all licensable radioactive material has been removed from your facility and the facility is authorized for unrestricted use in accordance with 10 CFR 20.1402, "Radiological Criteria for Unrestricted Use." An environmental assessment for this action is not required, since this action is categorically excluded under 10 CFR 51.22(c)(20)(ii).

Based on these conclusions no further remediation or actions with respect to NRC regulated material is required. Your facility is suitable for unrestricted use and NRC license number 53-27716-01 for your facility at Honolulu, Hawaii, is hereby terminated.

In accordance with 10 CFR 2.390 of the NRC's "Rules of practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

If you have questions or require clarification on any of the information stated above, we encourage you to contact us at 817-860-8132.

Sincerely,

/RA/

Jacqueline D. Cook, Senior Health Physicist
Nuclear Materials Licensing Branch

Docket: 030-35901
License: 53-27716-01
Control: 470994

Enclosure: As stated

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
53-27716-01Docket or Reference Number
030-35901

Amendment No. 05

Finlay Testing Laboratories, Inc.
707 Kakoi Street
Honolulu, HI 96819

In accordance with NRC Form 314 dated May 17, 2006, License No. 53-27716-01 is hereby terminated.



FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date: July 24, 2006

By: _____

Jacqueline D. Cook, Senior Health Physicist
Nuclear Materials Licensing Branch
Region IV
Arlington, Texas 76011