

U.S. NUCLEAR REGULATORY COMMISSION		Date: 7/25/2006
TELEPHONE CONVERSATION RECORD		Time: 10:30 AM
Mail Control or Report No(s).	139131	License No(s). 45-31174-01
		Docket No(s). 030-37309
<p>Name of Licensee: Volkert and Associates, Inc.</p> <p>Name of Participant(s): William D. McDowall II, P.E. Kathy Modes, NRC</p> <p>Telephone No. 804-452-2862</p> <p>Subject: new portable gauge license (NOTE: This will be used as the Documents Title in ADAMS)</p>		
<p>Summary: Please provide your fax number and email address for our records.</p> <p>Please re-submit following the exact wording in Appendix B of NUREG-1556, Volume 1, Revision 1. Your submission of 7/11/2006 had the incorrect leak test frequency of 6 months. The correct one is 12 months. We need you to confirm that when you are requesting the Model 3400 Series - you are requesting to use Models 3430, 3430-M, 3440, 3440-M, 3450 and 3451. You indicated that you would use equivalent training, but did not specifically state equivalent to Appendix D of NUREG-1556, Vol 1 Rev 1 as required by this guidance. Inventory records are to be retained for 5 years, not 3 as written in this application.</p> <p>Since Volkert is located in various states, please provide a copy of Volkert's agreement state license(s).</p>		
<p>Action Required: Wait for additional information</p>		
<p>Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available</p> <p><input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright</p> <p><input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date</p> <p>Prepared & SUNSI Review Completed By: K Modes / RA / Date: 7/25/2006</p>		