

(10-2003)

10 CFR 2.201

## SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Subhash C. Sabharwal, M.D.  
Cardiovascular Diseases  
17000 Hubbard Drive  
Suite 200  
Dearborn, MI 48126

2. NRC/REGIONAL OFFICE

UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION III  
2443 WARRENVILLE ROAD, SUITE 210  
LISLE, IL 60532-4352

REPORT 2006-001

3. DOCKET NUMBER(S)

030-37001

4. LICENSEE NUMBER(S)

21-32588-01

5. DATE(S) OF INSPECTION

June 14, 2006

## LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

\_\_\_\_\_ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- ☐ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

## Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title

Printed Name

Signature

Date

LICENSEE'S

REPRESENTATIVE

NRC INSPECTOR

Robert P. Hays

6/14/2006

(10-2003)  
10 CFR 2.201**Docket File Information  
SAFETY INSPECTION REPORT  
AND COMPLIANCE INSPECTION**

1. LICENSEE <b>Subhash C. Sabharwal, M.D.</b>		2. NRC/REGIONAL OFFICE <b>Region III</b>	
REPORT NUMBER(S) <b>2006-001</b>			
3. DOCKET NUMBER(S) <b>03037001</b>	4. LICENSE NUMBER(S) <b>21-32588-01</b>	5. DATE(S) OF INSPECTION <b>June 14, 2006</b>	
6. INSPECTION PROCEDURES USED <b>87130 (10/24/02)</b>	7. INSPECTION FOCUS AREAS <b>03.01- 03.07</b>		
<b>SUPPLEMENTAL INSPECTION INFORMATION</b>			
1. PROGRAM CODE(S) <b>02201</b>	2. PRIORITY <b>5</b>	3. LICENSEE CONTACT <b>Denise Smith, Office Mgr.</b>	4. TELEPHONE NUMBER <b>313/271-2580</b>
<input checked="" type="checkbox"/> Main Office Inspection		Next Inspection Date: <b>June 2011</b>	
<input type="checkbox"/> Field Office			
<input type="checkbox"/> Temporary Job Site			

**PROGRAM SCOPE**

The licensee was a private practice clinic authorized by the license to use any byproduct material permitted by 10 CFR 35.100, as needed for any uptake, dilution and excretion study permitted by 10 CFR 35.100, and 10 CFR 35.200, as needed (excluding xenon-133) for any imaging and localization study permitted by 10 CFR 35.200, at the location specified on the license. Licensed activities were conducted 2 days per week in the nuclear medicine suite. The nuclear medicine department was staffed with 2 part-time alternating nuclear medicine technologists (NMT) who routinely conducted 2-4 patient procedures on Wednesdays and Saturdays each week. The licensee received unit doses as ordered from a local nuclear pharmacy for cardiac rest and stress studies. All waste was held for decay-in-storage (DIS) or returned to the nuclear pharmacy as a limited quantity shipment.

**Performance Observations**

During the inspection, the licensee's NMT demonstrated/discussed: (1) package check-in procedures and wipe test counting; (2) dosimetry; (3) dose calibrator checks; (4) unit dose handling procedures; (5) security of licensed material; (6) radiation safety program audits; and (7) rad waste procedures.

7588