



ENCORE SPECIALTY, LLC

177 Edison Ave
Edison, NJ 08820
Tel: 732-205-1324
Fax: 732-205-1325

Date July 24, 2006

License No. 29-31053-01

ATTN: Document Control Desk, Washington, D.C. 20555

Copy To: Regional Administrator, Region I

Subject: ——— Reply to Notice of Violation — EA No. 06-171 ———

To Whom It May Concern:

During the NRC inspection conducted on June 14, 2006, one violation of NRC requirements was identified. The reason for the Violation was that Case was close but not locked because it was stored in the safe place and locked from all the sides, we thought it would be ok. The NRC Personal gave us the reasons why its required and we have taken a corrective steps. We have ordered a lock which will keep the case always lock when we are not using it. Now on every week we will inspect the Gauge machine to avoid any future violations. We should receive the locks for the Case by 07/31/06.

Thanks

Nalin R. Shah

Nalin R. Shah

IE07