

U.S. NUCLEAR REGULATORY COMMISSION TELEPHONE CONVERSATION RECORD		Date: 8/1/06 Time:
Mail Control	138885	License No(s). 37-01873-01
or Report No(s).		Docket No(s). 03003011
Name of Licensee:	Conemaugh Valley Memorial Hospital	
Name of Participant(s):	Randy Kutchman, Radiology/Nuc Med Manager	
Telephone No.	814-534-6032	
Subject: <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	Additional information for amendment request to add GliaSite	
Summary:	<p>1) The faxed letter dated July 21, 2006 (received on 8/1/06) includes all of the information requested in my e-mail dated July 14, 2006 except for confirmation that the licensee will report a leaking source to the NRC within five days of the leakage test to the locations specified and provide the information identified in 10 CFR 35.3067.</p> <p>2) As stated in my e-mail to Keith Ostrum on July 27, please confirm that the RSO and all medical physicists participating in GliaSite treatments will complete the vendor training program.</p>	
Action Required:	Await responses to the two issues above, then issue amendment.	
Document Availability: <div style="display: flex; justify-content: space-around; align-items: flex-end;"> <div style="text-align: center;"> <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Immediate Release </div> <div style="text-align: center;"> <input checked="" type="checkbox"/> Non-Sensitive Copyright <input checked="" type="checkbox"/> Normal Release </div> <div style="text-align: center;"> <input type="checkbox"/> Sensitive <input type="checkbox"/> Delay Release Date </div> <div style="text-align: center;"> <input type="checkbox"/> Non-Publicly Available <input type="checkbox"/> Sensitive Copyright </div> </div>		
Prepared & SUNSI Review Completed By: S. Gabriel / RA / Date: 8/1/06		