



August 1, 2006

Sandy Gabriel  
Senior Health Physicist  
Medical Branch; NRC Region 1  
U.S.N.R.C.  
475 Allendale Road  
King of Prussia, PA 19406-1415

K-8  
MS-16

Re: Request for Information  
License # 37-01873-01 03003011  
Mail Control No.: 138885

Dear Ms. Gabriel:

The following are the responses to the questions posed in your request for additional information.

5. f) "Safety Instructions" will provide instructions on how to safely handle contaminated unsealed materials, in addition to the instructions required by 10 CFR 35.410. In addition, we are committed to report any source leakage to the NRC within 5 days of the treatment and provide proper documentation thereof in accordance with 10 CFR 35.3067.
8. We confirm that the Radiation Safety Officer and all the medical physicists involved in the GliaSite treatments will also complete the vendor-training program.

It is hoped that the following responses meet your concerns. If you have any additional questions concerning our initial application or this response, please feel free to contact our medical physics consultant, Keith Ostrom, Associates in Medical Physics, LLC at 1-800-709-4855 Ext. 31 or Randy Kutchman, Operations Manager – Radiology/Nuclear Medicine at 1-814-534-6032.

Sincerely,

Richard M. Sukenik  
Vice-President  
Conemaugh Valley Memorial Hospital

1086 Franklin Street  
Johnstown, PA 15905-4398  
814-534-9000  
www.conemaugh.org

138885  
NMSS/RGNI MATERIALS-002