



**South Jersey Healthcare®**  
Changing Medicine. Changing Lives.

**Corporate Office**

2950 College Drive  
Suite 1E  
Vineland, New Jersey 08360  
(856) 641-8000

July 28, 2006

Region I USNRC  
Division of Materials Licensing  
475 Allendale Road  
King of Prussia PA 19605

Re: License # NRC 29-19311-01  
South Jersey Healthcare - Regional Medical Center

Q-8

MS-12

2006 AUG - 7 PM 12: 17

RECEIVED  
REGION I

Dear Sir or Madam:

Please amend our license to allow the use of Gadolinium - 153 for the purpose of attenuation correction source for a Vantage-equipped gamma camera.

<b>Manufacturer</b>	Isotope Products Laboratories (IPL)	North American Scientific(NAS)
<b>Model Number</b>	NES-8412	MED3601
<b>Registration Number</b>	CA0406S204S	CA0510S121S
<b>Single Source Strength</b>	250 mCi, not to exceed 300 mCi	250 mCi, not to exceed 300 mCi
<b>Maximum Possession Limit</b>	750 mCi	750 mCi
<b>Authorized user</b>	Robert W. Ciphers, M.D. (Current authorized user 35.200 on NRC License No. 29-19311-01)	

Please contact Jonathan Law (609) 652-3409 if there are any questions regarding this amendment request.

Sincerely,

Wayne Schiffner  
Executive Vice President

WCS:pr

139080  
NMSS/RGNI MATERIALS-C02