

GEOTECH ENVIRONMENTAL, INC.

ENVIRONMENTAL CONSULTANTS
125 NORTH ROUTE 73
MAPLE SHADE, NEW JERSEY 08052

TEL. (856) 667-6010 • FAX (856) 667-7892

August 11, 2006

United States Nuclear Regulatory Commission
Washington, D. C. 20555

Attention: Document Control Desk

Re: Reply to a Notice of Violation Resulting from
NRC Inspection No. 03033171/2006001, EA No. 06-168
(License No. 29-28286-02)

Dear Sir/Madam:

Pursuant to the provisions of 10 CFR 2.201, this letter will provide explanations and statements regarding the Notice of Violation (NOV) received from the NRC, dated July 19, 2006, citing two violations of the NRC requirements.

Our responses to each of the two categories of violations are presented below in the same order as in the NOV.

- A. Based on the security requirements outlined in 10 CFR 30.34(i), our RSO has provided a minimum of two independent physical controls, (chains and steel cables with locks) to secure the nuclear gauges from unauthorized removal during transportation of the gauges to and from the job sites. These security measures will also be implemented during temporary job site storage.
- B. According to the section entitled "Radiation Safety Program – Instruments" outlined in NUREG – 1556, Vol. 1, Rev. 1, dated November 2001, we have made arrangements with U.S. Environmental Rental Corporation to have access to a radiation survey meter.

IE07

We trust that the above detailed responses to NOV are satisfactory and will enable us to continue using our gauges in accordance with NRC requirements and regulations. In addition, we would very much welcome a visit from the NRC Inspector to examine our files and records and verify that we are complying with the Commission's regulations and the license conditions.

Yours very truly,

GEOTECH ENVIRONMENTAL, INC.



Murat Arkan, P.E.
Project Engineer
(RSO)



Cetin A. Okcuoglu, P.E.
President

HMA:CAO:r

cc: Mr. John D. Kinneman, Chief (NRC-Region 1)