

<b>U.S. NUCLEAR REGULATORY COMMISSION</b>		Date: 9/19/06
<b>TELEPHONE CONVERSATION RECORD</b>		Time: various
Mail Control 139053 or Report No(s).	License No(s). 37-13831-01	Docket No(s). 03003203
<p>Name of Licensee: Penn State Milton S. Hershey Medical Center</p> <p>Name of Participant(s): Mike Erdman, Associate Health Physicist</p> <p>Telephone No. 717-531-8574</p> <p>Subject: Additional information for amendment request  <small>(NOTE: This will be used as the Documents Title in ADAMS)</small></p>		
<p>Summary: This telephone conversation record represents a series of e-mails and telephone conversations on this date. I informed Mike that the following additional information is required:</p> <ul style="list-style-type: none"> <li>a) Description of the method used to assure that the HDR unit and CT simulator cannot be energized at the same time. For example, will a two-position switch be used to provide power to only one unit at a time?</li> <li>b) HDR shielding calculations, including workload assumptions, to document that existing shielding is sufficient for proposed HDR workload.</li> </ul> <p>In response to a), Mike contacted Varian and determined that they will either install a 2-way switch or wire the control consoles in synonymous fashion, so that operation of either machine electrically locks out the other.</p> <p>Mike agreed to fax written responses to both items by Thursday, September 21. He also stated that the time frame of the HDR project has been delayed to accomodate installation of the CT simulator. He anticipates first HDR clinical use in December or January.</p>		
<p>Action Required: Await response, then issue amendment</p>		
<p>Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available</p> <p><input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright</p> <p><input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date</p> <p>Prepared &amp; SUNSI Review Completed By: S. Gabriel / RA / Date: 09/21/06</p>		