

**From:** Penny Lanzisera  
**To:** glenn.jennings@gmail.com  
**Date:** Thu, May 3, 2007 10:24 AM  
**Subject:** Request for Additional Information for St. Clare's-Denville Amendment

St. Clare's Hospital  
License No. 29-10245-02  
Docket No. 030-09606  
Mail Control No. 140236

Mr. Jennings, it is unclear whether the following spot-checks are performed in accordance with 10 CFR 35.643:

1. Clock (date and time) in the unit's computer; and
2. Decayed source activity in the unit's computer.

Please describe how this information is confirmed on each day of device use.

You may fax the information to my attention at 610-337-5269. Please reference Mail Control No. 140236 in your response.

Penny Lanzisera  
US NRC, Region 1

**Mail Envelope Properties** (4639F0A3.D9C : 17 : 55940)

**Subject:** Request for Additional Information for St. Clare's-Denville Amendment  
**Creation Date** Thu, May 3, 2007 10:24 AM  
**From:** Penny Lanzisera

**Created By:** PAN@nrc.gov

**Recipients**  
gmail.com  
glenn.jennings (glenn.jennings@gmail.com)

**Post Office**

**Route**  
gmail.com

<b>Files</b>	<b>Size</b>	<b>Date &amp; Time</b>
MESSAGE	1208	Thursday, May 3, 2007 10:24 AM

**Options**  
**Expiration Date:** None  
**Priority:** Standard  
**ReplyRequested:** No  
**Return Notification:** None

**Concealed Subject:** No  
**Security:** Standard