

U. S. ATOMIC ENERGY COMMISSION
BYPRODUCT MATERIAL LICENSE
Supplementary SheetPage 1 of 1 PagesLicense Number 47-260-2
(H64)

Amendment No. 10

Union Carbide Olefins Company
Special Instrumentation Division
437 MacCorkle Avenue
South Charleston, West Virginia

Attention: H. T. Sessions

In response to letter dated May 15, 1962 from H. T. Sessions, License No. 47-260-2
is amended as follows:

Item 4 is amended to extend the expiration date from August 31, 1962 to August 31, 1964.

The symbol below the license number is changed from (H62) to (H64).

Item 8.D. is amended to increase the amount of Cesium 137 which the licensee may
possess at any one time from 2 sources of 500 millicuries each, total 1 curie, to
2 sources of 500 millicuries each and 2 sources of 1 curie each, total 3 curies.

Condition 14. C. is amended to read:

14. C. If the test reveals the presence of 0.005 microcurie or more of removable
contamination, the licensee shall immediately withdraw the sealed source
from use and shall cause it to be decontaminated and repaired or to be
disposed of in accordance with Commission regulations. A report shall be
filed within five days of the test with the Director, Division of
Licensing and Regulation, U. S. Atomic Energy Commission, Washington 25,
D. C., describing the equipment involved, the test results and the corrective
action taken. A copy of such report shall also be sent to the Director,
Region II, Division of Compliance, U. S. Atomic Energy Commission, 50 Seventh
Street, Northeast, Atlanta 23, Georgia.

DUPLICATED
FOR DIV. OF COMPLIANCE

For the U. S. Atomic Energy Commission

Original Signed by
Robert E. Brinkman

by _____ Isotopes Branch

Division of Licensing and Regulation
Washington 25, D. C.

MAY 31 1962

Date

REB 5/31/62

A-14