



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

July 19, 2007

Docket No. 03021288
Control No. 140647

License No. 29-20876-01

Edward A. Christman, Ph.D., CHP
Radiation Safety Officer
University of Medicine and Dentistry of New Jersey
School of Osteopathic Medicine
40 East Laurel Road
Suite 1040
Stratford, NJ 08084

SUBJECT: UNIVERSITY OF MEDICINE AND DENTISTRY OF NEW JERSEY, LICENSE
AMENDMENT, CONTROL NO. 140647

Dear Dr. Christman:

This refers to your license amendment request. Enclosed with this letter is the amended license.

Please review the enclosed document carefully and be sure that you understand and fully implement all the conditions incorporated into the amended license. If there are any errors or questions, please notify the U.S. Nuclear Regulatory Commission, Region I Office, Licensing Assistance Team, (610) 337-5239, so that we can provide appropriate corrections and answers.

An environmental assessment for this action is not required, since this action is categorically excluded under 10 CFR 51.22(c)(14).

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Academic, and Industrial Uses of Nuclear Material**; then **Regulations, Guidance, and Communications**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-888-293-6498. The GPO is open from 7:00 a.m. to 8:00 p.m. EST, Monday through Friday (except Federal holidays).

Thank you for your cooperation.

Sincerely,

Original signed by Stephen Hammann

Stephen Hammann
Health Physicist
Commercial and R&D Branch
Division of Nuclear Materials Safety

Enclosure:
Amendment No. 11

E. Christman 2
University of Medicine and Dentistry of New Jersey

DOCUMENT NAME: C:\FileNet\ML072050092.wpd

SUNSI Review Complete: SHammann

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	<input checked="" type="checkbox"/> N	DNMS/RI	<input type="checkbox"/>	DNMS/RI	<input type="checkbox"/>		<input type="checkbox"/>
NAME	SHammann/ <i>STH2</i> /							
DATE	07/19/07							

OFFICIAL RECORD COPY

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p>Licensee</p> <p>1. University of Medicine and Dentistry of New Jersey School of Osteopathic Medicine</p> <p>2. 40 East Laurel Road, #1040 Stratford, New Jersey 08084</p>	<p>In accordance with the application dated June 4, 2007,</p> <p>3. License number 29-20876-01 is amended in its entirety to read as follows:</p> <p>4. Expiration date April 30, 2016</p> <p>5. Docket No. 030-21288</p>
--	---

6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Hydrogen 3	A. Any	A. 500 millicuries
B. Carbon 14	B. Any	B. 20 millicuries
C. Phosphorus 32	C. Any	C. 200 millicuries
D. Phosphorus 33	D. Any	D. 200 millicuries
E. Sulfur 35	E. Any	E. 120 millicuries
F. Calcium 45	F. Any	F. 2 millicuries
G. Chromium 51	G. Any	G. 50 millicuries
H. Iron 59	H. Any	H. 1 millicurie
I. Copper 64	I. Any	I. 10 millicuries
J. Iodine 125	J. Any	J. 50 millicuries
K. Iodine 131	K. Any	K. 10 millicuries

9. Authorized use:

A. Through K. Research and development as defined in 10 CFR 30.4.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
29-20876-01Docket or Reference Number
030-21288

Amendment No. 11

CONDITIONS

10. Licensed material may be used or stored only at the licensee's Science Center, 2 Medical Center Drive, Stratford, New Jersey.
11. A. Licensed material listed in Item 6. shall be used by, or under the supervision of; Edward A. Christman, Ph.D. or Venkat Venkataraman, Ph.D. Licensed material listed in items 6.A. through 6. I. shall be used by, or under the supervision of; Michael Anakin, S. Biswas, Ph.D., Sergei Borukhov, Ph.D., Salvatore Caradonna, Ph.D., Rocco Carsia, Ph.D., Katrina Cooper, Ph.D., Ronald Ellis, Ph.D., Gary Goldberg, Ph.D., Michael Henry, Ph.D., C.E. Hock, Ph.D., Hristo B. Houbaviy, Ph.D., Peitin Liu, Ph.D., William McAllister, Ph.D., Eric Moss, Ph.D., R. Nagele, Ph.D., Theresa Scholl, Ph.D., T.P. Stein, Ph.D., Randy Strich, Ph.D., Dmitry Teminkov, Ph.D. or K. Yin, Ph.D.

B. The Radiation Safety Officer for this license is Edward A. Christman, Ph.D.
12. The licensee shall not use licensed material in or on human beings.
13. The licensee shall not use licensed material in field applications where it is released except as provided otherwise by specific condition of this license.
14. Experimental animals, or the products from experimental animals, that have been administered licensed materials shall not be used for human consumption.
15. The licensee is authorized to hold byproduct material with a physical half-life of less than 120 days for decay-in-storage before disposal without regard to its radioactivity if it:
 - A. Monitors byproduct material at the surface before disposal and determines that its radioactivity cannot be distinguished from the background radiation level with an appropriate radiation detection survey meter set on its most sensitive scale and with no interposed shielding; and
 - B. Removes or obliterates all radiation labels, except for radiation labels on materials that are within containers and that will be managed as biomedical waste after they have been released from the licensee; and
 - C. Maintains records of the disposal of licensed materials for 3 years. The record must include the date of disposal, the survey instrument used, the background radiation level, the radiation level measured at the surface of each waste container, and the name of the individual who performed the disposal.
16. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
29-20876-01Docket or Reference Number
030-21288

Amendment No. 11

17. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

- A. Application dated December 23, 2005 [ML060090480]
- B. Letter dated March 1, 2006 [ML060760024]



For the U.S. Nuclear Regulatory Commission

Date July 19, 2007

By

Original signed by Stephen Hammann

Stephen Hammann
Commercial and R&D Branch
Division of Nuclear Materials Safety
Region I
King of Prussia, Pennsylvania 19406