

**Reactor Operating Events  
Event Notification Report****~~\*\*\* Not For Public Distribution \*\*\*~~  
Event Notification Report**

U.S. Nuclear Regulatory Commission - Operations Center  
Event Reports For EN No (36997)  
(May include 2-day hold)

**36997**

TOP

General Information or Other	Event Number: 36997
Rep Org: NORTH CAROLINA DIV OF RAD PROTECT Licensee: HUMBOLDT Region: 2 City: RALEIGH State: NC County: License #: 092-0750-1 Agreement: Y Docket: NRC Notified By: ACCATTATO HQ OPS Officer: CHAUNCEY GOULD	Notification Date: 05/12/2000 Notification Time: 16:30 [ET] Event Date: 05/12/2000 Event Time: 09:15 [EDT] Last Update Date: 05/12/2000
Emergency Class: NON EMERGENCY 10 CFR Section: 20.2201(a)(1)(i) - LOST/STOLEN LNM>1000X	Person (Organization): BRIAN BONSER (R2 ) DON COOL (NMSS ) BRENT CLAYTON (R3 )

**Event Text**

HUMBOLDT NOTIFIED THE NORTH CAROLINA DIVISION OF RADIATION PROTECTION OF A MISSING MOISTURE DENSITY GAUGE.

A HUMBOLDT MOISTURE DENSITY GAUGE, MODEL #5001C, SERIAL #1966, WAS SHIPPED FROM THE MANUFACTURER IN RALEIGH, NC, TO COPLAY, PA, VIA THEIR CARRIER, ROADWAY EXPRESS. AFTER FIVE DAYS THE CUSTOMER NOTIFIED HUMBOLDT THAT THEY HAD NOT RECEIVED THE GAUGE. HUMBOLDT THEN CALLED ROADWAY AND FOUND OUT THAT THE GAUGE WAS MISSING IN TRANSIT. A TRACE WAS PUT OUT BY ROADWAY FOR THE GAUGE AND THEY BELIEVE IT MAY HAVE BEEN TRANSFERRED TO ANOTHER TRAILER IN WINSTON-SALEM, NC WHICH WAS SCHEDULED TO GO TO INDIANAPOLIS, IN. THE SUSPECTED TRAILER HAS BEEN LOCATED IN INDIANA, AND THEY WILL TAKE IT BACK TO THEIR HUB IN INDIANAPOLIS TO UNLOAD IT TO VERIFY IF THE GAUGE IS THERE. THE GAUGE CONTAINED 10 MILLICURIES OF CESIUM-137 AND 40 MILLICURIES AMERICIUM/BERYLLIUM. THEY ARE ALSO CHECKING OTHER TRAILERS THAT LEFT FROM THE WINSTON/SALEM, NC HUB.

\* \* \* UPDATE ON 5/12/00 @ 2244 BY ACCATTATO TO GOULD \* \* \*

THE GAUGE WAS RECOVERED IN INDIANAPOLIS AND HAS BEEN REDIRECTED TO THE ORIGINAL CUSTOMER.

Google

Google Search

Search Internet Search www.nrc.gov

B-13

**Information Security Reminder**

This page on NRC's Internal Web site may contain sensitive information.  
Check with the information owner before distributing outside the agency.

Last Updated on 3/26/2004 by NRRWebServices