

(10-2003)
10 CFR 2.201

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

*Cardiology Consultants
560 Mitchell, Suite 500
Petersburg, MI 49770*
REPORT *07-01*

2. NRC/REGIONAL OFFICE

*U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4351*

3. DOCKET NUMBER(S)

030-29838

4. LICENSEE NUMBER(S)

21-24901-01

5. DATE(S) OF INSPECTION

8/7/07

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



1. Based on the inspection findings, no violations were identified.



2. Previous violation(s) closed.



3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):



4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title

Printed Name

Signature

Date

LICENSEE'S
REPRESENTATIVE

NRC INSPECTOR

Kenneth J. Lambert

*Kenneth J. Lambert**8/7/07*

	Initial		Announced	<input checked="" type="checkbox"/>	Unannounced	<input checked="" type="checkbox"/>	Routine		Special	
--	---------	--	-----------	-------------------------------------	-------------	-------------------------------------	---------	--	---------	--

NRC FORM 591M PART 3
(10-2003) 10 CFR 2.201

Docket File Information
**SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

U.S. NUCLEAR REGULATORY COMMISSION

1. LICENSEE Cardiology Consultants, P.C. 560 Mitchell, Suite 500 Petosky, MI REPORT NOS 07-01	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III, 2443 Warrenville Road Suite 210 Lisle Illinois, 60532-4352
--	---

3. DOCKET NUMBER(S) 030-29838	4. LICENSE NUMBER(S) 21-24901-01	5. DATE(S) OF INSPECTION 08/07/2007
6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS 03.01-.08	8. INSPECTOR K. Lambert

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02201	2. PRIORITY 5	3. LICENSEE CONTACT James Botti, M.S., RSO	4. TELEPHONE NUMBER 734-662-3197
------------------------------------	-------------------------	--	--

☒ Main Office Inspection

Field Office _____

Temporary Job Site _____

Next Inspection Date: 08/2012

PROGRAM SCOPE

This was a routine unannounced inspection at a small cardiology practice which is authorized for 35.200 materials. The practice conducts 3-5 cardiac stress tests per day using Tc-99m Cardiolite. The facility is staffed Monday thru Friday, with one ~~one~~ full time technologist. While the licensee is authorized to use radioactive material in Suites 300 and 500, all material is used in Suite 500. Radioactive Material has not been used in Suite 300.

The inspection consisted of interviews with licensee personnel, a review of selected records, a tour of the nuclear medicine department and independent measurements. The inspection included observations of security of byproduct material, use of personnel monitoring, dose calibrator QA checks, areas surveys, and radioactive material packages receipt surveys. The inspector also observed licensee nuclear medicine personnel prepare, assay and administer several unit doses for various imaging procedures. The inspector reviewed dosimetry records, inventory records, package receipt survey results, waste storage records, annual radiation program review records and training records. No violations or issues were identified during the inspection.