

U.S. NUCLEAR REGULATORY COMMISSION		Date: 09/29/2007
TELEPHONE CONVERSATION RECORD		Time: 10:55
Mail Control N/A or Report No(s).	License No(s). 34-32570-01	Docket No(s). 030-37029
<p>Name of Licensee: Baker Inspection Group</p> <p>Name of Participant(s): Joseph Wolfe, RSO Baker Inspection Group Scott Chapin, Thermo-Fisher Scientific</p> <p>Telephone No. Joe Wolfe: 304-675-1150; X413 Scott Chapin: 1-800-875-1578</p> <p>Subject: Niton X-ray Fluorescence Analyzer Held Under General License (NOTE: This will be used as the Documents Title in ADAMS)</p>		
<p>Summary:</p> <p>Mr. Wolfe called to ask if a Niton X-ray Fluorescence Analyzer Held Under General License, containing approximately 6 mCi of Cd-109 was required to be held under a specific license. The device was an XL II Series with a Serial Number of U2123.</p> <p>I reviewed SSDR RI-164-D-102-B which indicated that Field Portable X-ray Fluorescence Analyzers, Model XL-II Series, that indicated that the device could be distributed to persons holding a general license, and the device had to be leak tested every 6-months.</p> <p>Mr. Chapin indicated that a device with that serial number was issued under a general license.</p> <p>Mr. Wolfe confirmed that the device had been routinely leak tested.</p>		
<p>Action Required: No action. Place Telecon in Baker Inspection Group docket file and ADAMS</p>		
<p>Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available</p> <p><input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright</p> <p><input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date</p> <p>Prepared & SUNSI Review Completed By: R. C. Ragland / RA / Date: 9/28/2007</p>		