

10 A
63

E. R. Squibb & Sons, Inc.

P.O. Box 191
New Brunswick, New Jersey 08903
201-545-1300

CHARLES L. KROLL, Sc. D., director
regulatory operations

RECEIVED - FLEAD	
Date	12/30/80
Log	Dec-5 REN
By	SD
Orig. To	
Action Compl.	12/30/80



December 18, 1980

Radioisotope Licensing Branch
Division of Fuel Cycle and Material Safety
U. S. Nuclear Regulatory Commission
Washington, D. C. 20555

Gentlemen:

Reference is made to our specific license 29-00134-05G, pursuant to 10 CFR 32.71, to manufacture or distribute by-product material for use under the general license of 10 CFR 31.11.

We now wish to request renewal of this license which expires on January 31, 1981, for the following products, all of which contain Iodine 125 not exceeding 10 microcuries in each prepackaged unit:

1. Angiotensin I IMMUTOPE Kit, 200 test kits, List 09421
2. Digoxin CLASP RIA Kit, 100 test kit, List G1330
3. ESTRIOI-SQUIBB RADIOIMMUNOASSAY Kit, 100 test kit, List H0810
4. HTSH CLASP RIA Kit, 100 test kit, List G2120
5. THYROSTAT-3 Diagnostic Test Kit, 25 test kit, List 09026
6. THYROSTAT-3 Diagnostic Test Kit, 100 test kit, List 09028
7. T3 CLASP RIA Kit, 100 test kit, List G2105
8. T3-SQUIBB RADIOIMMUNOASSAY Kit, 100 test kit, List H0820
9. T4 CLASP RIA Kit, 100 test kit, List G2110
10. THYROSTAT-FTI Diagnostic Test Kit, 25 test kit, List 09127
11. THYROSTAT-FTI Diagnostic Test Kit, 100 test kit, List 09152
12. THYROSTAT FTI Diagnostic Test Kit, 5 x 100 test kit, List 09172

Applicant	325 404
Stock No.	A570(3G)
Renewal	12/30/80
Signature	Jackson
Number	06242

COPIES SENT TO OFF. OF
INSPECTION AND ENFORCEMENT

A/10

December 18, 1980

The printed labeling components which contain the information required under §32.71(c)(1) and (2), §32.71(d), and §32.71(e) are enclosed for each product. For the purposes of this application, only those portions of the brochures, which accompany the packages, containing information pertinent to §32.71(d) and (e) are being submitted for review; these are highlighted in the attached brochures.

In addition, a check for \$570.00 is enclosed to cover the required renewal fee.

We trust this information is adequate to support renewal of our license 29-00139-05G.

Sincerely,



C. L. Kroll

CLK:js
Enclosure

06242



200 ml

List 09421

¹²⁵I ANGIOTENSIN I BUFFER

Approx. 0.005 μ g Angiotensin I per 200 ml

For *in vitro* Diagnostic Use • See insert

Not for Internal or External Use in Humans or Animals

SODIUM AZIDE: 0.1% ADDED • REFRIGERATE

Total As
Act.: <1.7 microcuries ¹²⁵I of Noon
LOT EXP. EST.
NO.: DATE:

E. R. Squibb & Sons, Inc.

Princeton, NJ 08540

Made in USA

C6673/09421



Angiotensin I IMMUTOPE® Kit

List 09421

For *in vitro* Diagnostic Use • **WARNING: NOT FOR INJECTION**

Not for Internal or External Use in Humans or Animals

Contains sufficient material for 200 tubes

REFRIGERATE • See insert

E. R. Squibb & Sons, Inc., Princeton, NJ 08540

Made in USA

C5149/09421

CONTENTS:

- 1 bottle (200 ml) ¹²⁵I ANGIOTENSIN I BUFFER
<1.7 microcuries per bottle
- 6 bottles (2 ml ea.) ANGIOTENSIN I STANDARD
(1 each of 0, 50, 100, 200, 300, and
500 pg/25 μ l, respectively)
- 1 vial (10 ml) ANGIOTENSIN I ANTISERUM
- 1 bottle (210 tabs.) ANGIOTENSIN I ADSORBENT CHARCOAL
TABLETS
- 1 vial (5 ml) ANGIOTENSIN I PLASMA BUFFER (pH 7.4)
- 1 vial (2 ml) DIMERCAPROL SOLUTION
- 1 vial (330 mg) 8-HYDROXYQUINOLINE SULFATE

**NEW
ASSAY
PROCEDURE**

SQUIBB IN VITRO

Angiotensin I IMMUTOPE® Kit

For Quantitative Measurement of
Plasma Renin Activity by
Radioimmunoassay

For *IN VITRO* Diagnostic Use
For Professional Use Only



06242



100 ml

List G1330

¹²⁵I DIGOXIGENIN

For use with Digoxin CLASP® RIA Kit
For *in vitro* Diagnostic Use • See insert

Not for Internal or External Use in Humans or Animals
Store at 2° to 8° C • SODIUM AZIDE: 0.1%

Total
Act.: < 7 microcuries ¹²⁵I
LOT
NO.:

As
of
EXP.
DATE:

Noon
EST

E. R. Squibb & Sons, Inc., Princeton, NJ 08540
Made in USA

C6648 / G1330



Digoxin CLASP® RIA Kit

List G1330

Squibb Digoxin Radioimmunoassay Kit

For *in vitro* Diagnostic Use • **WARNING: NOT FOR INJECTION**

Not for Internal or External Use in Humans or Animals

Contains sufficient material for 100 tubes

STORE AT 2° to 8° C • See insert

CONTENTS:

1 bottle (100 ml) ¹²⁵I DIGOXIGENIN
< 7 microcuries ¹²⁵I per bottle

6 bottles (1 ml ea.) DIGOXIN STANDARD
(1 each of 0, 0.5, 1.0, 2.0, 3.0
and 5.0 ng per ml)

1 bottle (1 ml) DIGOXIN CONTROL

DIGOXIN STANDARDS and CONTROL
contain human serum

E. R. Squibb & Sons, Inc., Princeton, NJ 08540

Made in USA

C5146 / G1330



J3-431C

Digoxin CLASP[®] RIA Kit

Revised May 1980

For Quantitative Measurement of Serum Digoxin
Levels by Radioimmunoassay

For *IN VITRO* Diagnostic Use
For Professional Use Only

DETERMINATION OF SERUM DIGOXIN LEVELS BY RADIOIMMUNOASSAY

Measurement of body constituents or administered compounds by the technique of radioimmunoassay offers a bioanalytical tool that combines the extreme sensitivity of radioisotope methodology with the extreme specificity of immunological techniques. The procedure requires a specific antibody, a radiolabeled antigen, a pure sample of the antigen to serve as a reference standard, and a means of separation of free antigen from antibody-bound antigen. The procedure follows the basic principle of saturation analysis, where there is competition between labeled and unlabeled antigen for a fixed number of antibody binding sites. As the concentration of unlabeled antigen (the substance actually being measured) increases, less of the added radiolabeled antigen will be bound to the antibody. After a specified period of time, the bound component is separated from the free component. The radioactivity in each bound fraction is then measured. The absolute quantity of unlabeled antigen in the sample being analyzed is determined by comparing the assay results to a standard curve prepared with known amounts of the unlabeled antigen.

In the Digoxin CLASP RIA Kit, antibody to digoxin (produced in New Zealand white rabbits by administration of digoxin coupled to a protein

DIGOXIN CHEMICAL AND BIOLOGICAL PROPERTIES

Chemical Properties: Digoxin is a pure glycoside obtained from the leaves of *Digitalis lanata*. Like all cardiac glycosides, digoxin consists of a steroidal portion or aglycone, and a glycosidic portion, consisting of three digitoxose sugar residues. Digoxin is formed upon partial hydrolysis of the naturally occurring Lanatoside C found in *Digitalis lanata*. It differs from digitoxin by the presence of an extra hydroxyl group at the C-12 position. Because of this structural difference, digoxin shows increased polarity and decreased lipid solubility, resulting in a marked difference in the pharmacokinetics of the two compounds.

The pharmacologic activity of cardiac glycosides is contained exclusively in the steroidal (aglycone or genin) portion of the molecule. The sugars possess no intrinsic activity, but they enhance the pharmacologic activity of the aglycone several times, presumably by increasing solubility or enhancing the ability of the drug to penetrate cell membranes. The pharmacologically active aglycone portion of the digoxin molecule, devoid of the sugar residues, is referred to as digoxigenin, and is the radiolabeled component of the Digoxin CLASP RIA Kit.

Biological Properties: Digoxin is well absorbed from the gastrointestinal tract, with approximately 80 percent of an oral dose being even-



100 ml

List H0810

ESTRIOL PREMIX

¹²⁵I Estriol and Separant Mixture

For *in vitro* Diagnostic Use • See insert

Not for Internal or External Use in Humans or Animals

SODIUM AZIDE: 0.06% added

REFRIGERATE • SHAKE WELL BEFORE USE

Total	As	Noon
Act.: <8.6 microcuries ¹²⁵ I	of	EST

LOT	EXP.
NO.:	DATE:

E. R. Squibb & Sons, Inc., Princeton, NJ 08540

Made in USA

C6723 / H0810



ESTRIOL SQUIBB RADIOIMMUNOASSAY KIT

For *in vitro* Diagnostic Use

WARNING: NOT FOR INJECTION

Not for Internal or External Use in Humans or Animals

Contains sufficient material for 100 tubes

REFRIGERATE • See insert

CONTENTS:

1 bottle (100 ml) ESTRIOL PREMIX

<8.6 microcuries ¹²⁵I per bottle

1 bottle (50 ml) ESTRIOL ANTISERUM

6 vials (1.2 ml ea.) ESTRIOL STANDARD (1 each of 0, 1, 4, 8, 16 and 32 ng per ml)

ESTRIOL STANDARDS contain human serum

E. R. Squibb & Sons, Inc., Princeton, NJ 08540 USA

C5152 / H0810

List H0810

06242

COVER
PMS 266
Purple

drop out white



SQUIBB

E. R. Squibb & Sons, Inc.
Princeton, NJ 08540

Printed in USA

Issued October 1980

J3-486

OK for production
idw 11/24/80
12/17/80

Revised 11/20/80 P 2 of 20 pp.
f

20% tint PMS 266 Purple

drop out white

SQUIBB IN VITRO

drop out white


ESTRIOL-SQUIBB RADIOIMMUNOASSAY KIT

For Quantitative Measurement of
Unconjugated Estriol Levels in
Serum by Radioimmunoassay

For IN VITRO Diagnostic Use

drop out white

Medotopes®



black

P-3-21 20 PT
Exhibit marked
~~11-10~~
12-17-80

PMS 266
Purple

For *IN VITRO* Diagnostic Use

PMS 266 Purple

Table of ContentsPage

Function of the Test	4
Principles of the Test	6
Figure I/Competitive Binding Reaction	6
Reagents	8
Warnings	9
Precautions	10
Chemical Hazard	11
Storage	11
Specimen Collection and Preparation	12
Test Procedure	13
Materials Needed	13
Procedural Precautions	13
Procedure	14
Results	16
Table I/Typical Standard Curve Data	
Calculated in % Bound	18
Figure II/Typical Standard Curve Graph,	
% Bound vs. Estriol Concentration	19
Table II/Typical Standard Curve Data	
Calculated in B/Bo	20
Figure III/Typical Standard Curve Graph,	
B/Bo vs. Estriol Concentration	21

Table of ContentsPage

Limitations of the Test Procedure	22
Expected Values	23
Table III/Normal Range Data	24
Figure IV/Predicted Estriol Levels	26
Figure V/Correlation Data	27
Specific Performance Characteristics	28
Standard Curve	28
Sensitivity	28
Specificity	28
Figure VI/Cross-Reactivity Graph	29
Table IV/Cross-Reactivity Data	30
Recovery Efficiency	31
Table V/Recovery Efficiency	31
Parallelism	31
Figure VII/Parallelism	32
Precision	33
Table VI/Precision Data	33
References	34

J3-486

18.4 of 20/88.
 Estriol Manual
~~4/28/80~~
 12/17/80

FUNCTION

Careful prenatal methods of fetal especially important chronic hypertension intrauterine growth vascular diseases with a past history methods for evaluation and/or exposure unconjugated system is an in fetoplace. at estriol or a succinate of fetal dis


HO 3

12.5 of 20 00.
 Estriol Manual
 1/24/80
 12/17/80

PMS 266
 purple

FUNCTION OF THE TEST

Careful pregnancy management requires dependable methods of fetoplacental monitoring. This monitoring is especially important in such high-risk circumstances as chronic hypertension, preeclampsia, diabetes mellitus, intrauterine-growth retardation, postmaturity, cardiovascular disease, hormonal disorders, and in patients with a past history of problem pregnancies.^{1,2} Direct methods for evaluation of fetal status require hospitalization and/or expensive equipment. The determination of unconjugated or "free" estriol in the maternal blood system is an inexpensive, indirect method for monitoring fetoplacental status, because consistently low levels of estriol or a sudden drop in estriol levels may be indicative of fetal distress.^{1,3}



PMS 266
 purple

- Page
- 22
- 23
- 24
- 26
- 27
- 28
- 28
- 28
- 28
- 29
- 30
- 31
- 31
- 31
- 32
- 33
- 33
- 34

Estriol begins to appear in the maternal blood system in the ninth week of gestation. By the end of the third trimester, the serum level of unconjugated estriol has risen to as high as 20 to 30 ng/ml. Virtually all of the estriol found in the maternal blood system is synthesized in the fetoplacental unit.⁴ The fetal adrenals produce dehydroepiandrosterone sulfate (DHEAS) which is enzymatically hydroxylated at the 16 α position by the fetal liver. This 16 α OH-DHEAS is aromatized to estriol in the placenta and released into the maternal blood system. The half-life of free estriol in the maternal blood system is only 20 to 30 minutes^{5,7} because the maternal liver conjugates estriol to a mixture of sulfates and glucuronides for facile urinary excretion. Because of this short half-life, changes in fetoplacental status are rapidly reflected.

Other assay procedures quantitate the amount of total estriol, free and conjugated, in the blood sample. However, the amount of conjugated estriol has been found to be sensitive to renal excretion.^{8,9} Therefore, these procedures introduce unnecessary, complicating factors into the diagnostic accuracy of the method.

The ESTRIOL-SQUIBB RADIOIMMUNOASSAY KIT measures only free (unconjugated) estriol. The specificity of the antiserum included in the kit obviates the need for any pretreatment of the serum sample, thereby substantially reducing sample handling.


Estriol levels are expressed in nanograms (ng, or 10⁻⁹ g) of estriol per milliliter of serum (ng/ml).

PMS 266 Purple

6 **PRINCIPLES OF THE TEST**
 The technique of radioimmunoassay offers a bioanalytical tool that combines the extreme sensitivity of radioisotope methodology with the extreme specificity of immunological techniques. The procedure requires a specific antibody, a radiolabeled antigen, a pure sample of the antigen to serve as a reference standard, and a means of separation of free antigen from antibody-bound antigen. The procedure follows the basic principle of saturation analysis, where there is competition between labeled and unlabeled antigen for a fixed number of antibody binding sites. As the concentration of unlabeled antigen (the substance actually being measured) increases, less of the radiolabeled antigen will be bound to the antibody. This competitive reaction is depicted in Figure 1. After a specified period of time, the free and bound components of the mixture are separated, and the radioactivity of the bound components is measured. The absolute quantity of unlabeled antigen in the sample being analyzed is determined by comparing the assay result to a standard curve prepared with known amounts of the unlabeled antigen.

PMS 266 Purple

FIGURE 1/The competitive binding reaction.



7 In the ESTRIL-SQUIBB RADIOIMMUNOASSAY KIT, antibody to estril (produced in New Zealand white rabbits by administration of a derivative of estril that was coupled at the 6 position to a protein carrier) is the specific antibody, a purified tyrosyl-containing estril derivative iodinated with ¹²⁵I serves as the radiolabeled antigen, and purified estril is the reference standard. Separation of the free and bound radiolabeled antigen is achieved by a double antibody accelerator system consisting of goat anti-rabbit gamma-globulin and polyethylene glycol. In order to reduce the number of pipetting steps involved in the assay procedure, the radiolabel and the separant have been combined into a single reagent.
 The half-life of ¹²⁵I is 60.2 days. The isotope decays in a complex fashion with emission of x-rays and gamma rays whose radiation energies are 27.4 keV and 35.4 keV, respectively. These energies are well within the detection capability of modern solid crystal gamma counters. There is no beta emission.

10 **Preca**
 Observ
 tive ma
 • There
 const
 radio
 • Wear
 • mate
 • Work
 sorbe
 • Wipe
 the α
 waste
 • Solid
 area i
 been
 until t
 defac
 with r
 • The u
 local
 be dis

 Becau
 these re
 in sc
 icant
 been fo
 with lice

8
pms 266
Purple

REAGENTS

The ESTRIOL-SQUIBB RADIOIMMUNOASSAY KIT contains sufficient material for 100 tubes. Each kit contains the following components:

1. Estriol Antiserum: 50 ml, including less than 34 microliters of estriol antiserum per bottle, bovine serum albumin and sodium chloride (carriers), sodium barbitol and tris(hydroxymethyl)-aminomethane (buffering agents), hydrochloric acid (pH adjustment), and sodium azide (a preservative) in water.
2. Estriol Premix (125 I Estriol and Separant Mixture): 100 ml, with a total radioactivity of less than 8.6 microcuries of 125 I per bottle, in an aqueous matrix containing goat anti-rabbit gamma-globulin (the second antibody), polyethylene glycol 4000 (the accelerator), normal rabbit serum, bovine serum albumin, and sodium chloride (carriers), sodium barbitol and tris(hydroxymethyl)-aminomethane (buffering agents), hydrochloric acid (pH adjustment), and sodium azide (a preservative).
3. Estriol Standards: six vials, 1.2 ml each, containing 0, 1, 4, 8, 16, and 32 ng of estriol per ml in a matrix of human serum including sodium azide (a preservative) and a trace quantity of ethanol (a solubilizer).

THE ESTRIOL PREMIX MUST BE SHAKEN WELL IMMEDIATELY BEFORE USE.

Reagents within each kit are specific to the lot number on the kit; therefore, only reagents from kits with the same lot number may be used interchangeably.

All reagents and samples must be equilibrated to room temperature before use.

266
pms 266
Purple

Warnings

For *In Vitro* Diagnostic Use.

Restricted Device: Federal law restricts the sale, distribution, or use of this device to, by, or on the lawful order of, a health professional.

Note: This radioactive material may be received, acquired, possessed, and used only by physicians, veterinarians in the practice of veterinary medicine, clinical laboratories or hospitals and only for in vitro clinical or laboratory tests not involving internal or external administration of the material, or the radiation therefrom, to human beings or animals. Its receipt, acquisition, possession, use, and transfer are subject to the regulations and a general license of the US Nuclear Regulatory Commission or of a State with which the Commission has entered into an agreement for the exercise of regulatory authority.

E.R. Squibb & Sons, Inc.

In vitro clinical laboratory testing with the ESTRIOL-SQUIBB RADIOIMMUNOASSAY KIT requires only a general license from the US Nuclear Regulatory Commission. The general license is issued to any physician, clinical laboratory, or hospital obtaining a validated registered USNRC Form 483. This form must be submitted in triplicate to the USNRC. The possessor of a general license is subject to the conditions and limitations under 10 CFR 31.11. (A specific license is available from the USNRC for quantities larger than 200 microcuries.)

P. 7 of 20 pp.
Estriol Manual
12/24/80
12/17/80

p. 8 of 20 pg.
Estriol Manual
11/24/80
12/17/80

pms 266
purple

10

Precautions

Observe the following safety rules in handling radioactive material:

- There should be no pipetting by mouth, smoking, or consumption of any food or drink in areas where radioactive materials are permitted.
- Wear gloves when handling radioactive materials.
- Wash hands thoroughly after handling radioactive materials.
- Working areas should be covered with disposable absorbent paper.
- Wipe up all spills quickly and thoroughly and discard the contaminated materials with the radioactive waste.
- Solid waste may be stored in a specifically designated area in a covered metal or plastic container that has been identified with a USNRC radiation caution label until the waste decays to a safe level. Then, after defacing the labeling, the material may be discarded with nonradioactive waste.
- The user should determine whether USNRC and/or local regulations permit the liquid radioactive waste to be discarded through the sanitary sewerage system.

Because Estriol Standards contain human serum, these reagents, and patient samples, should be handled in the same manner as any potentially infective biological material. The human serum in the standards has been found to be nonreactive for HBsAg when tested with licensed reagents.

pms 266
purple

Chemical Hazard

- All of the reagents in this kit contain the commonly used bacteriostatic agent, sodium azide. Disposal of these materials through the laboratory plumbing system could lead to the formation of highly explosive copper and lead azides. In order to reduce the possibility of an explosion, it is recommended that large amounts of water be used to flush excess reagents through the plumbing system.

pms 266
purple

Storage

The kit should be refrigerated upon receipt.

All reagents should be stored in their original containers, with their original closures. Do not transfer these reagents to other containers or change the closures for storage purposes.

SPECIMEN COLLECTION AND PREPARATION

An appropriate quantity of blood should be collected from a peripheral vein using a standard blood collection tube that does not contain an anticoagulant. The patient does not need to be in a fasting state when the blood is withdrawn. In order to avoid the observation of spurious changes in estriol values, the position of the patient during blood collection and the method of blood collection should be standardized.

Serum, not plasma, should be used in performing the test. The variables introduced into the procedure through the use of plasma could make the results unreliable.

It is preferable to separate the serum sample on the same day that the blood is withdrawn from the patient. Serum samples may be stored in a refrigerator for up to 24 hours, and should be frozen if stored for longer periods. Frozen samples should be equilibrated to room temperature and mixed well before use. Samples should not be refrozen.

The use of hemolyzed or lipemic samples is not recommended.

The patient's history should be thoroughly scrutinized to determine if any diagnostic or therapeutic radioisotopes have been administered within the week or two immediately prior to the estriol determination. If the patient has been treated with radioisotopes or if the patient's history is unavailable, the activity of the serum sample should be checked in a scintillation well counter that is set for ^{125}I . If after a few seconds it appears that the count rate is above background, 20 microliters of the serum sample should be accurately counted to determine whether the radioactivity contained in this volume could significantly affect assay results.

pms 266
Purple

TEST PROCEDURE

Materials Needed

In addition to the reagents supplied with the ESTRIOI-SQUIBB RADIOIMMUNOASSAY KIT, the following materials are required:

- 12 x 75 mm plastic test tubes
- Centrifuge capable of 3000 to 4000 rpm
- Well-type gamma scintillation counter (discriminator settings of 20 to 50 keV)
- Refrigerator
- Vortex mixer
- Pipettor and tips for 50 microliters (μl) [and, optionally 500 μl]
- Repeating pipettors capable of delivering 1 ml and 500 μl
- Container for radioactive waste
- Test tube racks
- Controlled-temperature water bath (37°C)

Procedural Precautions

Good laboratory practice dictates that an appropriate number of controls be analyzed in conjunction with each set of samples.

All reagents and samples must be equilibrated to room temperature before use.

pms 266
Purple

Procedure
The technician performs the test on samples, starting with the first sample.

1. Mark the tube for the curve.
2. Add 50 μl of sample.

Tubes

- 1 and 2
- 3 and 4
- 5 and 6
- 7 and 8
- 9 and 10
- 11 and 12
- 13 and 14
- 15 and 16

*If desired, follow:

- a) Mark the tube
- b) Add 50 μl of sample
- c) Add 50 μl of reagent
- d) Treat the sample

p: 9 of 20 of
Estriol Manual
11/24/80
12/17/80

P. 10 of 20 pp.
 Estriol manual
 12/17/80

PMS 266
 Purple

13

14

Procedure

The technique described below is based upon the performance of duplicate analyses for all standards, samples, and controls. Read the entire procedure before starting the test.

1. Mark a series of 12 x 75 mm plastic test tubes with numerals 1 through 14. The first two tubes are used for measuring total counts and the remaining 12 tubes are required for generation of a standard curve. Two additional tubes will be needed for each clinical sample or control to be assayed.*
2. Add 50 μ l of Estriol Standard or 50 μ l of the clinical sample to the respective tubes as follows:

Tubes	Sample to be Added
1 and 2	None
3 and 4	50 μ l of 0 ng/ml Standard
5 and 6	50 μ l of 1 ng/ml Standard
7 and 8	50 μ l of 4 ng/ml Standard
9 and 10	50 μ l of 8 ng/ml Standard
11 and 12	50 μ l of 16 ng/ml Standard
13 and 14	50 μ l of 32 ng/ml Standard
15 and 16	50 μ l of Clinical Sample

*If desired, nonspecific binding can be determined as follows:

- a) Mark two tubes as NSB.
- b) Add 50 μ l of "0" Standard to each tube.
- c) Add 500 μ l of distilled water to each tube.
- d) Treat these two tubes in the same manner as tubes 3 through 14 by completing steps 4 through 10.

3. Add 500 μ l of Estriol Antiserum to tubes 3 through 14.
4. Add 1.0 ml of Estriol Premix to each tube. **THIS REAGENT MUST BE SHAKEN WELL IMMEDIATELY BEFORE USE.**
5. Vortex each tube.
6. Set tubes 1 and 2 aside and incubate the remaining tubes in a controlled-temperature water bath at 37° C for 45 minutes.
7. Centrifuge the incubated tubes for 5 to 10 minutes at 3000 to 4000 rpm.
8. Decant the supernatant by gently inverting each tube once, discarding the liquid into a radioactive waste container. Keeping the tube inverted, touch the rim on an absorbent paper and gently tap the tube.
9. Measure the radioactivity in all tubes in a standard well-type gamma scintillation counter (discriminator settings of 20 to 50 keV) for one minute or for a fixed amount of time that is sufficient to eliminate counting statistics as an important source of variability.
10. Subtract background cpm and record net cpm for each tube.

15

11/24/80
12/17/80

pms 266
purple

16

RESULTS

1. It is possible to plot either % Bound or B/Bo or net cpm versus concentration on linear-linear graph paper. Use the following formula to calculate % Bound for each tube and record the results on a worksheet.

$$\% \text{ Bound} = \frac{\text{PELLET COUNTS (net cpm)} \times 100}{\text{TOTAL COUNTS (net cpm)}}$$

Example:

Background cpm: 150
Tube #3 PELLET COUNTS (gross cpm): 52,398
TOTAL COUNTS (gross cpm): 98,227

$$\begin{aligned} \% \text{ Bound} &= \frac{52,398 - 150}{98,227 - 150} \times 100 \\ &= \frac{52,248}{98,077} \times 100 = 53.3 \end{aligned}$$

Alternatively, use the following formula to calculate B/Bo values for each tube and record the results on a worksheet.

$$\text{B/Bo} = \frac{\text{PELLET COUNTS (net cpm)}}{\text{Avg. PELLET COUNTS of "0" Std. (net cpm)}}$$

Example:

Background cpm: 150
Patient #1 PELLET COUNTS (gross cpm): 25,150
"0" Std. - Avg. PELLET COUNTS
(gross cpm): 50,150

$$\begin{aligned} \text{B/Bo} &= \frac{25,150 - 150}{50,150 - 150} \\ &= \frac{25,000}{50,000} = 0.50 \end{aligned}$$

17

2. Prepare a standard curve by plotting either % Bound or B/Bo or net cpm against the concentrations of the standards. A typical set of data calculated in terms of % Bound is given in Table I and graphed in Figure II. The same set of data calculated in terms of B/Bo is listed in Table II and plotted in Figure III. These curves are provided for guidance only and **should not** be used in calculating the estril levels in the clinical samples.
3. Determine the concentration (ng/ml) of estril in each sample by referring to the standard curve prepared earlier. The observed % Bound or B/Bo or net cpm for each sample will correspond to a specific estril concentration. Samples with estril concentrations of greater than 32 ng/ml should be accurately diluted with "0" standard and reassayed, if an exact value is required.

18

TABLE I/Typical Terms of % Bound

TOTAL COUNT	
Concentration	
0	
0	
1	
1	
4	
4	
8	
8	
16	
16	
32	
32	

18


TABLE I/Typical Standard Curve Data Calculated in Terms of % Bound

TOTAL COUNT		94,683 net cpm		
Standard Concentration (ng/dl)	Pellet Counts (net cpm)	% Bound	Avg. % Bound	
0	47,080	49.7	49.2	
0	46,169	48.8		
1	42,694	45.1	44.6	
1	41,739	44.1		
4	34,254	36.2	36.1	
4	34,039	36.0		
8	28,752	30.4	30.4	
8	28,770	30.4		
16	23,900	25.2	25.2	
16	23,848	25.2		
32	19,781	20.9	21.0	
32	19,894	21.0		

pms 266
 purple

19

FIGURE III/A graph of % Bound vs. Estril Concentration for a typical standard curve obtained with the ESTRIL-SQUIBB RADIOIMMUNOASSAY KIT. As indicated, a patient sample having a mean per-cent bound of 27.5 would have an estril concentra-tion of 11.4 ng/ml.



graph pms 266
 balance of
 color black




TABLE II/Typical Standard Curve Data Calculated in Terms of B/Bo

TOTAL COUNT		94,683 net cpm		
Bo Counts (mean avg.)		46,624 cpm		
Standard Concentration (ng/ml)	Pellet Counts (net cpm)	B/Bo	Avg. B/Bo	
0	47,080	1.01		
0	46,169	0.99	1.00	
1	42,694	0.91		
1	41,739	0.89	0.90	
4	34,254	0.72		
4	34,039	0.72	0.72	
8	28,752	0.60		
8	28,770	0.60	0.60	
16	23,900	0.49		
16	23,848	0.48	0.48	
32	19,781	0.39		
32	19,894	0.40	0.39	

PMS 266
purple

FIGURE III/A graph of B/Bo vs. Estriol Concentration for a typical standard curve obtained with the ESTRIOL-SQUIBB RADIOIMMUNOASSAY KIT. As indicated, a patient sample having a mean B/Bo of 0.55 would have an estriol concentration of 10.4 ng/ml.



P.13 of 20
Estriol manual
11/24/80
12/17/80

graph PMS 266
purple
balance of
copy black



LIT
Box
unc
a t:
fex
sev
lew
one
log
a s
les:
in c
ro
gi
wet
wet
dist
in v
feta
estr
nate
pati
ced
ultra
firm

LIMITATIONS OF THE TEST PROCEDURE

Because the ranges for normal and abnormal levels of unconjugated estriol overlap, it is important to establish a trend for each patient through periodic sampling. The frequency of sampling will depend upon the type and severity of the particular problem. Unconjugated estriol levels have been observed to drop precipitously from one measurement to the next without apparent pathologic cause.¹ One group of investigators² suggests that a single determination must be at least 40 to 45 percent less than the mean of the previous three determinations in order to be more than 99 percent confident that the drop is indicative of fetal distress. Observed unconjugated estriol values of less than 3 ng/ml in the last 12 weeks of pregnancy and less than 4 ng/ml in the last six weeks of pregnancy are highly correlated with fetal distress.³ Because there have been numerous cases^{2,11-14} in which normal estriol levels were not indicative of fetal distress and other examples^{14,15} of women with low estriol levels who gave birth to vigorous, healthy neonates, medical intervention should be based upon the patient's entire clinical picture. Other diagnostic procedures such as fetal heart monitoring, amniocentesis, ultrasonography and stress testing may be used to confirm estriol results.

PMS 266
Purple

EXPECTED VALUES

A total of 119 serum samples from women in the ninth to forty-second week of gestation were assayed. The results are summarized in Table III. Eight of the nine samples from women in the ninth to thirteenth week of gestation were found to contain less than 1 ng of estriol per ml. The individual data points were utilized to compute 95 percent confidence intervals for predicted estriol levels. Because the highest standard included in the ESTRIOI-SQUIBB RADIOIMMUNOASSAY KIT contains 32 ng of estriol per ml, any sample having a larger value is reported as being ">32.00". For the purpose of this analysis, these samples were assigned a value of 32. Therefore, those predictions, which are listed in Table III and graphically portrayed in Figure IV, are conservative.

P. 14 of 20 pg.
Estriol manual
1.6.80
12/17/80

TABLE III/Normal Range Data*

Weeks Gestation	Number of Samples	Range (ng/ml)	Mean (ng/ml)	Predicted Levels (ng/ml)	95% Confidence Limits	
					Lower (ng/ml)	Upper (ng/ml)
9	2	<1.0	<1.0	—	—	—
10	2	<1.0	<1.0	—	—	—
12	2	<1.0-2.3	1.6	—	—	—
13	3	<1.0	<1.0	—	—	—
16	1	5.0	5.0	1.9	0.8	4.4
20	3	3.1-4.4	3.8	2.7	1.2	6.2
21	1	2.9	2.9	3.0	1.3	0.8
22	4	2.9-4.4	3.7	3.3	1.4	7.4
23	1	2.4	2.4	3.6	1.6	8.1
24	3	2.2-4.4	3.0	3.9	1.7	8.9
25	3	2.2-5.9	3.7	4.3	1.9	9.7
26	3	2.8-5.9	4.7	4.7	2.1	10.6
27	1	4.5	4.5	5.1	2.3	11.6
28	2	4.5-4.6	4.6	5.6	2.5	12.6
29	2	4.6-5.4	5.0	6.1	2.7	13.8
30	5	3.8-8.9	6.3	6.7	3.0	15.1
31	2	4.9-6.0	5.4	7.3	3.3	16.5
32	2	12.1-13.5	12.8	8.0	3.6	18.0
34	9	4.4-13.2	7.9	9.6	4.3	21.6
35	15	7.3-17.8	10.5	10.5	4.7	23.6
36	8	6.7-25.0	13.2	11.5	5.1	25.9
37	7(1)*	9.4->32.0	19.2	12.6	5.6	28.3
38	9	9.7-30.6	18.5	13.8	6.1	31.0
39	11(1)*	11.8->32.0	18.8	15.1	6.7	33.9
40	8(1)*	7.1->32.0	16.6	16.5	7.3	37.2
41	7	7.2-27.1	16.1	18.0	8.0	40.7
42	3	7.8-25.5	17.6	19.7	8.7	44.6

*Samples containing less than 1 ng or greater than 32 ng of estriol per ml are reported as "<1.0" or ">32.0", respectively. The limit values were used to calculate the results appearing in this table.

* Number of samples with a value of >32.0.


P. 13 of 20 88.
Estriol Manual
11/24/80
12/17/80

Squibb PDD-RHWJ3-485-10/3/80-AA's 11/14/80
9 on 10 Hel. Lite with 11 Hel. Bold x 18 picas
Disc 0022-Hel on 5-8

P. 16 of 20 pp.
Estriol manual
12/24/80
12/17/80

26


FIGURE IV/Expected range of estriol levels as measured by the ESTRIOL-SQUIBB RADIOIMMUNOASSAY KIT as a function of time of gestation. The mean and the 55% confidence limits are shown. (see the footnotes in Table III)



Benday

Graphs PMS 266 Purple
balance of copy black

FIGURE V/Regression analysis of 110 samples analyzed by the ESTRIOL-SQUIBB RADIOIMMUNOASSAY KIT and by the NEN method. The limit values of 1.0, 32.0, and 36.0 were used to calculate the results appearing in this figure.



In another study the sera of 110 pregnant women were analyzed by the ESTRIOL-SQUIBB RADIOIMMUNOASSAY KIT and by the RIANEN™ Estriol [¹²⁵I] Radioimmunoassay Kit sold by New England Nuclear. The regression analysis of these data is shown in Figure V.

	Estriol-SQUIBB	NEN
Range (ng/ml)	1.0 - 32.0	1.0 - 36.0
Mean (ng/ml)	10.18	10.95
Correlation Coefficient		0.94
Intercept		-0.64
Slope		1.00

Graphs on board

28

SPECIFIC PERFORMANCE CHARACTERISTICS

Standard Curve

Typical standard curves are shown in Figures II and III (see pages 19 and 21).

Sensitivity

The lowest estriol-containing standard supplied with the kit contains 1 ng/ml.


Specificity

The cross-reactivity of the antiserum with various steroids is depicted in Figure VI and tabulated in Table IV.

The relatively large cross-reactivity of the antiserum with 16-epiestriol and estetrol will be of little practical consequence because serum levels of these steroids are very low.^{16,17}

pms 266 purple

FIGURE VI/Cross-reactivity of various steroids in the ESTRIOL-SQUIBB RADIOIMMUNOASSAY KIT.



x—x Aldosterone, Cholesterol, Corticosterone, Cortisol, Cortisone, 11-Deoxycortisol, Dehydroepiandrosterone, Estradiol, 17α-Hydroxyprogesterone, Progesterone, and Testosterone

29

pms 266 purple

graph pms 266 purple
balance of copies
black

30

TABL
the E

Steroid

Aldos

Chole

Cortic

Cortic

11-De

Dehydro

16-Ep

Estetr

Estrac

Estric

Estric

Estric

16α+

17α+

Proge

Testo

*Pe

of

**Pe

p. 18 of 20 pg.
 Estriol manual
 1/24/82
 12/17/80

And all purple

TABLE IV/Cross-Reactivity of Various Steroids with the ESTRIOL-SQUIBB RADIOIMMUNOASSAY KIT

Steroid	Percent Cross-Reactivity at a Concentration of 10 µg/ml
Aldosterone	< 0.01
Cholesterol	< 0.01
Corticosterone	< 0.01
Cortisol	< 0.01
Cortisone	< 0.01
11-Deoxycortisol	< 0.01
Dehydroepiandrosterone	< 0.01
16-Epiestriol	3.33*
Estetrol	13.73*
Estradiol	< 0.01
Estriol-3-glucuronide	0.18
Estriol-16-glucuronide	0.02
Estriol-3-sulfate	0.40
Estrone	0.01
16α-Hydroxyestrone	0.10**
17α-Hydroxyprogesterone	< 0.01
Progesterone	< 0.01
Testosterone	< 0.01

*Percent cross-reactivity at a concentration of 0.1 µg/ml

**Percent cross-reactivity at a concentration of 1 µg/ml

Recovery Efficiency

An important criterion of assay validity is the recovery of exogenous estriol from serum samples. Various amounts of exogenous estriol were added to a serum pool composed of samples taken from normal pregnant women having an endogenous estriol content of 2.49 ng/ml. Each sample was then assayed a total of 12 times over three separate runs. The results, which are presented in Table V, demonstrate quantitative recovery of the exogenous estriol.

TABLE V/Recovery Efficiency

Endogenous Concentration (ng/ml)	Amount Added (ng/ml)	Total Expected (ng/ml)	Total Found (ng/ml)	Percent Recovery*
2.49	2.50	4.95	5.38	108.7
2.49	5.00	7.49	7.45	99.5
2.49	10.00	12.49	12.05	96.5
2.49	20.00	22.49	23.03	102.4

$$\text{*Percent Recovery} = \frac{\text{Total Found}}{\text{Total Expected}} \times 100$$


Parallelism

This type of study indicates whether patient samples and standards behave similarly toward the radiolabel and the antiserum, thereby validating the matrix of the standards and the use of the zero standard as a diluent for patient samples. A serum sample having a known analyte concentration is serially diluted with zero standard and the original sample and its dilutions are assayed. A plot of observed dose against expected dose should theoretically yield a 45° line through the origin. Such a study was performed with three separate samples containing high levels of estriol. The results, depicted in Figure VII, demonstrate the validity of the matrix for the standards and the use of the zero standard as a diluent.

P. 19 of 20 pp.
 Estril manual
 11/24/80
 12/17/80

32

FIGURE VII/The parallelism study: A graph of observed dose versus expected dose. The theoretical 45° line is shown.



Estril - SQUIBB			
	Sample 1 (•)	Sample 2 (○)	Sample 3 (▲)
Slope			
Correlation	1.059	0.926	0.966
Intercept	0.22	0.40	0.12
Correlation			
Coefficient	0.999	0.997	0.993

pros 266 purple

Precision

Intra-assay CV data were obtained by performing a single run that consisted of 20 consecutive tubes each of human serum pools containing low, medium and high concentrations of estril. In order to obtain inter-assay CV data, each of these controls was also assayed in duplicate in 12 runs done over a period of two months. The observed results yielded the reproducibility data shown in Table VI.

TABLE VI/Precision Data

Parameter	Serum Pool		
	Low	Medium	High
INTRA-ASSAY			
Mean (ng/ml)	5.19	10.06	22.16
Std. Dev. (ng/ml)	0.36	0.89	1.46
Range (ng/ml)	4.58-5.68	8.53-11.65	18.75-23.83
Coefficient of Variation (%)	7.0	8.8	6.6
INTER-ASSAY			
Mean (ng/ml)	5.27	10.53	20.53
Std. Dev. (ng/ml)	0.53	0.90	1.93
Range (ng/ml)	4.34-6.43	8.64-12.08	17.49-25.18
Coefficient of Variation (%)	10.0	8.5	9.4

33

34

R

pms 266
purple

34

REFERENCES

1. Boster JE, Abraham GE: The applications of steroid hormone radioimmunoassays to clinical obstetrics. Obstet Gynecol 46: 489-499, 1975
2. Bashore RA, Westlake JR: Plasma unconjugated estriol values in high-risk pregnancy. Amer J Obstet Gynecol 128: 371-379, 1977
3. Gorwill RH, Sarda IR: Hormonal studies in pregnancy. II. Unconjugated estriol in maternal peripheral vein, cord vein, and cord artery serum at delivery in pregnancies complicated by intrauterine growth retardation. Amer J Obstet Gynecol 127: 17-20, 1977
4. Loriaux DL et al: Estrone sulfate, estrone, estradiol and estriol plasma levels in human pregnancy. J Clin Endocrinol Metab 35: 887-891, 1972
5. Sandberg AA, Slaunwhite WR: Studies on phenolic steroids in human subjects. VII. Metabolic fate of estriol and its glucuronide. J Clin Invest 44: 694-702, 1965
6. Tulchinsky D, Abraham GE: Radioimmunoassay of plasma estriol. J Clin Endocrinol Metab 33: 775-782, 1971
7. Podoba V et al: Radioimmunoassay for estrogens in maternal blood in late pregnancy: values for normal and complicated pregnancy. Amer J Obstet Gynecol 117: 321-330, 1973
8. Timonen S et al: Urinary volume and excretion of oestrogens in late pregnancy. Acta Endocrinol 49: 393-402, 1965

9. Goebelsmann U, Jaffe RB: Oestriol metabolism in pregnant women. Acta Endocrinol 66: 679-693, 1971
10. Katagiri H et al: Estriol in pregnancy. IV. Normal concentrations, diurnal and/or episodic variations, and day-to-day changes to unconjugated and total estriol in late pregnancy plasma. Amer J Obstet Gynecol 124: 272-280, 1976
11. MacLeod SC et al: The value of urinary oestriol measurements during pregnancy. Aust New Zeal J Obstet Gynaecol 7: 25-38, 1967
12. Reid S et al: Urinary oestriol excretion in pregnancies complicated by proteinuria. Aust New Zeal J Obstet Gynaecol 8: 189-196, 1968
13. Kloppe A: The assessment of fetoplacental function by estriol assay. Obstet Gynecol Survey 23: 813-838, 1968
14. Kloppe A: Assessment of fetoplacental function by hormone assay. Amer J Obstet Gynecol 107: 807-827, 1970
15. France JT, Liggins GC: Placental sulfatase deficiency. J Clin Endocrinol Metab 29: 138-141, 1969
16. Adlercruetz H, Luukkainen T: Identification and determination of oestrogens in various biological materials in pregnancy. Ann Clin Res 2: 365-380, 1970
17. Tulchinsky D, et al: Plasma estetrol as an index of fetal well-being. J Clin Endocrinol Metab 40: 560-567, 1975

35

p 20 of 20 in
Estriol manual
11/11/80
12/17/80



50 ml
¹²⁵I HTSH REAGENT

For use with HTSH CLASP® RIA Kit
For *in vitro* Diagnostic Use • See directions
Not for Internal or External Use in Humans or Animals
SODIUM AZIDE: 0.1%
STORE BELOW -10°C • SHAKE WELL BEFORE USE

Total As
Act.: <3 microcuries ¹²⁵I of

Noon
EST

LOT
NO.: EXP.
DATE:

E. R. Squibb & Sons, Inc., Princeton, NJ 08540
Made in USA

List G2120

G6318B / G2120



HTSH CLASP® RIA Kit

Squibb HTSH Radioimmunoassay Kit
For *in vitro* Diagnostic Use • See enclosed directions
Not for Internal or External Use in Humans or Animals
Contains sufficient material for 100 tubes
STORE BELOW -10°C

List G2120

CONTENTS:
1 bottle (50 ml) ¹²⁵I HTSH REAGENT
<3 microcuries/bottle
1 bottle (4 ml) HTSH RIA STANDARD
0 µIU/ml
5 bottles (2 ml ea.) HTSH RIA STANDARD
(1 each of 2.5, 5.0, 10.0, 25.0 and
50.0 µIU/ml)
1 bottle (20 ml) HTSH ANTISERUM

E. R. Squibb & Sons, Inc., Princeton, NJ 08540
Made in USA

M5370A / G2120



J3-445B

Revised June 1979

HTSH CLASP[®] RIA Kit

**For Quantitative Measurement of Serum Human
Thyroid Stimulating Hormone (HTSH) Levels by Radioimmunoassay**

**For *IN VITRO* Diagnostic Use
For Professional Use Only**

06242



50 ml

THYROSTAT-3 LIOTHYRONINE I 125 BUFFER SOLUTION

0.03 microcurie or less per ml • SODIUM AZIDE: 0.1%

FOR LABORATORY USE ONLY with THYROSTAT-3


Adsorbent Tablets • See accompanying directions

For *in vitro* Diagnostic Use • Shake gently before use

Not for Internal or External Use in Humans or Animals

Total	As	Noon
Act. <1.9 microcuries I 125	of	EST
LOT	EXP.	
NO.	DATE	

E. R. Squibb & Sons, Inc.
Princeton, NJ 08540 Made in USA C6033B / 09026



25 TEST PKG.

THYROSTAT-3 DIAGNOSTIC TEST KIT

For *in vitro* Diagnostic Use

For Evaluation of Thyroid Function

FOR LABORATORY USE ONLY

Not for Internal or External

Use in Humans or Animals

REFRIGERATE AT 2° to 8°C

CONTENTS:

1 vial (28 tabs.) THYROSTAT-3 ADSORBENT TABLETS

1 bottle (50 ml) THYROSTAT-3 LIOTHYRONINE I 125

BUFFER SOLUTION <1.9 microcuries per bottle

NOTE: Buffer solution in this carton must

be used with the accompanying Adsorbent Tablets.

See accompanying directions

E. R. Squibb & Sons, Inc.

Princeton, NJ 08540



Made in USA

C6043A/09026

LOT NO.
EXP. DATE

REFRIGERATE AT 2° to 8° C

THYROSTAT-3[®] DIAGNOSTIC TEST KIT



25 TEST PKG.

List 09026

THYROSTAT-3[®] DIAGNOSTIC TEST KIT

For *in vitro* Diagnostic Use

For Evaluation of Thyroid Function

FOR LABORATORY USE ONLY • See accompanying directions

Not for Internal or External Use in Humans or Animals


CONTENTS:

1 vial (28 tabs.) THYROSTAT-3 ADSORBENT TABLETS

1 bottle (50 ml) THYROSTAT-3 LIOTHYRONINE I 125 BUFFER SOLUTION
<1.9 microcuries per bottle

Note: Buffer solution in this carton must
be used with the accompanying Adsorbent Tablets.

REFRIGERATE AT 2° to 8° C



THYROSTAT-3[®] DIAGNOSTIC TEST KIT

06242



200 ml

List 09028

THYROSTAT[®]-3 LIOTHYRONINE I 125 BUFFER SOLUTION

0.03 microcurie or less per ml • SODIUM AZIDE: 0.1%
FOR LABORATORY USE ONLY with THYROSTAT-3

Adsorbent Tablets • See accompanying directions.

For *in vitro* Diagnostic Use • Shake gently before use


Not for Internal or External Use in Humans or Animals

Total	As	Noon
Act. < 7.1 microcuries I 125	of	EST
LOT	EXP.	
NO.	DATE	

E. R. Squibb & Sons, Inc.
Princeton, NJ 08540


Made in USA

C6032B / 09028



SQUIBB
100 TEST PKG.

CAUTION
RADIOACTIVE MATERIAL



List 09028

THYROSTAT[®]-3 DIAGNOSTIC TEST KIT

For *in vitro* Diagnostic Use
For Evaluation of Thyroid Function
FOR LABORATORY USE ONLY
Not for Internal or External Use in Humans or Animals
REFRIGERATE AT 2° to 8°C

CONTENTS:
1 vial (105 tablets) THYROSTAT-3 ADSORBENT TABLETS
1 bottle (200 ml) THYROSTAT-3 LIOTHYRONINE I 125 BUFFER SOLUTION < 7.1 microcuries per bottle
1 vial (1.0 ml) THYROSTAT-3 NORMAL CONTROL SERUM
1 vial (1.0 ml) THYROSTAT-3 HYPER CONTROL SERUM
8 THYROSTAT-3 TOTAL COUNT RESIN TUBES

Note: All reagents in this kit must be used with the accompanying Adsorbent Tablets • See accompanying directions

E. R. Squibb & Sons, Inc.
Princeton, NJ 08540

Made in USA

C6042A/09028

LOT NO.:
EXP. DATE:



J3-358B

THYROSTAT®-3 DIAGNOSTIC TEST KIT

Revised February 1980

For Quantitative Measurement of Serum Liothyronine
(T_3) Uptake for the Evaluation of Thyroid Function

For *IN VITRO* Diagnostic Use
For Professional Use Only

PRINCIPLES OF THE TEST

The currently accepted principles underlying the T_3 (synonymous with liothyronine and triiodothyronine) uptake test are as follows:

The circulating thyroid hormones produced by the thyroid gland are bound to specific plasma proteins known collectively as thyroxine-binding proteins (TBP), and are in equilibrium with a small fraction of the free thyroid hormones circulating in the plasma. A change in the number of unoccupied TBP binding sites will alter the free thyroid hormonal level in the blood, which can be indirectly measured through the use of the T_3 uptake test.

In the T_3 uptake test, a supply of labeled exogenous thyroid hormone (in the Liothyronine I 125 Buffer Solution) is added to the patient's serum together with a secondary binding site (the Thyrostat-3 Adsorbent Tablet). A portion of the liothyronine ^{125}I will become bound to the binding sites of the TBP that are not occupied by the thyroxine whereas some, not bound to the TBP, will become bound to the adsorbent.

When the thyroid gland produces an excess of thyroid hormones, as in hyperthyroidism, the number of unoccupied TBP binding sites is reduced and a greater proportion of the added hormone will become bound to the adsorbent. Conversely, when thyroid hormone production is decreased, as in hypothyroidism, the number of unoccupied TBP binding sites is increased and a greater proportion of the added liothyronine ^{125}I will become bound to the TBP resulting in a decreased uptake by the adsorbent. (In normal pregnancy, although the free thyroid hormone level is normal, there is an increased production of TBP, so that more binding sites are available resulting in the binding of a greater proportion of the added liothyronine ^{125}I to these sites as in hypothyroidism.)

Therefore, the use of the T_3 uptake test provides an estimate of the unoccupied binding sites of the TBP in a given serum sample. This, in turn, gives an indirect estimate of the amount of endogenous circulating thyroid hormone, and therefore an indirect but reliable indication of thyroid function.

In summary, a large T_3 uptake indicates hyperthyroidism, while a small T_3 uptake indicates hypothyroidism (or normal pregnancy).

RATIONALE FOR USE

The T_3 uptake test represents a significant advance in the search for a simple and reliable test of thyroid function. The test is an *in vitro* procedure which avoids any exposure of the patient to ionizing radiation. Equally important is the fact that the test is diagnostically significant in the presence of unrelated nonthyroidal factors which are known to complicate in-

terpretation of other thyroid function tests. Although other thyroid function tests may be affected for considerable periods of time by the prior administration of most iodine-containing preparations, the T_3 uptake test is not so affected at the normal dose level at which these drugs are used. Anxiety, hypertension, congestive heart failure, or administration of mercurial agents also have no effect on the test.

The technique readily falls within the scope of any hospital or office laboratory with ordinary isotope facilities and is simple, rapid, and inexpensive enough to be used as a general screening test. Moreover, the test is consistently reliable when repeated at frequent intervals.

Note: While the T_3 uptake test is a very useful aid in the evaluation of thyroid function, it should not be used as the sole basis for such an evaluation. In any patient, the clinical state is probably the best indication of thyroid status, and any laboratory test must be interpreted with caution when test results do not agree with clinical evidence.

The Thyrostat-3 test offers further advantages in the performance of the T_3 uptake test. Unlike many of the T_3 uptake procedures employing anion exchange resins, Thyrostat-3 test results are not significantly affected by variations in time or temperature during contact with the Thyrostat-3 Adsorbent Tablet (test results are essentially unchanged at normally encountered room temperatures ranging between 20° and 25° C.).

The use of ^{125}I rather than ^{131}I considerably lengthens the shelf-life of the liothyronine employed in the test because of the longer half-life of ^{125}I and the fact that it emits no beta rays to affect the stability of the liothyronine. Moreover, with ^{125}I labeled material, radiation exposure to the technician is lowered. Radioactivity is well within good counting range of modern equipment, and *in vitro* counting is quite efficient.

The half-life of ^{125}I is 60.2 days. The isotope decays in a complex fashion with emission of x-rays and gamma rays whose radiation energies are 27.4 keV and 35.4 keV, respectively. There is no beta emission.

REAGENTS

The Thyrostat-3 *in vitro* diagnostic test for quantitative measurement of serum liothyronine (T_3) uptake for the evaluation of thyroid function is available in 25- and 100-test kits.

The Thyrostat-3, 25-test kit provides 25 plastic test tubes; 1 vial of Adsorbent Tablets (28 tablets); 1 bottle of Liothyronine I 125 Buffer Solution (50 ml, containing 0.03 μCi I 125 or less per ml with preservatives and buffer); and 1 Control Pak

J3-358B

J3-358B



100 ml

List G2105

¹²⁵I T3 REAGENT

For use with T3 CLASP® RIA Kit
For in vitro Diagnostic Use • See directions
Not for Internal or External Use in Humans or Animals
SODIUM AZIDE: 0.1%
Store at 2° to 8° C • SHAKE WELL BEFORE USE

Total	As	Noon
Act.: <4.3 microcuries	of	EST
LOT	EXP.	
NO:	DATE:	

E. R. Squibb & Sons, Inc., Princeton, NJ 08540
Made in USA

C6303B / G2105



T3 CLASP® RIA Kit

List G2105

Squibb T3 Radioimmunoassay Kit
For in vitro Diagnostic Use • See enclosed directions
Not for Internal or External Use in Humans or Animals
Contains sufficient material for 100 tubes
Store at 2° to 8° C

CONTENTS:
1 bottle (100 ml) ¹²⁵I T3 REAGENT
<4.3 microcuries ¹²⁵I per bottle
6 bottles (1.5 ml ea.) T3 RIA STANDARD
(1 each of 10, 50, 100, 250, 500 and
1000 ng per dl)

E. R. Squibb & Sons, Inc., Princeton, NJ 08540
Made in USA

M5371B / G2105



J3-447B

Revised February 1980

T3 CLASP[®] RIA Kit

For Quantitative Measurement of Serum Liothyronine Levels by Radioimmunoassay

**For *IN VITRO* Diagnostic Use
For Professional Use Only**

DETERMINATION OF SERUM LIOTHYRONINE (T₃) LEVELS BY RADIOIMMUNOASSAY

Measurement of body constituents or administered compounds by the technique of radioimmunoassay offers a bio-analytical tool that combines the extreme sensitivity of radioisotope methodology with the extreme specificity of immunological techniques. The procedure requires a specific antibody, a radiolabeled antigen, a pure sample of the antigen to serve as a reference standard, and a means of separation of free antigen from antibody-bound antigen. The procedure follows the basic principle of saturation analysis, where there is competition between labeled and unlabeled antigen for a fixed number of antibody-binding sites. As the concentration of unlabeled antigen (the substance actually being measured) increases, less of the added radiolabeled antigen will be bound to the antibody. After a specified period of time, the

combined with a normal serum thyroxine concentration is evidence for T₃ thyrotoxicosis, a disease state described by Sterling and co-workers in 1970.¹ Although liothyronine levels are usually below the normal range in hypothyroidism, some overlap with euthyroid levels has been found. Therefore, liothyronine assay results have limited use in the confirmation of hypothyroidism.

Earlier methods for the assay of serum liothyronine concentration (competitive protein binding^{2,3} and gas chromatography⁴) were tedious, time consuming, and required large volumes of serum. The development of specific liothyronine antibodies⁵ and the utilization of liothyronine-binding protein inhibitors^{6,7} led to the development of specific, accurate, and relatively simple RIA procedures for the determination of serum liothyronine levels.

The utilization of solid-phase technology in the T3 CLASP



100 ml

List H0820



¹²⁵I T3 and Separant Mixture

For *in vitro* Diagnostic Use • See insert

Not for Internal or External Use in Humans or Animals

SODIUM AZIDE: 0.1% added

REFRIGERATE • SHAKE WELL BEFORE USE

Total

Act.: < 6 microcuries ¹²⁵I

LOT

NO.:

As

of

EXP.

DATE:

Noon

EST

E. R. Squibb & Sons, Inc., Princeton, NJ 08540

Made in USA

C6747 / H0820



List H0820



For *in vitro* Diagnostic Use • **WARNING: NOT FOR INJECTION**

Not for Internal or External Use in Humans or Animals

Contains sufficient material for 100 tubes

REFRIGERATE • See insert

CONTENTS:

1 bottle (100 ml) T3 PREMIX < 6 microcuries ¹²⁵I per bottle

1 bottle (50 ml) T3 ANTISERUM

6 vials (1.2 ml ea.) T3 STANDARD (1 each of 0, 50, 100, 250, 500 and 1000 ng of triiodothyronine per dl)

T3 STANDARDS contain human serum

Restricted Device: Federal law restricts the sale, distribution, or use of this device to, by, or on the lawful order of, a health professional

E. R. Squibb & Sons, Inc., Princeton, NJ 08540

Made in USA

C5165 / H0820

T3-SQUIBB RADIOIMMUNOASSAY Kit

For Quantitative Measurement of Serum
Triiodothyronine Levels by Radioimmunoassay

For *IN VITRO* Diagnostic Use

FUNCTION OF THE TEST

Measurement of serum levels of triiodothyronine (T_3) is an important adjunct in the determination of thyroid function. It is especially useful in the confirmation of hyperthyroidism and in the detection of T_3 thyrotoxicosis. An elevated serum triiodothyronine level accompanied by an elevated serum thyroxine level is a strong indication of hyperthyroidism. In patients exhibiting clinical hyperthyroidism, an elevated serum triiodothyronine level combined with a normal serum thyroxine concentration is evidence for T_3 thyrotoxicosis, a disease state described by Sterling and co-workers in 1970.¹ Disagreement exists concerning the applicability of triiodothyronine measurement to the detection of hypothyroidism. In all^{2,3} or most⁴ patients with unequivocal clinical hypothyroidism, levels of triiodothyronine are below normal. However, in borderline cases, triiodothyronine levels are sometimes³ or often² normal. Thus, a reduced serum level of triiodothyronine is not as clear an indicator of hypothyroidism as a reduced serum level of thyroxine (T_4) or an elevated level of thyroid-stimulating hormone (TSH).⁴


Chemical and Biological Principles

The basal membrane of the thyroid gland traps inorganic iodide and actively transports it into the gland where it is processed into organic iodide through conversion of the tyrosyl residues of the thyroglobulin to moniodotyrosine and diiodotyrosine. Oxidative coupling of two diiodo residues yields T_4 , whereas oxidative coupling of a moniodo

DEC 17 1980

residue to a diiodo residue yields T_3 . Approximately a two month supply of T_4 is stored within the thyroglobulin. When needed, thyroglobulin is enzymatically hydrolyzed to release these hormones into the circulatory system. However, most of the T_3 in circulation is normally derived from peripheral deiodination of T_4 in the liver.

The release of T_4 and T_3 from the thyroid is markedly influenced by pituitary thyroid-stimulating hormone which in turn is influenced by hypothalamic thyrotropin-releasing hormone (TRH). Normally, increased blood levels of T_4 and T_3 act to decrease the amount of TSH secreted, thereby reducing the production and release of T_4 and T_3 . Decreased blood levels of T_4 and T_3 produce the opposite effect, leading to increased production and secretion of T_4 and T_3 . In this manner a normal circulating thyroid hormone balance is maintained. Most of the circulating T_4 and T_3 in the blood is bound to serum proteins, i.e., thyroxine-binding globulin (TBG), thyroxine-binding prealbumin (TBPA), and albumin. A small fraction of triiodothyronine (~0.3%) is free. This free triiodothyronine is considered to be the metabolically active form in its effect on target tissue. The free fraction of T_3 is influenced not only by the amount of thyroxine-binding proteins in the blood, but also by the rate of peripheral deiodination of T_4 .



Triiodothyronine

DEC 17 1980

PRINCIPLES OF THE TEST

The technique of radioimmunoassay offers a bioanalytical tool that combines the extreme sensitivity of radioisotope methodology with the extreme specificity of immunological techniques. The procedure requires a specific antibody, a radiolabeled antigen, a pure sample of the antigen to serve as a reference standard, and a means of separation of free antigen from antibody-bound antigen. The procedure follows the basic principle of saturation analysis, where there is competition between labeled and unlabeled antigen for a fixed number of antibody binding sites. As the concentration of unlabeled antigen (the substance actually being measured) increases, less of the radiolabeled antigen will be bound to the antibody. This competitive reaction is depicted in Figure 1. After a specified period of time, the free and bound components of the mixture are separated, and the radioactivity of the bound components is measured. The absolute quantity of unlabeled antigen in the sample being analyzed is determined by comparing the assay results to a standard curve prepared with known amounts of the unlabeled antigen.




Figure 1. The competitive binding reaction.

DEC 17 1980

In the T3-SQUIBB RADIOIMMUNOASSAY Kit, antibody to triiodothyronine (produced in New Zealand white rabbits by administration of a derivative of triiodothyronine that was coupled to a protein carrier) is the specific antibody, purified triiodothyronine containing ^{125}I serves as the radiolabeled antigen, and purified triiodothyronine is the reference standard. Specific components in the T3 Premix (see REAGENTS section) promote the release of triiodothyronine from the serum binding proteins. Separation of the free and bound radiolabeled antigen is achieved by a double antibody accelerator system consisting of goat anti-rabbit gamma-globulin and polyethylene glycol. In order to reduce the number of pipetting steps involved in the assay procedure, the radiolabel and the separant have been combined into a single reagent.

The half-life of ^{125}I is 60.2 days. The isotope decays in a complex fashion with emission of x-rays and gamma rays whose radiation energies are 27.4 keV and 35.4 keV, respectively. These energies are well within the detection capability of modern solid crystal gamma counters. There is no beta emission.

Triiodothyronine levels are expressed in nanograms (ng, or 10^{-9} g) of triiodothyronine per deciliter of serum (ng/dl).

REAGENTS

The T3-SQUIBB RADIOIMMUNOASSAY Kit contains sufficient material for 100 tubes. Each kit contains the following components:

1. T3 Antiserum: 50 ml, including less than 5 microliters of T_3 antiserum per bottle, bovine serum albumin and normal rabbit serum (carriers), tris(hydroxymethyl)aminomethane and acetic acid (buffering agents), and sodium azide (a preservative) in water.

DEC 17 1980

2. T3 Premix (^{125}I T3 and Separant Mixture): 100 ml, with a total radioactivity of less 6 microcuries of ^{125}I per bottle; in an aqueous matrix containing goat anti-rabbit gamma-globulin (the second antibody); polyethylene glycol 4000 (the accelerator); disodium edetate dihydrate, sodium salicylate, and ammonium 8-anilino-1-naphthalene sulfonate (inhibitors of TBP binding); bovine serum albumin (a protein carrier); tris(hydroxymethyl)aminomethane and acetic acid (buffering agents); and sodium azide (a preservative).
3. T3 Standards: six vials, 1.2 ml each, containing 0, 50, 100, 250, 500, and 1,000 ng of triiodothyronine per dl in a matrix of human serum including sodium azide (a preservative) and trace amounts of ethanol and ammonium hydroxide (solubilizers).

THE T3 PREMIX MUST BE SHAKEN WELL IMMEDIATELY BEFORE USE.

Reagents within each kit are specific to the lot number on the kit; therefore, only reagents from kits with the same lot number may be used interchangeably.

All reagents and samples must be equilibrated to room temperature before use.

Warnings

For *IN VITRO* Diagnostic Use.

Restricted Device: Federal law restricts the sale, distribution, or use of this device to, by, or on the lawful order of, a health professional.

Note: This radioactive material may be received, acquired, possessed, and used only by physicians, veterinarians in the practice of veterinary medicine, clinical laboratories, or hospitals and only for in vitro clinical or laboratory tests not involving internal or external administration of the material, or the radiation therefrom, to human beings or animals. Its receipt, acquisition, possession, use, and transfer are subject to the regulations and a general license of the US Nuclear Regulatory Commission or of a State with which the Commission has entered into an agreement for the exercise of regulatory authority.

E.R. Squibb and Sons, Inc.

DEC 17 1960

In vitro clinical laboratory testing with the T3-SQUIBB RADIO-IMMUNOASSAY KIT requires only a general license from the US Nuclear Regulatory Commission. The general license is issued to any physician, clinical laboratory, or hospital obtaining a validated registered USNRC Form 483. This form must be submitted in triplicate to the USNRC. The possessor of a general license is subject to the conditions and limitations under 10 CFR 31.11. (A specific license is available from the USNRC for quantities larger than 200 microcuries.)

Precautions

Observe the following safety rules in handling radioactive material:

- There should be no pipetting by mouth, smoking, or consumption of any food or drink in areas where radioactive materials are permitted.
- Wear gloves when handling radioactive materials.
- Wash hands thoroughly after handling radioactive materials.
- Working areas should be covered with disposable absorbent paper.
- Wipe up spills quickly and thoroughly and discard the contaminated materials with the radioactive waste.
- Solid waste may be stored in a specifically designated area in a covered metal or plastic container that has been identified with a USNRC radiation caution label until the waste decays to a safe level. Then, after defacing the labeling, the material may be discarded with nonradioactive waste.
- The user should determine whether USNRC and/or local regulations permit the liquid radioactive waste to be discarded through the sanitary sewerage system.

DEC 17 1980

Because T3 Standards contain human serum, these reagents, and patient samples, should be handled in the same manner as any potentially infective biological material. The human serum in the standards has been found to be nonreactive for HB_s Ag when tested with licensed reagents.

Chemical Hazard

- All of the reagents in this kit contain the commonly used bacteriostatic agent, sodium azide. Disposal of this material through the laboratory plumbing system could lead to the formation of highly explosive copper and lead azides. Therefore, it is recommended that large amounts of water be used to flush excess reagents through the plumbing system.

Storage

The kit should be refrigerated upon receipt.

All reagents should be stored in their original containers, with their original closures. Do not transfer these reagents to other containers or change the closures for storage purposes.

SPECIMEN COLLECTION AND PREPARATION

An appropriate quantity of blood should be collected from a peripheral vein using a standard blood collection tube that does not contain anticoagulant.

The patient does not need to be in a fasting state when the blood is withdrawn.

Serum, not plasma, should be used in performing the test. The variables introduced into the procedure through the use of plasma could make the results unreliable.

It is preferable to separate the serum sample on the same day that the blood is withdrawn from the patient. Serum samples may be stored in

a refrigerator for up to 24 hours, and should be frozen if stored for longer periods. Frozen samples should be equilibrated to room temperature and mixed well before use. Samples should not be refrozen.

The use of hemolyzed or lipemic samples is not recommended.

The patient's history should be thoroughly scrutinized to determine if any diagnostic or therapeutic radioisotopes have been administered within the week or two immediately prior to the T_3 determination. If the patient has been treated with radioisotopes or if the patient's history is unavailable, the activity of the serum sample should be checked in a scintillation well counter that is set for ^{125}I . If after a few seconds it appears that the count rate is above background, 20 microliters of the serum sample should be accurately counted to determine whether the radioactivity contained in this volume could significantly affect assay results.

TEST PROCEDURE

Materials Needed

In addition to the reagents supplied with the T3-SQUIBB RIA Kit, the following materials are required:

- 12 X 75 mm polystyrene test tubes
- Centrifuge capable of 3000 to 4000 rpm
- Well-type gamma scintillation counter (discriminator settings of 20-50 keV)
- Refrigerator
- Vortex mixer
- Pipettor and tips for 100 microliters (μl) [and, optionally 500 μl]
- Repeating pipettors capable of delivering 1 ml and 500 μl
- Container for radioactive waste
- Test tube racks

Precedural Precautions.

Good laboratory practice dictates that an appropriate number of controls be analyzed in conjunction with each set of samples.

All reagents and samples must be equilibrated to room temperature before use.

Procedure

The technique described below is based upon the performance of duplicate analyses for all standards, samples and controls. Read the entire procedure before starting the test.

1. Mark a series of 12 X 75 mm plastic test tubes with numerals 1 through 14. The first two tubes are used for measuring total counts and the remaining 12 tubes are required for generation of a standard curve. Two additional tubes will be needed for each clinical sample or control to be assayed.*
2. Add 100 μ l of T3 Standard or 100 μ l of the clinical sample to the respective tubes as follows:

<u>Tubes</u>	<u>Sample to be Added</u>
1 and 2	None
3 and 4	100 μ l of 0 ng/dl Standard
5 and 6	100 μ l of 50 ng/dl Standard
7 and 8	100 μ l of 100 ng/dl Standard
9 and 10	100 μ l of 250 ng/dl Standard
11 and 12	100 μ l of 500 ng/dl Standard
13 and 14	100 μ l of 1000 ng/dl Standard
15 and 16	100 μ l of Clinical Sample

3. Add 1.0 ml of T3 Premix to each tube. THIS REAGENT MUST BE SHAKEN WELL IMMEDIATELY BEFORE USE.
4. Add 500 μ l of T3 Antiserum to tubes 3 through 14.
5. Vortex each tube.
6. Set tubes 1 and 2 aside and incubate the remaining tubes at ambient temperature for 2 hours.

*If desired, nonspecific binding can be determined as follows:

- a) Mark two tubes as NSB.
- b) Add 100 μ l of "0" Standard to each tube.
- c) Add 500 μ l of distilled water to each tube.
- d) Add 1.0 ml of T3 Premix to each tube.
- e) Treat these two tubes in the same manner as tubes 3 through 14 by completing steps 5 through 10.

7. Centrifuge the incubated tubes for 5 to 10 minutes at 3000 to 4000 rpm.
8. Decant the supernatant by gently inverting each tube once, discarding the liquid into a radioactive waste container. Keeping the tube inverted, touch the rim on an absorbent paper and gently tap the tube.
9. Measure the radioactivity in all tubes in a standard well-type gamma scintillation counter (discriminator settings of 20 to 50 keV) for one minute or for a fixed amount of time that is sufficient to eliminate counting statistics as an important source of variability.
10. Subtract background cpm and record net cpm for each tube.

RESULTS

1. It is possible to plot either % Bound or B/Bo or net cpm versus concentration on semilog graph paper. Use the following formula to calculate % Bound for each tube and record the results on a worksheet.

$$\% \text{ Bound} = \frac{\text{PELLET COUNTS (net cpm)} \times 100}{\text{TOTAL COUNTS (net cpm)}}$$

Example:

Background cpm: 150
Tube #3 PELLET COUNTS (gross cpm): 52,398
TOTAL COUNTS (gross cpm): 98,227

$$\begin{aligned} \% \text{ Bound} &= \frac{52,398 - 150}{98,227 - 150} \times 100 \\ &= \frac{52,248}{98,077} \times 100 = 53.3 \end{aligned}$$

Alternately, use the following formula to calculate B/Bo values for each tube and record the results on a worksheet.

$$B/B_0 = \frac{\text{PELLET COUNTS (net cpm)}}{\text{Avg. PELLET COUNTS OF "0" Std. (net cpm)}}$$

Example:

Background cpm: 150
 Patient #1 PELLET COUNTS (gross cpm): 25,150
 "0" Std. - Avg. PELLET COUNTS (gross cpm): 50,150

$$B/B_0 = \frac{25,150 - 150}{50,150 - 150}$$

$$= \frac{25,000}{50,000} = 0.50$$

2. Prepare a standard curve by plotting either % Bound or B/B₀ or net cpm against the concentrations of the standards. A typical set of data calculated in terms of % Bound is given in Table I and graphed in Figure II. The same set of data calculated in terms of B/B₀ is listed in Table II and plotted in Figure III. These curves are provided for guidance only and should not be used in calculating the T₃ levels in the clinical samples.
3. Determine the concentration (ng/dl) of T₃ in each sample by referring to the standard curve prepared earlier. The observed % Bound or B/B₀ or net cpm for each sample will correspond to a specific T₃ concentration. Samples with T₃ concentrations of greater than 1000 ng/dl should be accurately diluted with "0" standard and reassayed, if an exact value is required.

TABLE I

Typical Standard Curve Data
Calculated in Terms of % Bound

TOTAL COUNT		69,700 cpm	
Standard Concentration (ng/dl)	Pellet Counts (net cpm)	% Bound	Avg. % Bound
0	38,551	55.3	
0	38,606	55.4	55.4
50	31,081	44.6	
50	31,652	45.4	45.0
100	26,712	38.3	
100	26,706	38.3	38.3
250	17,841	25.6	
250	17,860	25.6	25.6
500	11,522	16.5	
500	11,625	16.7	16.6
1000	7,274	10.4	
1000	7,189	10.3	10.4

TABLE II

Typical Standard Curve Data
Calculated in Terms of B/Bo

TOTAL COUNT		69,700 cpm	
Bo Counts (net avg.)		38,578 cpm	
Standard Concentration (ng/dl)	Pellet Counts (net cpm)	B/Bo	Avg. B/Bo
0	38,551	1.00	
0	38,606	1.00	1.00
50	31,081	0.80	
50	31,652	0.82	0.81
100	26,712	0.69	
100	26,706	0.69	0.69
250	17,841	0.46	
250	17,860	0.46	0.46
500	11,522	0.30	
500	11,625	0.30	0.30
1000	7,274	0.19	
1000	7,189	0.19	0.19

LIMITATIONS OF THE TEST PROCEDURE

In an individual having normal levels of TBG, the measurement of total T_3 yields an accurate diagnosis of thyroid status. However, there are many circumstances in which the level of TBG is not normal. For example, pregnancy or estrogen therapy cause increased synthesis of TBG and a concomittant increase in total T_3 whereas androgenic steroids have the opposite effect. Because of this variation in TBG levels, interpretation of T_3 results should be tempered by the determination of TBG binding capacity via the T_3 uptake assay.*

In addition, depressed levels of triiodothyronine have been observed in a wide variety of serious, nonthyroidal illnesses such as hepatic cirrhosis, anorexia nervosa, chronic renal failure, and disseminated malignancy,⁵⁻⁷ after surgery,⁸ and during caloric restriction.⁵

EXPECTED VALUES

Serum samples from 110 normal volunteers were assayed with the T3-SQUIBB RADIOIMMUNOASSAY Kit obtaining the following results:

Arithmetic Mean	115.3 ng/dl
Standard Deviation	17.5 ng/dl
Observed Range	70 - 155 ng/dl

Each sample was analyzed in duplicate. The frequency distribution of the patient values is depicted in Figure IV. With these results, the percentile estimate method yielded a normal range of 86 to 151 ng/dl, with 95% confidence.

It is recommended that each laboratory establish its own range of normal triiodothyronine values.

* T_3 uptake may be determined by use of the Squibb THYROSTAT®-3 Diagnostic Test Kit.

In another study 46 patient samples were analyzed by the T3-SQUIBB RADIOIMMUNOASSAY Kit, the T3RIA (PEG) Diagnostic Kit of Abbott Laboratories, and the IMMOPHASETM T-3 ¹²⁵I Radioimmunoassay Test System of Corning Medical. The regression analyses of these data are shown in Figures V and VI.

SPECIFIC PERFORMANCE CHARACTERISTICS

Standard Curve

Typical standard curves are shown in Figures II and III.

Sensitivity

The lowest triiodothyronine-containing standard supplied with the kit contains 50 ng/dl.

Specificity

The cross-reactivity of the antiserum with various substances is tabulated in Table III.

TABLE III

Cross-Reactivity of Various Substances with the T3-SQUIBB RADIOIMMUNOASSAY Kit

<u>Substance</u>	<u>Concentration</u>	<u>Percent Cross-Reactivity</u>
3,5-Diiodothyronine (T ₂)	50 µg/dl	0.5
Diphenylhydantoin	10 mg/dl	0
Iodoacetic acid	50 µg/dl	0
Phenyl butazone	10 mg/dl	0
Sodium salicylate	50 mg/dl	0
Reverse Triiodothyronine (rT ₃)	50 µg/dl	0.04
L-Thyroxine (T ₄)	40 µg/dl	0.05
	23 µg/dl	0

Recovery Efficiency

An important criterion of assay validity is the recovery of exogenous triiodothyronine from serum samples. Various amounts of exogenous T₃ were added to a serum pool having an endogenous T₃ content of 51 ng/dl and each sample was then assayed in duplicate. The results, which are presented in Table IV, demonstrate quantitative recovery of the exogenous triiodothyronine.

TABLE IV

Recovery Efficiency

Endogenous Concentration (ng/dl)	Amount Added (ng/dl)	Total Expected (ng/dl)	Total Found (ng/dl)	Percent Recovery*
51	50	101	107	105.9
51	100	151	161	106.6
51	250	301	325	108.0
51	500	551	590	107.1

$$* \text{Percent Recovery} = \frac{\text{Total Found}}{\text{Total Expected}} \times 100$$

Parallelism

This type of study indicates whether patient samples and standards behave similarly toward the radiolabel and the antiserum, thereby validating the matrix of the standards and the use of the zero standard as a diluent for patient samples. A serum sample having a known analyte concentration is serially diluted with zero standard and the original sample and its dilutions are assayed. A plot of observed dose against expected dose should theoretically yield a 45° line through the origin. Such a study was performed with three separate samples containing high levels of triiodothyronine. The results, depicted in Figure VII, demonstrate the validity of the matrix for the standards and the use of zero standard as a diluent.

Precision

Intra-assay CV data were obtained by performing a single run that consisted of 20 consecutive tubes each of human serum pools containing low, medium and high concentrations of thyroxine. In order to obtain inter-assay CV data, each of these controls was also assayed in duplicate in 25 runs done over a period of two months. The observed results yielded the reproducibility data shown in Table V.

TABLE V
Precision Data

Parameter	Serum Pool		
	Low	Medium	High
INTRA-ASSAY			
Mean (ng/dl)	79	158	313
Std. Dev. (ng/dl)	3.6	5.2	6.2
Range (ng/dl)	74-84	152-170	308-330
Coefficient of Variation (%)	4.6	3.3	2.0
INTER-ASSAY			
Mean (ng/dl)	74	156	320
Std. Dev. (ng/dl)	4.6	7.0	8.5
Range (ng/dl)	64-82	145-170	305-340
Coefficient of Variation (%)	6.2	4.5	2.7

REFERENCES

1. Sterling K, et al: T_3 thyrotoxicosis: Thyrotoxicosis due to elevated serum triiodothyronine levels. JAMA 213: 571-575, 1970
2. Burke CW, Eastman CJ: Thyroid hormones. Brit Med Bull 30: 93-99, 1974
3. Eastman CJ, et al: The radioimmunoassay of triiodothyronine and its clinical application. J Clin Pathol 28: 225-230, 1975
4. Utiger RD: Serum triiodothyronine in man. Ann Rev Med 25: 289-302, 1974
5. Cavalieri RR, Rapoport B: Impaired peripheral conversion of thyroxine to triiodothyronine. Ann Rev Med 28: 57-65, 1977 and references cited therein
6. McLarty DG et al: Thyroid-hormone levels and prognosis in patients with serious non-thyroidal illness. Lancet II: 275-276, 1975
7. Spector DA et al: Thyroid function and metabolic state in chronic renal failure. Ann Int Med 85: 724-730, 1976

DEC 17 1980

8. Burr WA et al: Serum triiodothyronine and reverse triiodothyronine concentrations after surgical operation. Lancet II: 1277-1279, 1975

E. R. Squibb & Sons, Inc.
Princeton, NJ 08540




Figure II. A graph of % Bound vs. T3 Concentration for a typical standard curve obtained with the T3-SQUIBB RADIOIMMUNOASSAY KIT. As indicated, a patient sample having a mean % Bound of 32.5 would have a triiodothyronine concentration of 154 ng/dl.




Figure III. A graph of B/B_0 vs. T3 Concentration for a typical standard curve obtained with the T3-SQUIBB RADIOIMMUNOASSAY KIT. As indicated, a patient sample having a mean B/B_0 of 0.75 would have a triiodothyronine concentration of 75 ng/dl.




Figure IV. Frequency distribution of 110 samples from normal volunteers.




Figure V. Regression analysis of 46 samples analyzed by the T3-SQUIBB Radioimmunoassay Kit and the Abbott Laboratories T3RIA (PEG) Diagnostic Kit.




Figure VI. Regression analysis of 46 samples analyzed by the T3-SQUIBB Radioimmunoassay Kit and the Corning Medical IMMOPHASE™ T-3¹²⁵I Radioimmunoassay Test System.




Figure VII. The parallelism study: A graph of observed dose versus expected dose. The theoretical 45° line is shown.



Store at 2° to 8° C



100 ml

List G2110

¹²⁵I T4 REAGENT

For use with T4 CLASP® RIA Kit
For *in vitro* Diagnostic Use • See directions
Not for Internal or External Use in Humans or Animals
SODIUM AZIDE: 0.1%
SHAKE WELL BEFORE USE

Total
Act.: <3.96 microcuries ¹²⁵I
LOT
NO.:

As
of
EXP
DATE:

Non
EST

E. R. Squibb & Sons, Inc., Princeton, NJ 08540
Made in USA

C6286A / G2110



T4 CLASP® RIA Kit

List G2110

Squibb T4 Radioimmunoassay Kit

For *in vitro* Diagnostic Use • See enclosed directions

Not for Internal or External Use in Humans or Animals

Contains sufficient material for 100 tubes

Store at 2° to 8° C

CONTENTS:

1 bottle (100 ml) ¹²⁵I T4 REAGENT
<3.96 microcuries ¹²⁵I per bottle
5 bottles (1 ml ea.) T4 RIA STANDARD
(1 each of 1, 5, 10, 15 and 25 µg per dl)
1 bottle (1 ml) T4 RIA CONTROL

E. R. Squibb & Sons, Inc., Princeton, NJ 08540
Made in USA

M5367A / G2110

06242



J3-438A

Revised February 1980

T4 CLASP® RIA Kit

For Quantitative Measurement of Serum Thyroxine Levels by Radioimmunoassay

**For *IN VITRO* Diagnostic Use
For Professional Use Only**

DETERMINATION OF SERUM THYROXINE LEVELS BY RADIOIMMUNOASSAY

Measurement of body constituents or administered compounds by the technique of radioimmunoassay offers a bio-analytical tool that combines the extreme sensitivity of radioisotope methodology with the extreme specificity of immunological techniques. The procedure requires a specific antibody, a radiolabeled antigen, a pure sample of the antigen to serve as a reference standard, and a means of separation of free antigen from antibody-bound antigen. The procedure follows the basic principle of saturation analysis, where there is competition between labeled and unlabeled antigen for a fixed number of antibody binding sites. As the concentration of unlabeled antigen (the substance actually being measured) increases, less of the added radiolabeled antigen will be bound to the antibody. After a specific period of time, the

other confirmatory tests such as TSH assay, serum liothyronine (T_3) uptake by the competitive protein-binding technique (CPB), T_3 by RIA, free thyroxine, or free thyroxine index (FTI) may be helpful in determining thyroid status.

Until recently the most commonly used method for the determination of total serum thyroxine utilized the thyroxine-binding properties of serum TBG in the CPB technique. The use of the more recently introduced radioimmunoassay technique for serum thyroxine determination offers several advantages over the CPB technique. Extraction of the thyroxine from serum is eliminated, thereby eliminating the problem of variation in extraction efficiency. The serum sample size required for RIA (10 μ l) is much less than that used in the CPB technique; cases in which limited quantities of serum are available (e.g., pediatric or elderly patients), are more readily accommodated. Use of high-affinity antiserum in the RIA



50 ml REFRIGERATE AT 2° to 8° C List 09127

THYROSTAT®FTI BUFFER SOLUTION

Contains T4 Binding Globulin and I¹²⁵-Thyroxine
FOR LABORATORY USE ONLY with Thyrostat-FTI
Adsorbent Tablets • See accompanying directions
For *in vitro* Diagnostic Use • Shake well before using
SODIUM AZIDE: 0.1%

Not for Internal or External Use in Humans or Animals

Total	As	Noon
Act. <2.6 microcuries I ¹²⁵	of	EST
LOT	EXP.	
NO.	DATE	

E. R. Squibb & Sons, Inc.
Princeton, NJ 08540

Made in USA C6031B / 09127



25 TEST PKG.

THYROSTAT®FTI DIAGNOSTIC TEST KIT

For *in vitro* Diagnostic Use
For Evaluation of Thyroid Function
FOR LABORATORY USE ONLY

Not for Internal or External Use in Humans or Animals

REFRIGERATE AT 2° to 8° C

CONTENTS:

- 1 vial (28 tabs.) THYROSTAT-FTI ADSORBENT TABLETS
- 1 bottle (50 ml) THYROSTAT-FTI BUFFER SOLUTION
- <2.6 microcuries I¹²⁵ per bottle
- 1 vial (3.0 ml) THYROSTAT-FTI CONTROL SERUM
- 1 vial (25 ml) THYROSTAT-FTI EXTRACTION ALCOHOL

List 09127

NOTE: All reagents in this kit must be used with the accompanying
Adsorbent Tablets • See accompanying directions

E. R. Squibb & Sons, Inc.
Princeton, NJ 08540

Made in USA

C6040A / 09127



200 ml REFRIGERATE AT 2° to 8° C List 09152

THYROSTAT®FTI BUFFER SOLUTION

Contains T4 Binding Globulin and I¹²⁵-Thyroxine
FOR LABORATORY USE ONLY with Thyrostat-FTI
Adsorbent Tablets • See accompanying directions
For *in vitro* Diagnostic Use • Shake well before using
SODIUM AZIDE: 0.1%


Not for Internal or External Use in Humans or Animals

Total	As	Noon
Act. <9.7 microcuries I ¹²⁵	of	EST
LOT	EXP.	
NO.	DATE	

E. R. Squibb & Sons, Inc.
Princeton, NJ 08540

Made in USA

C6034B / 09152



100 TEST PKG.

THYROSTAT®FTI DIAGNOSTIC TEST KIT

For *in vitro* Diagnostic Use
For Evaluation of Thyroid Function
FOR LABORATORY USE ONLY

Not for Internal or External Use in Humans or Animals
REFRIGERATE AT 2° to 8° C

CONTENTS:

- 1 vial (105 tabs.) THYROSTAT-FTI ADSORBENT TABLETS
- 1 bottle (200 ml) THYROSTAT-FTI BUFFER SOLUTION
- <9.7 microcuries I¹²⁵ per bottle
- 1 vial (3.0 ml) THYROSTAT-FTI CONTROL SERUM
- 1 vial (100 ml) THYROSTAT-FTI EXTRACTION ALCOHOL

List 09152

NOTE: All reagents in this kit must be used with the accompanying
Adsorbent Tablets • See accompanying directions

E. R. Squibb & Sons, Inc.
Princeton, NJ 08540

Made in USA

C6041A / 09152



500 TEST KIT

List 09172

THYROSTAT®-FTI DIAGNOSTIC TEST KIT

For *in vitro* Diagnostic Use

For Evaluation of Thyroid Function

FOR LABORATORY USE ONLY • See accompanying directions

Not for Internal or External Use in Humans or Animals

CONTENTS:

500 plastic test tubes

5 vials (105 tabs. each) THYROSTAT-FTI ADSORBENT
TABLETS

5 bottles (200 ml each) THYROSTAT-FTI BUFFER
SOLUTION

<9.7 microcuries I 125 per bottle

5 vials (3.0 ml each) THYROSTAT-FTI CONTROL SERUM

5 vials (100 ml each) THYROSTAT-FTI EXTRACTION
ALCOHOL

Note: All reagents in this kit must be used with the accompanying
Adsorbent Tablets

REFRIGERATE AT 2° to 8° C

E. R. Squibb & Sons, Inc.

Princeton, NJ 08540

Made in USA

C8041A / 09172



THYROSTAT®-FTI DIAGNOSTIC TEST KIT

Revised February 1980

For Quantitative Measurement of Free Thyroxine Index
for the Evaluation of Thyroid Function

For *In Vitro* Diagnostic Use
For Professional Use Only

PRINCIPLES OF THE TEST

In the absence of abnormalities in serum thyroxine-binding proteins (TBP) concentration, the measurement of total serum thyroxine provides an accurate means of assessing thyroid status. In conditions where TBP concentration is altered, it is also necessary to obtain an estimate of the unoccupied binding sites of the TBP, in order to obtain an accurate indication of thyroid status. The Thyrostat-FTI procedure provides a means for simultaneous measurement of total serum thyroxine and estimation of unoccupied binding sites of the TBP. The result obtained by this procedure provides an indirect measure of free thyroid hormone, which correlates closely with thyroid status, regardless of serum TBP concentration.

The first step in the assay procedure involves liberation of the bound thyroxine and extraction of the major portion of the liberated and free thyroxine with an alcoholic solvent. Subsequent to the extraction step, the procedure follows the basic principle of saturation analysis and requires specific thyroxine-binding proteins, a radiolabeled derivative of thyroxine, and a normal control serum. With the saturation analysis technique, there is competition between labeled and unlabeled thyroxine for a fixed number of protein binding sites. As the concentration of unlabeled thyroxine (obtained from the extraction step) increases, less of the radiolabeled thyroxine will be retained by the binding proteins. The quantity of thyroxine which is not bound by the specific binding proteins is measured by introducing an adsorbent material that will only bind free thyroxine. The relative amounts of free and bound thyroxine are determined by isolating the adsorbent and measuring the radioactivity associated with it. The estimation of unoccupied binding sites of the TBP is achieved by adding a small quantity of unextracted patient serum to the incubation medium prior to addition of the adsorbent material. The amount of radiolabeled thyroxine taken up by the adsorbent material will be influenced by the number of unoccupied TBP binding sites present in the patient serum sample. As the number of unoccupied binding sites increases, the amount of radioactivity taken up by the adsorbent material will decrease. The net effect of addition of unextracted patient serum is adjustment of assay results to compensate for alterations in TBP concentration. Serum thyroxine measurements adjusted for TBP concentration provide an indirect measure of serum-free thyroxine, which is highly correlated with thyroid status. An "index" of free thyroid hormone is obtained by comparing adsorbent uptake of radiolabeled thyroxine in a

J3-415C

patient sample with adsorbent uptake of radiolabeled thyroxine, for a normal control serum sample assayed in an identical manner. When the thyroid gland produces an excess of thyroid hormones, as in hyperthyroidism, the increased level of free thyroxine will be reflected by an increased uptake of radiolabeled thyroxine by the adsorbent material, when compared with the uptake for a normal control serum. Conversely, when thyroid hormone production is decreased, as in hypothyroidism, the decreased level of free thyroxine will be reflected by a decreased adsorbent uptake of radiolabeled thyroxine, compared with the uptake for a normal control serum.

In the Thyrostat-FTI test, the thyroxine is extracted from serum with ethanol, and the alcoholic extract is mixed with ^{125}I -labeled thyroxine bound to thyroxine-binding proteins. An organic adsorbent is used to separate bound thyroxine from free thyroxine. The normal control serum is of animal origin.

NOMENCLATURE

It has been suggested by an *ad hoc* committee of the American Thyroid Association that the measurement of total serum thyroxine by the classic Murphy-Pattee "radiotransin assay" be identified as "Thyroxine (displacement)," abbreviated " $T_4(\text{D})$," and expressed as micrograms per 100 milliliters of serum ($\mu\text{g}/100\text{ ml}$).

When there are pronounced alterations in the binding capacity of the transport proteins, the $T_4(\text{D})$ test, used in conjunction with the commonly used T_3 test (Resin Triiodothyronine uptake-RT₃U), provides an indirect measure of the concentration of free T_4 , which is highly correlated with the thyroid state.² This indirect measure of free T_4 is commonly referred to as the "free thyroxine index," with the American Thyroid Association suggested designation being "Thyroxine-resin T_3 index," abbreviated $T_4\text{-RT}_3$. The $T_4\text{-RT}_3$ is the mathematical product of the results of the $T_4(\text{D})$ and RT₃U tests.

Nomenclature has not been suggested for tests that provide a direct measurement of "free thyroxine index," as opposed to the calculation procedure described above. For purposes of simplicity, the direct measurement of "free-thyroxine index," as embodied in the Thyrostat-FTI procedure, will be abbreviated FTI(M).

RATIONALE FOR USE

The measurement of total serum thyroxine (T_4) by the $T_4(\text{D})$ procedure represents a significant advance in the *in vitro*

J3-415C