

## U.S. NUCLEAR REGULATORY COMMISSION

Amendment No. 01

**MATERIALS LICENSE**

Corrected Copy

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p>Licensee</p> <p>1. Indiana Department of Transportation Fort Wayne District</p> <p>2. 5333 Hatfield Road Fort Wayne, IN 46808</p>	<p>In accordance with application dated January 11, 2002,</p> <p>3. License number 13-26345-01 is renewed in its entirety to read as follows:</p> <p>4. Expiration date February 29, 2012</p> <p>5. Docket No. 030-32467 Reference No.</p>	
<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Cesium-137</p> <p>B. Americium-241</p>	<p>7. Chemical and/or physical form</p> <p>A. Sealed sources registered either with NRC under 10 CFR 32.210 or with an Agreement State and incorporated in a compatible gauging device as specified in Item 9 of this license.</p> <p>B. Sealed sources registered either with NRC under 10 CFR 32.210 or with an Agreement State and incorporated in a compatible gauging device as specified in Item 9 of this license.</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. No single source to exceed the maximum activity specified in the certificate of registration issued by NRC or an Agreement State. <b>Total activity not to exceed 162 millicuries.</b></p> <p>B. No single source to exceed the maximum activity specified in the certificate of registration issued by NRC or an Agreement State. <b>Total activity not to exceed 792 millicuries.</b></p>

## 9. Authorized Use:

- A. and B. To be used in Troxler Model 3400 series surface moisture/density gauges.
- B. To be used in Troxler Model 3241 asphalt content gauges.

CONDITIONS

10. Licensed material may be stored at 5333 Hatfield Road, Fort Wayne, Indiana and may be used at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.
11. A. The Radiation Safety Officer (RSO) for this license is Phil Ellet

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number

13-26345-01

Docket or Reference Number

030-32467

Amendment No. 01

Corrected Copy

- B. Before assuming the duties and responsibilities as RSO for this license, future RSOs shall have successfully completed one of the training courses described in Criteria in Section 8.8 of NUREG-1556, Volume 1, dated May 1997.
12. Licensed material shall only be used by, or under the supervision and in the physical presence of, individuals who have received the training described in application dated January 11, 2002.
13. A. (1) The source(s) specified in Item(s) 7.A. through 7.B. shall be tested for leakage and/or contamination at intervals not to exceed 6 months. Any source received from another person which is not accompanied by a certificate indicating that a test was performed within 6 months before the transfer shall not be put into use until tested.
- (2) Notwithstanding the periodic leak test required by this condition, any licensed sealed source is exempt from such leak tests when the source contains 100 microcuries or less of beta and/or gamma emitting material or 10 microcuries or less of alpha emitting material.
- B. Any source in storage and not being used need not be tested. When the source is removed from storage for use or transfer to another person, it shall be tested before use or transfer.
- C. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the source shall be removed from service and decontaminated, repaired, or disposed of in accordance with Commission regulations. A report shall be filed within 5 days of the date the leak test result is known with the U.S. Nuclear Regulatory Commission, Region III, 801 Warrenville Road, Lisle, Illinois 60537, ATTN: Chief, Nuclear Materials Safety Branch. The report shall specify the source involved, the test results, and corrective action taken. Records of leak test results shall be kept in units of microcuries and shall be maintained for inspection by the Commission. Records may be disposed of following Commission inspection.
- D. The licensee is authorized to collect leak test samples for analysis by Troxler or tests for leakage and/or contamination shall be performed by persons specifically licensed by the Commission or an Agreement State to perform such services.
13. Sealed sources or source rods containing licensed material shall not be opened or sources removed or detached from source rods or gauges by the licensee.
14. Except for maintaining labeling as required by 10 CFR Part 20 or 71, the licensee shall obtain authorization from NRC before making any changes in the sealed source, device, or source-device combination that would alter the description or specifications as indicated in the respective Certificates of Registration issued either by the Commission pursuant to 10 CFR 32.210 or by an Agreement State.
15. The licensee shall conduct a physical inventory every 6 months, or at other intervals approved by NRC, to account for all sources and/or devices received and possessed under the license.
16. The licensee is authorized to transport licensed material only in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**License Number  
13-26345-01Docket or Reference Number  
030-32467Amendment No. 01  
Corrected Copy

17. Each portable gauge shall have a lock or outer locked container designed to prevent unauthorized or accidental removal of the sealed source from its shielded position. The gauge or its container must be locked when in transport, storage, or when not under the direct surveillance of an authorized user.
18. Any cleaning, maintenance, or repair of the gauges that requires detaching the source or source rod from the gauge shall be performed only by the manufacturer or other persons specifically licensed by the Commission or an Agreement State to perform such services.
19. A. If the licensee uses unshielded sealed sources extended more than 3 feet below the surface, the licensee shall use surface casing that extends from the lowest depth to 12 inches above the surface and other appropriate procedures to reduce the probability of the source or probe becoming lodged below the surface. If it is not feasible to extend the casing 12 inches above the surface, the licensee shall implement procedures to ensure that the cased hole is free of obstruction before making measurements.
- B. If a sealed source or a probe containing sealed sources becomes lodged below the surface and it becomes apparent that efforts to recover the sealed source or probe may not be successful, the licensee shall notify the U. S. Nuclear Regulatory Commission and submit the report required by 10 CFR 30.50(b)(2) and (c). The licensee shall not abandon the sealed source or probe without obtaining the Commission's prior written consent.
20. When performing tests at temporary job sites, the authorized user shall not leave the moisture/density gauge unattended. Upon completion of tests the device shall be locked in the licensee's vehicle or a secure building to prevent unauthorized use, loss or theft.
21. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents including any enclosures, listed below. The Nuclear Regulatory Commission's regulations shall govern unless the statements, representations and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application dated January 11, 2002, and;
- B. Facsimile dated January 31, 2002.

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date JUN 29 2011By Michael G. Herr, CHP  
Materials Licensing Branch  
Region III