

From: Ullrich, Elizabeth
Sent: Friday, July 15, 2011 8:42 AM
To: Nag, Jay - Middlebury, CT; Paul Steinmeyer (kpstein@radpro.com)
Subject: two questions re: termination of license

License No. 06-00221-08
Docket No. 030-03759
Control No. 575491

Jay and Paul,

1. Page 21 states that RSA planned to take an LSC to their facility by the end of June. Was this device moved?
2. In the past, Chemtura/Uniroyal had a number of gas chromatographs with Ni63 ECDs. Although these have not been listed on the license for some time, the ECDs are a common item that can be possessed under the general license in 10 CFR 31.5. Does Chemtura have any Ni63 ECDs under a general license remaining at the site? How about other possible items under a general license – Po210 static eliminators, EXIT signs? If so, let me know what you plan to do with such items; there are restrictions on how items under a general license may be transferred, and to whom they may be transferred.

Please respond to this item in a hard-copy letter, not by email. Contact me if you have any questions at this email or by telephone as listed below.

Thanks,
Betsy
Betsy Ullrich, MS, CHP
Senior Health Physicist
Commercial and R&D Branch
Division of Nuclear Material Safety
Region I
Nuclear Regulatory Commission
(610) 337-5040
Elizabeth.Ullrich@nrc.gov