



# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory  
Certificate #0972-01

## Certificate of Calibration

**Cost Center / Customer:** DIV20 / DON BANNON

**Mail Stop:** B51

**Manufacturer/Model:** SWRI / 1000 OHM

**Description:** RESISTOR BOX

**Serial Number:** 171001

**Asset Number:** 003096

**Procedure:** RESISTORS 3458A METHOD - 26 FEB 09

**Work Order:** 303099178

**Date Issued:** 11-Jan-2011

**Date Calibrated:** 11-Jan-2011

**\* Date Due :** 11-Jul-2011

**\*\* Results:** FOUND-LEFT

**Temperature:** 74.0 °F

**Humidity:** 40 %RH

**Barometer:** N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.


\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.


Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of  $k=2$ .

**Remarks:**  $\pm 1.0\%$

### Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
001505	HEWLETT-PACKARD	3458A/OPT 002	MULTIMETER	5-Feb-2010	5-Feb-2011

  
Walt Hill  
Laboratory Manager

  
Bob Trollinger  
Metrology Technician

Southwest Research Institute  
Calibration Laboratory  
Measurement Report

Work Order:	303099178	Mfr:	SWRI	Technician:	blt
Asset No:	003096	Model:	1000 Ohm	Type Data:	Found-left
Serial No:	171001	Type:	Resistor Box	Cal Date:	11-Jan-11
Remarks: Limits of $\pm 1.0$ % per customer request.					

Function/Range	Test Point	TI Reading	Difference	+/- Limit	+/- Uncertainty	Result	% Limit
Resistance	$\Omega$	$\Omega$	$\Omega$	$\Omega$	$\Omega$		
Green to White	1000.00	1000.52	0.52	10	0.016	Pass	5%
END OF REPORT							