



UNITED STATES
ATOMIC ENERGY COMMISSION
DIVISION OF COMPLIANCE
REGION I
970 BROAD STREET
NEWARK, NEW JERSEY 07102

AREA CODE: 201-
TELEPHONE: 645-3962

IN REPLY REFER TO:

8/19/69

TO PRN

FROM CFS

B/M
Evaluation of
Beryllium Corporation
Hazleton, Pa
Lic No 37-07676-02

The only use under this license is the operation of an instrument containing 250 mc Sb-124. There is no direct handling of byproduct material.

Safety procedures and control appeared adequate for the potential hazards involved. Surveys are conducted monthly, and when replacement sources are received. Personnel exposures are minimal.

In the judgment of the inspector, there are no thus significant potential health and safety hazards; it is recommended that the time for re-inspection be extended by eighteen months to January, 1974. This license has been correctly classified as E(2) IV.

Charles F. Neams, RADIATION SPECIALIST