

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Columbus Imaging Associates dba Columbus Diagnostic Imaging 790 Creekview Drive Columbus, IN 47201 REPORT Nos 2011-001		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III, 2443 Warrenville Road Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-34441	4. LICENSE NUMBER(S) 13-26797-01	5. DATE(S) OF INSPECTION 7/18/2011	

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

☐ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- ☐ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Tara L. Weidner		7/18/2011
BRANCH CHIEF	Tamara Bloomer		8/11/11

SUNSI Review Completed By: /RA/ TLWeidner

☒ Public
 ☒ Non-Sensitive

Initial		Announced	X	Unannounced	X	Routine		Special	Increased Controls
---------	--	-----------	---	-------------	---	---------	--	---------	--------------------

NRC FORM 591M PART 3
(10-2011) 10 CFR 2.201

Docket File Information
**SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

U.S. NUCLEAR REGULATORY COMMISSION

1. LICENSEE Columbus Diagnostic Imaging REPORT NO(S): 2011-001	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III, 2443 Warrenville Road Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-34441	4. LICENSE NUMBER(S) 13-26797-01	5. DATE(S) OF INSPECTION July 18, 2011
6. INSPECTION PROCEDURES USED 87131	7. INSPECTION FOCUS AREAS 3.01 – 3.07	8. INSPECTOR T. Weidner

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM 2200	2. PRIORITY 3	3. LICENSEE CONTACT Robert DeWeese, M.D. - RSO	4. TELEPHONE NUMBER 812-376-1000
---------------------------	-------------------------	--	--

☒ Main Office Inspection
☐ Field Office
☐ Temporary Job Site Inspection

Next Inspection Date: July 2014

PROGRAM SCOPE

Columbus Diagnostic is a small outpatient facility authorized for 35.100, 35.200, 35.300, and 31.11. The facility currently operates two days per week. One nuclear medicine technologist (NMT) operating one camera performs approximately 20 studies per month. Studies using technetium-99m unit doses include bone, lung, HIDA, gastric emptying and iodine-123 for uptakes. No heart studies are performed at this facility. Iodine-131 was used for hyperthyroid treatments three times in 2010. Appropriate written directives were prepared. No NARM or 31.11 materials are currently used at this facility. Unit doses are obtained from a licensed radiopharmacy. All doses are assayed in a dose calibrator prior to administration. Sealed source inventory includes 4 cesium-137 sources and one cobalt-57 source.

The Radiation Safety Officer is one of the authorized users on the license. At least one of the authorized users is on site on a daily basis. Advanced Medical Physics provides consulting services on a quarterly basis. Services include program audits, equipment QC, sealed source inventory and leak tests, dosimetry review, training, etc.

Performance Observations

On the day of the inspection, no procedures were scheduled and the NMT was not present. The inspection consisted of a telephonic interview of the NMT, tour of the nuclear medicine department, independent surveys, and a review of selected records.

An exit meeting was held with Rhonda Lennon, Practice Manager to review the findings of the inspection. No violations were identified and a clear Form 591M Part 1 was issued in the field.