

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Heart To Heart
Roseville, Michigan
REPORT NUMBER(S) 2011-001

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Lisle, IL 60532

3. DOCKET NUMBER(S)

030-35352

4. LICENSEE NUMBER(S)

21-32248-01

5. DATE(S) OF INSPECTION

August 11, 2011

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



1. Based on the inspection findings, no violations were identified.



2. Previous violation(s) closed.



3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied

_____ Non-cited violation(s) were discussed involving the following requirement(s):



4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura	<i>Deborah A. Piskura</i>	8/11/2011
Branch Chief	Tamara E. Bloomer	<i>Tamara Bloomer</i>	8/22/11

NRC FORM 591 M PART 3
(06-2010)
10 CFR 2.201

U.S. NUCLEAR REGULATORY COMMISSION

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE
Heart To Heart
25689 Kelly Road
Suite D
Roseville, MI 48066
REPORT NUMBER(S) 2011-001

2. NRC/REGIONAL OFFICE
U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4351

3. DOCKET NUMBER(S)
030-35352

4. LICENSE NUMBER(S)
21-32248-01

5. DATE(S) OF INSPECTION
August 11, 2011

6. INSPECTION PROCEDURES
87130

7. INSPECTION FOCUS AREAS
03.01-03.08

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM
02201

2. PRIORITY
5

3. LICENSEE CONTACT
Mukhtar A. Khan, M.D., RSO

4. TELEPHONE NUMBER
586-774-0700

☒ Main Office Inspection

Next Inspection Date: August 2016

Field Office Inspection

☐ Temporary Job Site Inspection _____

PROGRAM SCOPE

This licensee was a private cardiology office and authorized to use licensed material permitted by Sections 35.100 and 35.200. The nuclear medicine department was staffed with one technologist who performed approximately 35-40 diagnostic procedures monthly. The licensee received unit doses from a licensed radiopharmacy. The clinic administered studies on Tuesdays and Thursdays. The clinic retained the services of a consulting physicist who audited the radiation safety program on a quarterly basis.

This inspection consisted of interviews with licensee personnel, a review of selected records, a tour of the nuclear medicine department, and independent measurements. The inspector observed the administration of two cardiac stress testing procedures. The inspector observed licensee personnel demonstrate dose calibrator QA tests, package receipt and surveys, and area radiation surveys.