

From: Nguyen, Janice
Sent: Thursday, August 25, 2011 3:37 PM
To: pwaite@rrmc.org
Subject: NRC Request for Information (Mail Control Number 575770)

Licensee: Rutland Regional Medical Center
License No. 44-14121-01
Docket No. 030-07587
Mail Control 575770

Dear Mr. Waite,

As per our telephone conversation on August 25, 2011, we need the following additional information to support your amendment request for the new facility:

1. Please confirm the name and address of the new facility.
2. Please indicate whether any other entity currently uses radioactive material at your new address of use.
3. Please confirm the requested licensed material (e.g., 35.100, 35.200) and use (e.g., 35.100, 35.200) at your new facility.
4. Please indicate whether PET radionuclides will be used, and if so, please provide shielding evaluations for areas of use. In addition, please provide the manufacturer name and model number for any sealed sources that do not meet the criteria in 10 CFR 35.65 (e.g., greater than 30 mCi).
5. Please submit facility diagrams indicating the position of each area of use and storage and describe the type, dimensions, and thickness of shielding that you will use for these areas. Specifically, please indicate which areas will be used for radioactive material use and/or storage. If the waste storage area is not located within your main department, describe how you will secure the material. In addition, please describe any additional shielding used in radiopharmaceutical preparation and dispensing areas (e.g., lead glass L-block, specifically designed PET shielding, etc.). Finally, identify adjacent areas from use and storage locations, including areas above/below, and show that adequate steps have been taken to assure that radiation levels in unrestricted areas will not result in doses to individual members of the public in excess of those specified in 10 CFR 20.1301.
6. Please confirm that all equipment, instrumentation, and sealed sources for the new facility will be as described in the letters dated July 10, 2003 and December 4, 2005. These are the tie-downs for the existing facilities at 6 Commons Street, Rutland, VT and 1 Commons Street, Rutland, VT.

The response letter should be signed by a member of senior management. You may fax the response to my attention to 610-337-5269. Please reference Mail Control No. 575770 in your response.

Please let me know if you have any questions.

Thank you.

Sincerely,

Jan Nguyen

Janice Nguyen
Health Physicist
Nuclear Regulatory Commission Region I
475 Allendale Road
King of Prussia, PA 19406
(610) 337-5006
FAX (610) 337-5269