



(517) 546-6121
FAX (517) 546-1478

**HASTINGS TESTING ENGINEERS
AND ENVIRONMENTAL INC.**

"Testing to keep America on a firm foundation"

4841 GOLF CLUB ROAD • HOWELL, MI 48843

January 17, 2012

United States Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Suite #120
Lisle, IL 60532-4352

Attention: Frank P D Tran

Reference: NRC Request for Additional Information - Control Number 575693
Application for NRC License Renewal - License # 21-26296-01

Hastings Testing Engineers and Environmental Inc. commits to the following:

1. Before obtaining licensed materials, the proposed RSO will have successfully completed one of the training courses described in Criteria in the section entitled "Individual(s) Responsible for Radiation Safety Program and Their Training and Experience - Radiation Safety Officer in NUREG - 1556, Vol.1, dated November 2001.
2. Before using licensed materials, authorized users will have successfully completed one of the training courses described in Criteria in the section entitled "Training for Individuals Working in or Frequenting Restricted Areas" in NUREG-1556, Vol. 1, Rev 1, dated November 01, 2001.
3. Our firm will possess and use, or have access to and use, a radiation survey meter that meets the Criteria in the section entitled Radiation Safety Program - Instruments" in NUREG 1556, Vol. 1., dated November 2001.
4. Physical inventories will be conducted at intervals not to exceed six months, to account for all sealed sources and devices received and possessed under the license.
5. Our firm will maintain, for inspection by the NRC, documents demonstrating that unmonitored individuals are not likely to receive a radiation dose in excess of 10 percent of the allowable limits in 10 CFR Part #20, or the our firm will provide dosimetry processed and evaluated by an NVLAP approved processor that is exchanged at a frequency recommended by the processor.
6. Operating and emergency procedures will be development, implemented, and maintained that will meet the criteria in the section titled radiation Safety Program - Operating and Emergency Procedures in NUREG-1556, Vol. 1, Rev. 1, dated November 2001.

RECEIVED JAN 17 2012



(517) 546-8121
FAX (517) 546-1478


**HASTINGS TESTING ENGINEERS
AND ENVIRONMENTAL INC.**

"Testing to keep America on a firm foundation"

4641 GOLF CLUB ROAD • HOWELL, MI 48843

7. Our firm will implement and maintain procedures for routine maintenance on our gauges according to each manufacturer's recommendations and instructions.
8. Our firm will send our gauges to the manufacturer or other person authorized by the NRC or an Agreement State to perform non-routine maintenance or repair operations that require the removal of the source or source rod from the gauge.
9. Leak test will be performed at intervals approved by the NRC or Agreement State and specified in the Sealed Source and Device Registration Sheet. Leak Test will be performed by an organization authorized by the NRC or an Agreement State to provide leak testing services for other licensees or using a leak test kit supplied by an organization authorized by the NRC or Agreement State to provide leak test kits to other licensees and according to the kit supplier's instructions.

Sincerely,


Robert Hastings
President