

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Midwest Division – LRHC, LLC
15th and State Streets
Lexington, Missouri 64067

REPORT NUMBER(S) 12-01

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-31699

4. LICENSE NUMBER(S)

24-26183-01

5. DATE(S) OF INSPECTION

January // , 2012

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Andrew M. Brammick	Andrew M. Brammick	1/11/2012
BRANCH CHIEF	Tamara Bloomer	T. Bloomer	1/20/12

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Midwest Division – LRHC, LLC 15th and State Streets Lexington, Missouri 64067 REPORT NUMBER(S) 12-01		2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
3. DOCKET NUMBER(S) 030-31699	4. LICENSE NUMBER(S) 24-26183-01	5. DATE(S) OF INSPECTION January 11, 2012
6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS 03.01 - 03.08	

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02121	2. PRIORITY 5	3. LICENSEE CONTACT Jay D. Crockett, D.O. - RSO	4. TELEPHONE NUMBER (660) 259-2203
--	-----------------------------	---	--

<input checked="" type="checkbox"/> Main Office Inspection	Next Inspection Date:	1/1/2017
<input type="checkbox"/> Field Office Inspection		
<input type="checkbox"/> Temporary Job Site Inspection		

PROGRAM SCOPE

This was a routine inspection of a 25 bed critical access hospital that performed approximately 25 diagnostic nuclear medicine procedures per month. One part time nuclear medicine technologist employed by a contract group of technologists performed all patient procedures on Mondays, Wednesday, and Fridays. The licensee obtained unit doses of licensed material from an area nuclear pharmacy, and did not use xenon-133 or molybdenum/technetium generators. The licensee performed primarily cardiac, bone, and thyroid scans, and was not authorized to administer therapeutic doses of licensed material.

PERFORMANCE OBSERVATIONS

Interviews of available staff revealed an adequate level of understanding of emergency and material handling procedures and techniques. The licensee successfully described dose calibrator constancy checks, package receipt, daily surveys, and waste handling and disposal procedures. The inspector confirmed that these activities were successfully completed by reviewing selected records since the previous inspection. A nuclear medicine technologist employed by the hospital's contract group performed quarterly program audits that were adequate to oversee the program.

Licensed material was adequately secured and not readily accessible to members of the general public. The licensee possessed a radiation survey meter that was calibrated and operational. Independent measurements did not indicate readings in excess of Title 10 of the Code of Federal Regulations (10 CFR) Part 20 limits in restricted or unrestricted areas. Personal whole body and extremity dosimetry badges were observed being worn by the staff during the inspection, and records did not indicate doses in excess of 10 CFR Part 20 limits. Dosimetry records indicated that the highest annual whole body and extremity readings since the previous inspection were 254 millirem (mrem) and 2000 mrem, respectively.

No violations were identified during this inspection.