

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

American Health Network of Indiana, LLC
10333 N. Meridian Street
Suite 450
Indianapolis, IN 47112

REPORT NUMBER(S) 12-01

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-36851

4. LICENSE NUMBER(S)

13-32555-01

5. DATE(S) OF INSPECTION

January 11, 2012

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
- (Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Robert P. Hays		1/25/12
BRANCH CHIEF	Tamara E. Bloomer		1/26/12

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: American Health Network of Indiana, LLC 10333 N. Meridian Street Suite 450 Indianapolis, IN 47112 REPORT NUMBER(S) 12-01		2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352	
3. DOCKET NUMBER(S) 030-36851	4. LICENSE NUMBER(S) 13-32555-01	5. DATE(S) OF INSPECTION January 11, 2012	
6. INSPECTION PROCEDURES USED 87131	7. INSPECTION FOCUS AREAS 03.01-03.07		

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02120	2. PRIORITY 3	3. LICENSEE CONTACT Ed Wroblewski, RSO	4. TELEPHONE NUMBER (317) 338-2381
---------------------------------	----------------------	---	---

☐ Main Office Inspection Next Inspection Date: 01/11/2015

☒ Field Office Inspection 3631 North Morrison Road, Muncie, IN

☐ Temporary Job Site Inspection

PROGRAM SCOPE

The licensee was a private medical practice with three locations in the Indianapolis, Indiana vicinity with authorization by the license to use any byproduct material as needed permitted by 10 CFR 35.100, 35.200, and 35.300, not to exceed 1 curie.

At the Muncie, IN location, one full-time nuclear medicine technologist (NMT) routinely performs an average of 3-5 diagnostic studies per day with the majority of tests being cardiac rest and stress scans. No I-131 dosages are administered, but I-123 is administered for uptake studies. Licensed material is received as unit doses from an Indianapolis, IN nuclear pharmacy.

Performance Observations

During the inspection, the licensee's available NMT demonstrated/discussed: (1) survey meter use and calibration; (2) package receiving and check-in procedures; (3) unit dose prep and safe use; (4) wipe test counting and efficiency; (5) dose calibrator tests; (6) decay-in-storage waste procedures; (7) security of licensed material; (8) sealed source inventory and leak tests; (9) radiation safety program audits; (10) contamination events (none); (11) HAZMAT refresher training; and (13) dosimetry (<10% of annual limits).

The inspector performed independent and confirmatory radiation measurements which indicated results consistent with licensee survey records and postings.