



U.S.NRC  
United States Nuclear Regulatory Commission  
*Protecting People and the Environment*

TO: file

COMPANY: N/A

# PAGES: 1 TEL.: (734) 395-7361

EMAIL: N/A

TIME	DATE
2:15 pm	January 13, 2012

ORGANIZATION  
Rochester Medical Center, P.C.

ORGANIZATION  
Rochester Medical Center, P.C.

License No.: 21-26287-01

|Control No.: 575805

We have reviewed your requesting license renewal application and find that we are unable to continue this action until we have received information regarding the following:

1. The submitted facility diagram for this application lacks dimensions and does not clearly indicate uses for areas above, below, and adjacent to the radioactive materials use area. Please resubmit diagrams, drawn to scale and indicating what that scale is. The diagram also should any room numbers, and describe what is adjacent to the radioactive materials use areas.

**RESPONSE:** The contact person indicated, via phone conversation on January 13, 2012, that the requested diagram would be submitted via facsimile. The updated diagram was received via facsimile on January 27, 2012. Accordingly, no additional information is required for this item.

2. The application was silent as to whether PET is being used at this facility. If PET is being used, additional shielding calculations may be required.

**RESPONSE:** The licensee indicated, via phone conversation of January 13, 2012, that no PET isotopes would be used. No additional information is required.

3. Under 10 CFR 35.24, a licensee's management shall appoint a Radiation Safety Officer in writing. Please provide a current, signed RSO Memorandum of Understanding/Delegation of Authority, which conforms to the requirements specified in 10 CFR 35.24(b) and 10 CFR 35.24(e). No such written statement was included with the renewal application, for Deborah Arellano, CNMT, who is currently listed as an alternate Radiation Safety Officer (RSO). According to NRC policy, we limit the number of RSO's listed on the license to one. At this time, Dr. Chae will remain as RSO on the license.

**RESPONSE:** Licensee contact indicated, via phone conversation on January 13, 2012, that he would inform the licensee of the situation, and that they may request that Ms. Arellano be added as RSO at a later date. No additional information is required.

We have requested that you submit the referenced item —

- Updated facility diagram

- via facsimile, to (630) 515-1078. Please reference the Control No. 575805, as listed at the top of this memo. The requested information was received. No additional information is required at this time.

**For future reference, please always include the name, phone number and fax number of at least one person whom we may contact for additional information when reviewing your licensing correspondence and requests.**

Please submit the requested information within **14** days of this record. **Include reference control number 575805, Please FAX your response to my attention at (630) 515-1078.** You may also scan your response and send to me via email, as a pdf file.

Please direct any questions you have to me at (630) 829-9892 or [sara.forster@nrc.gov](mailto:sara.forster@nrc.gov).

NAME OF PERSON DOCUMENTING CONVERSATION  
Sara A.B. Forster

SIGNATURE

DATE

SIGNATURE Nana A. B. Foster DATE 01/31/2012