

SARA A.B. FORSTER  
MATERIALS LICENSING BRANCH

TELECON & FAX TRANSMITTAL



NUCLEAR REGULATORY COMMISSION  
REGION III  
2443 WARRENVILLE ROAD  
LISLE, ILLINOIS 60532-4351  
(630) 829-9892 FAX: (630) 515-1078

TO: File

COMPANY: N/A, see below

# PAGES: N/A TEL.: N/A

FAX #: N/A

CONVERSATION RECORD

TIME 9:30 am DATE 2/3/2012

NAME OF PERSON(S) CONTACTED  
James Botti, Physicist

TELEPHONE NO.  
(734) 662-3197

ORGANIZATION  
MPC Physics

REPRESENTED PERSON or PERSONS  
Daryl Melvin, M.D., Radiation Safety Officer

ORGANIZATION  
Thoracic Cardiovascular Institute

SUBJECT

License No.: 21-20313-01

Control No.: 576361

SUMMARY

We have conducted a preliminary review your license amendment request and find that we are unable to continue this action until we have received information regarding the following:

1. The correspondence dated 11/22/2011 indicated that leak test results for the sealed sources would be attached to the request to remove Treadmill Room #1. However no such leak tests were enclosed.

**RESPONSE:** Per the phone conversation, the licensee's contact person indicated that the sources were retained by the licensee and were not transferred to another licensee. The contact person indicated that he would look into the leak tests but as of 02/09/2012, no leak tests were submitted. However, no residual contamination was shown in the close out surveys for the treadmill room, and no additional information is required.

2. The closeout report indicated separate wipe test results for each of five separate areas of use. The survey results were individually listed by location, but all five surveys had identical results. Please explain the uniform survey data.

**RESPONSE:** The contamination surveys for the five locations listed were counted simultaneously, which resulted in the 4 dpm reading. No additional information is required, as the treadmill room is less than 150 square feet.

We have not requested that you submit any additional items at this time. No further action is required.

Should you wish to submit any additional information, please reference the Control No. 576361, listed at the top of this memo.

Please direct any questions you have to me at (630) 829-9892 or [sara.forster@nrc.gov](mailto:sara.forster@nrc.gov).

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Sara A.B. Forster

*Sara A.B. Forster*

02/09/2012