

**From:** Lodhi, Sattar  
**Sent:** Thursday, April 26, 2012 1:40 PM  
**To:** jlewis@uesorl.com  
**Subject:** Amendment Request to your License

Mr. Jim Lewis:

This refers to your request for an amendment to your NRC license. Your request did not include details of the device-specific training for your authorized users. We will need this information in order to complete our review. Current NRC policy requires that users of the gauges, including the proposed paint analyzer be provided adequate device-specific training in the use of the device and be authorized in writing by the licensee's Radiation Safety Officer.

Additionally, I noted that your license currently authorizes 20 portable gauges. All your devices are stored at Agreement State locations. As I explained to you during our telephone conversation of April 26, 2012, you do not need to include all of your inventory in the NRC license. An Agreement State Licensee need only request the number of devices that they estimate will be in use at any one time at locations that are within NRC jurisdiction. From our discussions it appears that you may not require more than 5 portable gauges on the license to satisfy your needs. Therefore, please confirm if authorization for 5 portable gauges and the two requested analyzers will be sufficient for your needs.

Please refer to Control Number 577391 in your response to this request. Please make sure that the letter is signed by a member of your senior management. You may send the signed response via e-mail to me (as pdf file). Thank you.

Sincerely,

Sattar Lodhi  
Senior Health Physicist  
U.S. Nuclear Regulatory Commission, Region I  
King of Prussia, PA 19406

(610) 337 5364  
484 868 0988 (cell)