

SHALOM
SERVICES CORPORATION



Corporate Office: 737 Southshire Drive,
St. Louis, MO 63125
Phone: (314)894-2640
Fax: (314)894-1066

5/3/12

Material Licensing Officer
Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, Illinois 60532

CHANGE OF STORING LOCATION

This is to notify the commission that the gauge with license No 24-32655-01 is currently being stored with R.M Wester and Associate, Inc. with License No 24-20091-01. The address is 215 Indacom, St. Peters, MO 63376.
The current leakage certificate is attached.

Sincerely yours,

Johnson Ojo *5/3/12*
Johnson Ojo RSO

RECEIVED MAY 08 2012

R. M. WESTER & ASSOCIATES, INC.

215 INDACOM DRIVE - ST. PETERS, MISSOURI 63376
(636) 928-9628 - FAX 928-9857

RADIOACTIVE SEALED SOURCE LEAK TEST REPORT

Test Date: May 2, 2012

Analytical Date: May 2, 2012

Source Identification:

Manufacturer: Humboldt

Model No.: HS 5001-EZ

Mach. S/N: 5665

Radionuclide: Cs-137

Activity: 10 mCi

Source S/N: 4757CN

Radionuclide: Am-241:Be

Activity: 40 mCi

Source S/N: NJ06146

Sample Submitted by: Kenneth Bachmann

Facility: Shalom Services

Address: 737 Southshire Drive

Saint Louis Missouri 63125

The identified sealed source listed above has been tested for leakage of radioactive materials as required by the United States Nuclear Regulatory Commission. The analysis of the wipe material used in testing the sealed source reveals the presence of $\leq 1.03 \text{ E-}04$ μCi of loose contamination for Cs-137 and $\leq 3.50 \text{ E-}05$ μCi of loose contamination for Am-241.

(XX) This source is acceptable for continued use.

() This source has been found to have a level of loose contamination greater than $0.005 \mu\text{Ci}$ of removable radioactive materials, and should be removed from service immediately.

(N/A) Operational and performance check of shutter mechanism satisfactory.

Next leak test due by: November 2, 2012

Analysis by: Kenneth Bachmann, Health Physicist

Date: May 2, 2012

Reviewed by: 

Date: May 2, 2012

SHALOM SERVICES CORP.
737 SOUTHSIDE DRIVE
ST. LOUIS, MO 63125

SAINT LOUIS MO 630

04 MAY 2012 PM 3 L



Material Licensing Officer
Nuclear Regulatory Commission
2443 Warrenton Road, Suite 210
Lisle, Illinois 60532

60532+4352

