

From: Nguyen, Janice
Sent: Monday, May 14, 2012 7:19 AM
To: ralvarez@heartcareassoc.com
Subject: NRC Request for Information (Mail Control Number 576909)

Licensee: HeartCare Associates of Connecticut, LLC
License No. 06-30702-01
Docket No. 030-35890
Mail Control 576909

Dear Ms. Alvarez,

Can you please confirm that you received this email?

As per our telephone conversation on May 11, 2012, we need the following additional information to support your license renewal application:

1. Please provide the manufacturer and model number for any sealed sources that do not meet the criteria in 10 CFR 35.65 (e.g. greater than 30 millicuries).
2. You indicated that there is one survey meter per location. Please describe what happens when a meter fails to operate, and/or is out for repair/calibration.
3. Please confirm that the purpose for 10 CFR 35.100 materials is uptake, dilution, and excretion studies for which a written directive is not required.
4. Please confirm that the purpose for 10 CFR 35.200 materials is imaging and localization studies for which a written directive is not required.
5. Your application did not include information regarding survey meter and dose calibration calibration. Please confirm the following statements:
 - a. "Radiation monitoring instruments will be calibrated by a person qualified to perform survey meter calibrations."
 - b. "We reserve the right to upgrade our survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used."
 - c. "Equipment used to measure dosages will be calibrated in accordance with nationally recognized standards or the manufacturer's instructions."
6. Please confirm that you do not currently provide PET service. If you do provide PET service, please submit facility diagrams, including all installed shielding.
7. For both of your facilities in Hamden and East Haven, please indicate what areas are adjacent (e.g. all four surrounding sides), including above and below, your areas of use.

Once we receive your response, we will continue our review of your license renewal. The response letter should be signed by a member of senior management. You may fax the response to my attention to 610-337-5269, or you may email the signed response in a pdf format. Please reference Mail Control No. 576909 in your response.

Please let me know if you have any questions.

Thank you.

Sincerely,

Jan Nguyen

Janice Nguyen
Health Physicist
Nuclear Regulatory Commission Region I
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406
(610) 337-5006
FAX (610) 337-5269