

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee

In accordance with letter dated

April 9, 2012,

1. Macomb Community College
c/o Sharon Lafferty
2. 14500 E. 12 Mile Road
Warren, MI 48088-3896

3. License number 21-32692-01 is amended in its entirety to read as follows:

4. Expiration date June 30, 2018

5. Docket No. 030-37745
Reference No.

6. Byproduct, source, and/or special nuclear material

7. Chemical and/or physical form

8. Maximum amount that licensee may possess at any one time under this license

- A. Technetium-99m

- A. Unsealed organic salt/liquid

- A. 1000 millicuries

- B. Cesium-137

- B. Sealed source (Eckert & Ziegler Model FL,PF and NES- Series Flood Sources)

- B. 500 millicuries

- C. Gallium-67

- C. Unsealed organic salt/liquid

- C. 100 millicuries

- D. Thallium-201

- D. Unsealed organic salt/liquid

- D. 100 millicuries

- E. Iodine-123

- E. Unsealed organic salt/liquid

- E. 100 millicuries

- F. Indium-111

- F. Unsealed organic salt/liquid

- F. 100 millicuries

- G. Cobalt-57

- G. Sealed sources (Eckert & Ziegler Model FL,PF and NES- Series Flood Sources)

- G. 20 millicuries

9. Authorized use:

- A. through G. For student instruction, related equipment testing and quality control studies.

CONDITIONS

10. Licensed material shall be received, used and stored only at the licensee's facilities located at 44575 Garfield Road, Clinton Township, Michigan.

11. The Radiation Safety Officer for this license is James M. Botti, M.S.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number

21-32692-01

Docket or Reference Number

030-37745

Amendment No. 03

12. Licensed materials shall be used by, or under the supervision of, Ka Kit Wong, M.D.
13. The licensee is authorized to transport licensed material only in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
14. Licensed material shall not be used in or on human beings.
15. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d) for establishing decommissioning financial assurance.
16. Licensed material shall not be used in or on human beings.
17. A. Sealed sources and detector cells shall be tested for leakage and/or contamination at intervals not to exceed 6 months or at such other intervals as specified by the certificate of registration referred to in 10 CFR 32.210.
- B. In the absence of a certificate from a transferor indicating that a leak test has been made within 6 months prior to the transfer, a sealed source or detector cell received from another person shall not be put into use until tested.
- C. Sealed sources need not be leak tested if:
- (i) they contain only hydrogen-3; or
 - (ii) they contain only a radioactive gas; or
 - (iii) the half-life of the isotope is 30 days or less; or
 - (iv) they contain not more than 100 microcuries of beta and/or gamma emitting material or not more than 10 microcuries of alpha emitting material; or
 - (v) they are not designed to emit alpha particles, are in storage, and are not being used. However, when they are removed from storage for use or transferred to another person, and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source or detector cell shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
- D. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations for inspection by the Commission. Records may be disposed of following Commission inspection.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
21-32692-01Docket or Reference Number
030-37745

Amendment No. 03

- E. Tests for leakage and/or contamination shall be performed by the licensee or by other persons specifically licensed by the Commission or an Agreement State to perform such services.
18. The licensee shall conduct a physical inventory every 6 months to account for all sources and/or devices received and possessed.
19. Sealed sources containing licensed material shall not be opened or sources removed from source holders by the licensee.
20. The licensee is authorized to hold radioactive material with a physical half life of less than or equal to 120 days for decay in storage before disposal in ordinary trash provided:
- A. Before disposal as ordinary trash, byproduct material shall be surveyed at the container surface with the appropriate meter set on its most sensitive scale and with no interposed shielding to determine that its radioactivity cannot be distinguished from background. All radiation labels shall be removed or obliterated.
- B. A record of each disposal permitted under this License Condition shall be retained for 3 years. The record must include the date of disposal, the date on which the byproduct material was placed in storage, the radionuclides disposed, the survey instrument used, the background dose rate, the dose rate measured at the surface of each waste container, and the name of the individual who performed the disposal.
21. The licensee shall not acquire licensed material in a sealed source or device that contains a sealed source unless the source or device has been registered with the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or with an Agreement State.
22. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application dated April 9, 2008; and,
- B. Letters dated April 7, 2008, June 16, 2008, and August 1, 2008.

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date MAY 24 2012

By


Toye L. Simmons
Materials Licensing Branch
Region III