

**From:** [Elliott, Robin](#)  
**To:** [Hall, Gary L CPT MIL USA](#)  
**Cc:** [Nguyen, Janice](#)  
**Subject:** License #32-04054-04, Renewal for Womack Army Medical Center  
**Date:** Thursday, May 24, 2012 9:52:00 AM

---

License no: 32-04054-04  
Docket no: 030-02631  
Mail control no: 576668

Dear Captain Hall,

Thank you for your response to my May 7, 2012 communication which I received on May 22, 2012. We have reviewed your submittal and have additional questions before we can complete the renewal of your license. As we discussed on the phone, May 24, 2012 the information is described below:

- Please confirm that the opposite side of the patient rooms from the hallway is bounded by the exterior of the building.
- Please confirm that in patients will have private rooms with private sanitary facilities.
- You stated in item 5.b.3) "In the case of the adjacent room(s), should it exceed regulatory limits, the room(s) shall be removed from use for patient care until either the I-131 patient is discharged or the measured exposure limits are such that the patient in the adjacent room(s) shall not exceed regulatory limits."
  - Please confirm that the adjacent room that is removed from use is posted as a restricted area.
  - Please confirm that the patient AND the radiation levels in the adjacent room(s) shall not exceed regulatory limits.
  - Please address whether portable shielding will be used and the means to assure it is maintained in place to keep exposures below regulatory levels.

Once we receive the additional information requested, we should be able to finalize the processing of your license renewal. You may respond to my attention in writing by letter or fax (610) 337-5269 referencing mail control number 576668.

Regards,

*Robin L. Elliott*

Health Physicist  
U.S. NRC Region I  
2100 Renaissance Blvd.  
King of Prussia, PA 19406  
(610) 337-5076 voice  
(610) 337-5269 fax  
robin.elliott@nrc.gov



Think Green - Not every email needs to be printed.