

U.S. NUCLEAR REGULATORY COMMISSION**Date:**

6/12/2012

TELEPHONE CONVERSATION RECORD

Mail Control or Report No(s).	577332	License No(s).	41-06832-06	Docket No(s).	03017797
--	--------	-----------------------	-------------	----------------------	----------

Name of Licensee: Tennessee Valley Authority

Name of Participant(s): Lee Miller, TVA
Randolph Ragland, NRC RI,

Telephone No. 423-751-3197; 610-337-5083

Subject: RAI License Renewal Application
(NOTE: This will be used as the Documents Title in ADAMS)

I noted that the cover letter dated March 22, 2012, for the license renewal application included a statement that "There are no regulatory commitments made by this submittal," was contrary to the license renewal process, in that NRC specifically requests licensee commitments as part of the license renewal process (e.g., the application gets tied down as a license condition). In addition, I noted that the application did not include a source misconnect test in the quarterly maintenance process and alerted Mr. Miller to the Sentinel Service Bulletin 2004 which recommends this test be added to the quarterly maintenance procedures..

Action Required: TVA will send a letter which indicates that TVA recognizes that the March 22, 2012, license application represents a regulatory commitment.

Document Availability: ☒ **Publicly Available** ☐ **Non-Publicly Available**

<input checked="" type="checkbox"/> Non-Sensitive	<input type="checkbox"/> Non-Sensitive Copyright	<input type="checkbox"/> Sensitive- Proprietary
<input type="checkbox"/> Sensitive – Privacy Act/ PII	<input type="checkbox"/> Sensitive – Internal	<input type="checkbox"/> Sensitive – Security-Related
<input type="checkbox"/> Immediate Release	Normal Release Date: 6/20/2012	Delayed Release Date:

SUNSI Review Completed By: Randolph C. Ragland, Jr. / RA /